

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133544

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133544BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	11/21/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133547

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133547BS							
BOD5	mg/L	198	194		98	1	84.6-115.4	11/21/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133558

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133558BS							
Orthophosphate as P	mg/L	0.5	0.50		100	1	90-110	11/21/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133608

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133608BS							
TSS	mg/L	550	536		98	1	90-110	11/25/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133556

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165165BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	11/21/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4937

Project: Pre Treatment Plant

Run No.: LB133582

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165176BS							
Ammonia as N	mg/L	1	0.99		99	1	90-110	11/22/2024