



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: P4946

Project: 2024 Blanket PO #LIND7460 - Unit Price List

RunNo.: LB133588

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	11/22/2024
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	11/22/2024
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	11/22/2024

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: P4946

Project: 2024 Blanket PO #LIND7460 - Unit Price List

RunNo.: LB133620

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	11/25/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/25/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	11/25/2024



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: P4946

Project: 2024 Blanket PO #LIND7460 - Unit Price List

RunNo.: LB133620

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------