

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PSEG04
 Lab Code: CHEM Case No.: P4946 SAS No.: P4946 SDG No.: P4946
 Instrument ID: MSVOA_Y Calibration Date(s): 11/19/2024 11/19/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 15:49 17:42
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY020338.D		RRF010 = VY020339.D		RRF020 = VY020340.D			
		RRF050 = VY020341.D		RRF100 = VY020342.D		RRF150 = VY020343.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Benzene	1.632	1.376	1.440	1.348	1.378	1.435	1.435	7.2	
Toluene	0.970	0.836	0.890	0.850	0.829	0.889	0.877	6	
Ethyl Benzene	2.319	1.973	2.002	2.051	1.932	2.063	2.057	6.7	
m/p-Xylenes	0.829	0.725	0.742	0.708	0.696	0.745	0.741	6.4	
o-Xylene	0.770	0.694	0.711	0.668	0.666	0.706	0.703	5.4	
1,2-Dichloroethane-d4	0.609	0.523	0.545	0.636	0.563	0.571	0.574	7.3	
Dibromofluoromethane	0.346	0.347	0.307	0.323	0.288	0.312	0.321	7.2	
Toluene-d8	1.397	1.186	1.208	1.398	1.130	1.259	1.263	8.9	
4-Bromofluorobenzene	0.459	0.381	0.384	0.430	0.369	0.413	0.406	8.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.