

Report of Analysis

Client:	Parkway Generation Operating LLC		Date Collected:	
Project:	2024 Blanket PO #LIND7460 - Unit Price List		Date Received:	
Client Sample ID:	VY1122SBS02		SDG No.:	P4946
Lab Sample ID:	VY1122SBS02		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY020404.D	1		11/22/24 11:04	VY112224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
71-43-2	Benzene	22.5		0.72	5.00	ug/Kg
108-88-3	Toluene	22.5		0.67	5.00	ug/Kg
100-41-4	Ethyl Benzene	22.6		0.62	5.00	ug/Kg
179601-23-1	m/p-Xylenes	45.5		1.40	10.0	ug/Kg
95-47-6	o-Xylene	22.6		0.70	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.1		50 - 163	106%	SPK: 50
1868-53-7	Dibromofluoromethane	53.2		54 - 147	106%	SPK: 50
2037-26-5	Toluene-d8	52.9		58 - 134	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.6		29 - 146	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	182000	7.719			
540-36-3	1,4-Difluorobenzene	298000	8.622			
3114-55-4	Chlorobenzene-d5	247000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	113000	13.352			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products