

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PSEG04
 Lab Code: CHEM Case No.: P4946 SAS No.: P4946 SDG NO.: P4946
 Lab File ID: VY020401.D Date Analyzed: 11/22/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	114938	13.359				
UPPER LIMIT	229876	13.859				
LOWER LIMIT	57469	12.859				
EPA SAMPLE NO.						
BW-1	94206	13.35				
VY1122SBL01	94757	13.36				
VY1122SBS02	113255	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.