

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	318680
		MA No.:	SDG No.: MH2GQ3
Instrument ID:	P7	Start Date :	11/27/2024
Run Batch:	LB133687	End Date :	11/27/2024

[illegible]

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		165Ho	Q	209Bi	Q		Q		Q		Q
ZZZZZZ	1853										
ZZZZZZ	1856										
ZZZZZZ	1859										
CCV002	1902	103		96							
CCB002	1905	103		105							
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1920										
MH2GQ3	1923	108		108							
MH2GQ3D	1926	107		108							
MH2GQ3L	1929	103		105							
MH2GQ3S	1932	105		105							
MH2GS3	1935	106		106							
MH2GS4	1938	106		107							
MH2GS5	1941	108		108							
MH2GS6	1944	109		108							
MH2GS7	1947	109		106							
MH2GS8	1950	105		97							
MH2GS9	1953	107		107							
MH2GT0	1956	107		107							
MH2GT1	1959	106		105							
MH2GT4	2009	102		102							
MH2GT5	2012	101		101							
MH2GT2	2015	105		106							
MH2GT3	2018	106		107							
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
CCV003	2033	97		92							
CCB003	2036	99		102							

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 318680 MA No.: SDG No.: MH2GQ3
 Instrument ID: P7 Start Date : 11/27/2024
 Run Batch: LB133687 End Date : 11/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		209Bi	Q		Q		Q		Q		Q
S0	1610	100									
S02	1616	99									
S03	1620	94									
S04	1623	97									
S05	1625	95									
S06	1628	95									
S07	1631	94									
S08	1634	82									
ICV001	1716	100									
ICB001	1726	97									
ICSA001	1730	96									
ICSAB001	1733	98									
CCV001	1736	92									
CCB001	1738	105									
ZZZZZZ	1742										
ZZZZZZ	1745										
ZZZZZZ	1748										
ZZZZZZ	1751										
ZZZZZZ	1755										
ZZZZZZ	1758										
PBS208	1801	102									
LCS208	1804	96									
ZZZZZZ	1808										
ZZZZZZ	1811										
ZZZZZZ	1814										
ZZZZZZ	1817										
ZZZZZZ	1820										
ZZZZZZ	1823										
ZZZZZZ	1826										
ZZZZZZ	1829										
ZZZZZZ	1832										
ZZZZZZ	1835										
ZZZZZZ	1838										
ZZZZZZ	1841										
ZZZZZZ	1844										
ZZZZZZ	1847										
ZZZZZZ	1850										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		209Bi	Q		Q		Q		Q		Q
ZZZZZZ	1853										
ZZZZZZ	1856										
ZZZZZZ	1859										
CCV002	1902	88									
CCB002	1905	101									
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1920										
MH2GQ3	1923	100									
MH2GQ3D	1926	100									
MH2GQ3L	1929	98									
MH2GQ3S	1932	99									
MH2GS3	1935	98									
MH2GS4	1938	102									
MH2GS5	1941	101									
MH2GS6	1944	93									
MH2GS7	1947	99									
MH2GS8	1950	92									
MH2GS9	1953	100									
MH2GT0	1956	99									
MH2GT1	1959	98									
MH2GT4	2009	98									
MH2GT5	2012	96									
MH2GT2	2015	100									
MH2GT3	2018	100									
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
CCV003	2033	86									
CCB003	2036	99									