

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4959 SAS No.: P4959 SDG No.: P4959

Instrument ID: BNA_N Calibration Date/Time: 11/28/2024 04:51

Lab File ID: BN035368.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.617		-1.4	20.0
Fluoranthene-d10	1.134	1.065		-6.1	20.0
2-Fluorophenol	1.001	1.030		2.9	20.0
Phenol-d6	1.204	1.273		5.7	20.0
Nitrobenzene-d5	0.244	0.241		-1.2	20.0
2-Fluorobiphenyl	1.512	1.520		0.5	20.0
2,4,6-Tribromophenol	0.284	0.280		-1.4	20.0
Terphenyl-d14	0.789	0.788		-0.1	20.0
1,4-Dioxane	0.382	0.375		-1.8	20.0

All other compounds must meet a minimum RRF of 0.010.