

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
 Data File : BN035355.D
 Acq On : 27 Nov 2024 18:33
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Nov 27 22:53:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 27 22:48:24 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.308	152	2226	0.400	ng	0.00
7) Naphthalene-d8	10.052	136	5857	0.400	ng	0.00
13) Acenaphthene-d10	13.967	164	4307	0.400	ng	0.00
19) Phenanthrene-d10	16.723	188	10513	0.400	ng	#-0.01
29) Chrysene-d12	20.974	240	11206	0.400	ng	0.00
35) Perylene-d12	23.067	264	11643	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.968	112	16991	3.007	ng	0.00
5) Phenol-d6	6.513	99	21633	3.053	ng	0.00
8) Nitrobenzene-d5	8.440	82	11759	2.307	ng	0.00
11) 2-Methylnaphthalene-d10	11.656	152	29957	2.870	ng	0.00
14) 2,4,6-Tribromophenol	15.475	330	10700	3.444	ng	0.00
15) 2-Fluorobiphenyl	12.574	172	52048	2.976	ng	0.00
27) Fluoranthene-d10	18.785	212	95675	2.971	ng	0.00
31) Terphenyl-d14	19.412	244	69164	2.941	ng	0.00
Target Compounds						
2) 1,4-Dioxane	2.996	88	6360	3.146	ng	99
3) n-Nitrosodimethylamine	3.285	42	5524	2.932	ng	# 95
6) bis(2-Chloroethyl)ether	6.759	93	17762	3.339	ng	99
9) Naphthalene	10.105	128	49151	3.214	ng	97
10) Hexachlorobutadiene	10.404	225	11039	2.463	ng	# 100
12) 2-Methylnaphthalene	11.732	142	36121	3.202	ng	98
16) Acenaphthylene	13.679	152	60734	3.300	ng	99
17) Acenaphthene	14.031	154	38676	3.206	ng	99
18) Fluorene	15.026	166	55613	3.134	ng	99
21) 4-Bromophenyl-phenylether	15.941	248	20321	3.034	ng	# 93
22) Hexachlorobenzene	16.040	284	23349	3.359	ng	99
23) Atrazine	16.227	200	15072	2.508	ng	96
24) Pentachlorophenol	16.388	266	11439	3.522	ng	86
25) Phenanthrene	16.773	178	94243	3.406	ng	100
26) Anthracene	16.860	178	87657	3.449	ng	99
28) Fluoranthene	18.812	202	125901	3.309	ng	100
30) Pyrene	19.179	202	129105	3.460	ng	100
32) Benzo(a)anthracene	20.956	228	126517	3.242	ng	99
33) Chrysene	21.009	228	127439	3.297	ng	99
34) Bis(2-ethylhexyl)phtha...	20.938	149	46251	2.259	ng	99
36) Indeno(1,2,3-cd)pyrene	25.108	276	150433	3.239	ng	99
37) Benzo(b)fluoranthene	22.456	252	136261	3.478	ng	# 94
38) Benzo(k)fluoranthene	22.494	252	134789	3.439	ng	# 93
39) Benzo(a)pyrene	22.980	252	114719	3.328	ng	# 93
40) Dibenzo(a,h)anthracene	25.123	278	119216	3.240	ng	96
41) Benzo(g,h,i)perylene	25.722	276	123841	3.164	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
Data File : BN035355.D
Acq On : 27 Nov 2024 18:33
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: Nov 27 22:53:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 27 22:48:24 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.308 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2






Tgt Ion:152 Resp: 2226
Ion Ratio Lower Upper
152 100
150 153.0 124.0 186.0
115 61.0 49.6 74.4



#2
1,4-Dioxane
Concen: 3.146 ng
RT: 2.996 min Scan# 16
Delta R.T. -0.007 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion: 88 Resp: 6360
Ion Ratio Lower Upper
88 100
43 20.6 17.2 25.8
58 55.4 44.5 66.7









#3
 n-Nitrosodimethylamine
 Concen: 2.932 ng
 RT: 3.285 min Scan# 50
 Delta R.T. -0.007 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion: 42 Resp: 5524
 Ion Ratio Lower Upper
 42 100
 74 162.0 124.9 187.3
 44 6.4 2.2 3.4#



#4
 2-Fluorophenol
 Concen: 3.007 ng
 RT: 4.968 min Scan# 289
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33

Tgt Ion: 112 Resp: 16991
 Ion Ratio Lower Upper
 112 100
 64 49.7 39.8 59.8
 63 27.1 21.0 31.6









#5
Phenol-d6
Concen: 3.053 ng
RT: 6.513 min Scan# 503
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	99	Resp:	21633
Ion	Ratio	Lower	Upper
99	100		
42	14.1	11.4	17.2
71	36.5	29.3	43.9



#6
bis(2-Chloroethyl)ether
Concen: 3.339 ng
RT: 6.759 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:	93	Resp:	17762
Ion	Ratio	Lower	Upper
93	100		
63	63.6	50.4	75.6
95	32.0	25.7	38.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.052 min Scan# 915
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	136	Resp:	5857
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.2	15.2
54	6.9	6.1	9.1
68	6.9	6.4	9.6



#8
Nitrobenzene-d5
Concen: 2.307 ng
RT: 8.440 min Scan# 764
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:	82	Resp:	11759
Ion	Ratio	Lower	Upper
82	100		
128	39.0	33.4	50.0
54	39.7	36.7	55.1





#9

Naphthalene

Concen: 3.214 ng

RT: 10.105 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:128 Resp: 49151

Ion Ratio Lower Upper

128 100

129 11.0 9.8 14.6

127 13.3 11.4 17.2



#10

Hexachlorobutadiene

Concen: 2.463 ng

RT: 10.404 min Scan# 948

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:225 Resp: 11039

Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.4 51.3 76.9









#11
2-Methylnaphthalene-d10
Concen: 2.870 ng
RT: 11.656 min Scan# 1081
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:152 Resp: 29957
Ion Ratio Lower Upper
152 100
151 20.6 16.6 25.0



#12
2-Methylnaphthalene
Concen: 3.202 ng
RT: 11.732 min Scan# 1096
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:142 Resp: 36121
Ion Ratio Lower Upper
142 100
141 88.7 72.2 108.4
115 37.7 31.4 47.0





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 13.967 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:164 Resp: 4307

Ion Ratio Lower Upper

164 100

162 102.0 82.2 123.2

160 49.7 40.1 60.1



#14

2,4,6-Tribromophenol

Concen: 3.444 ng

RT: 15.475 min Scan# 1578

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:330 Resp: 10700

Ion Ratio Lower Upper

330 100

332 0.0 0.0 0.0

141 33.7 26.6 40.0









#15
2-Fluorobiphenyl
Concen: 2.976 ng
RT: 12.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:172 Resp: 52048
Ion Ratio Lower Upper
172 100
171 35.0 29.0 43.4
170 23.1 19.8 29.8



#16
Acenaphthylene
Concen: 3.300 ng
RT: 13.679 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:152 Resp: 60734
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.2
153 12.9 10.4 15.6









#17
Acenaphthene
Concen: 3.206 ng
RT: 14.031 min Scan# 1443
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 38676
Ion Ratio Lower Upper
154 100
153 115.5 92.6 139.0
152 59.2 49.0 73.6



#18
Fluorene
Concen: 3.134 ng
RT: 15.026 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:166 Resp: 55613
Ion Ratio Lower Upper
166 100
165 98.5 79.7 119.5
167 13.3 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.723 min Scan# 1

Delta R.T. -0.012 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 10513

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.9 4.6 6.8#



#21

4-Bromophenyl-phenylether

Concen: 3.034 ng

RT: 15.941 min Scan# 1616

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:248 Resp: 20321


Ion Ratio Lower Upper

248 100

250 104.3 80.6 120.8

141 30.7 31.5 47.3#









#22
Hexachlorobenzene
Concen: 3.359 ng
RT: 16.040 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:284 Resp: 23349
Ion Ratio Lower Upper
284 100
142 33.3 26.7 40.1
249 30.2 24.6 36.8



#23
Atrazine
Concen: 2.508 ng
RT: 16.227 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:200 Resp: 15072
Ion Ratio Lower Upper
200 100
173 26.2 24.1 36.1
215 45.1 36.9 55.3





#24

Pentachlorophenol

Concen: 3.522 ng

RT: 16.388 min Scan# 1652

Delta R.T. -0.012 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:266 Resp: 11439

Ion Ratio Lower Upper

266 100

264 62.6 42.3 63.5

268 64.4 43.3 64.9



#25

Phenanthrene

Concen: 3.406 ng

RT: 16.773 min Scan# 1683

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:178 Resp: 94243


Ion Ratio Lower Upper

178 100

176 19.3 15.4 23.2

179 15.3 12.3 18.5









#26
 Anthracene
 Concen: 3.449 ng
 RT: 16.860 min Scan# 1690
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	178	Resp:	87657
Ion Ratio	Lower	Upper	
178	100		
176	18.6	15.0	22.6
179	15.1	12.6	18.8



#27
 Fluoranthene-d10
 Concen: 2.971 ng
 RT: 18.785 min Scan# 1882
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33

Tgt Ion:	212	Resp:	95675
Ion Ratio	Lower	Upper	
212	100		
106	11.4	9.2	13.8
104	6.4	5.3	7.9





#28

Fluoranthene

Concen: 3.309 ng

RT: 18.812 min Scan# 1888

Delta R.T. -0.005 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:202 Resp: 125901

Ion Ratio Lower Upper

202 100

101 9.1 7.4 11.0

203 17.3 13.7 20.5



#29

Chrysene-d12

Concen: 0.400 ng

RT: 20.974 min Scan# 2253

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:240 Resp: 11206


Ion Ratio Lower Upper

240 100

120 9.3 7.9 11.9

236 28.3 22.9 34.3









#30
 Pyrene
 Concen: 3.460 ng
 RT: 19.179 min Scan# 1967
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:202 Resp: 129105




Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.0	25.4
203	17.7	14.3	21.5




#31
 Terphenyl-d14
 Concen: 2.941 ng
 RT: 19.412 min Scan# 2017
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33

Tgt Ion:244 Resp: 69164

Ion	Ratio	Lower	Upper
244	100		
212	9.2	8.1	12.1
122	12.0	10.3	15.5









#32
Benzo(a)anthracene
Concen: 3.242 ng
RT: 20.956 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 126517
Ion Ratio Lower Upper
228 100
226 27.6 22.5 33.7
229 19.4 15.8 23.8



#33
Chrysene
Concen: 3.297 ng
RT: 21.009 min Scan# 2257
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:228 Resp: 127439
Ion Ratio Lower Upper
228 100
226 30.5 24.6 37.0
229 19.4 15.9 23.9









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 2.259 ng
 RT: 20.938 min Scan# 21
 Delta R.T. 0.009 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	149	Resp:	46251
Ion Ratio	Lower	Upper	
149	100		
167	28.3	22.2	33.4
279	3.2	2.7	4.1



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.067 min Scan# 2639
 Delta R.T. -0.000 min
 Lab File: BN035355.D
 Acq: 27 Nov 2024 18:33

Tgt Ion:	264	Resp:	11643
Ion Ratio	Lower	Upper	
264	100		
260	27.1	21.4	32.2
265	47.7	40.2	60.4





#36

Indeno(1,2,3-cd)pyrene

Concen: 3.239 ng

RT: 25.108 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 150433

Ion Ratio Lower Upper

276 100

138 23.7 19.4 29.0

277 24.9 19.8 29.6



#37

Benzo(b)fluoranthene

Concen: 3.478 ng

RT: 22.456 min Scan# 2430

Delta R.T. 0.003 min

Lab File: BN035355.D

Acq: 27 Nov 2024 18:33




Tgt Ion:252 Resp: 136261


Ion Ratio Lower Upper

252 100

253 22.0 19.6 29.4


125 9.0 9.6 14.4#







#38
Benzo(k)fluoranthene
Concen: 3.439 ng
RT: 22.494 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 134789
Ion Ratio Lower Upper
252 100
253 21.8 19.5 29.3
125 9.2 10.2 15.4#



#39
Benzo(a)pyrene
Concen: 3.328 ng
RT: 22.980 min Scan# 2609
Delta R.T. 0.003 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:252 Resp: 114719
Ion Ratio Lower Upper
252 100
253 21.9 20.2 30.4
125 10.1 10.9 16.3#









#40
Dibenzo(a,h)anthracene
Concen: 3.240 ng
RT: 25.123 min Scan# 31
Delta R.T. -0.003 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:278 Resp: 119216
Ion Ratio Lower Upper
278 100
139 16.1 14.2 21.4
279 23.7 20.5 30.7



#41
Benzo(g,h,i)perylene
Concen: 3.164 ng
RT: 25.722 min Scan# 3547
Delta R.T. -0.000 min
Lab File: BN035355.D
Acq: 27 Nov 2024 18:33

Tgt Ion:276 Resp: 123841
Ion Ratio Lower Upper
276 100
277 23.6 19.9 29.9
138 20.6 17.8 26.8

