

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	11/21/24
Project:	CTO WE13	Date Received:	11/22/24
Client Sample ID:	RW5-SP303-20241121	SDG No.:	P4959
Lab Sample ID:	P4959-03	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035299.D	1	11/22/24 12:25	11/26/24 02:38	PB165198

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.45		30 - 150		113%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.23		55 - 111		57%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.24		53 - 106		60%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.73	*	58 - 132		182%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	4030	7.315				
1146-65-2	Naphthalene-d8	10600	10.063				
15067-26-2	Acenaphthene-d10	6680	13.966				
1517-22-2	Phenanthrene-d10	14000	16.734				
1719-03-5	Chrysene-d12	880	21.008				
1520-96-3	Perylene-d12	7740	23.078				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products