

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4959 SAS No.: P4959 SDG No.: P4959

Instrument ID: BNA_N Calibration Date/Time: 11/28/2024 09:02

Lab File ID: BN035375.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.614		-1.9	50.0
Fluoranthene-d10	1.134	1.056		-6.9	50.0
2-Fluorophenol	1.001	1.006		0.5	50.0
Phenol-d6	1.204	1.212		0.7	50.0
Nitrobenzene-d5	0.244	0.244		0.0	50.0
2-Fluorobiphenyl	1.512	1.517		0.3	50.0
2,4,6-Tribromophenol	0.284	0.266		-6.3	50.0
Terphenyl-d14	0.789	0.787		-0.3	50.0
1,4-Dioxane	0.382	0.378		-1.0	50.0

All other compounds must meet a minimum RRF of 0.010.