

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112524\
 Data File : BN035279.D
 Acq On : 25 Nov 2024 11:28
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Nov 25 14:13:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 14:13:00 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.315	152	2200	0.400	ng	0.00
7) Naphthalene-d8	10.063	136	5326	0.400	ng	0.00
13) Acenaphthene-d10	13.976	164	2947	0.400	ng	0.00
19) Phenanthrene-d10	16.734	188	6895	0.400	ng	0.00
29) Chrysene-d12	21.008	240	586	0.400	ng	0.00
35) Perylene-d12	23.078	264	4692	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.968	112	1735	0.311	ng	0.00
5) Phenol-d6	6.521	99	2431	0.347	ng	0.00
8) Nitrobenzene-d5	8.451	82	1846	0.398	ng	0.00
11) 2-Methylnaphthalene-d10	11.667	152	2873	0.303	ng	0.00
14) 2,4,6-Tribromophenol	15.484	330	531	0.250	ng	0.00
15) 2-Fluorobiphenyl	12.608	172	721	0.060	ng	0.00
27) Fluoranthene-d10	18.792	212	6669	0.316	ng	0.00
31) Terphenyl-d14	19.419	244	4254	3.460	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.003	88	631	0.316	ng	100
3) n-Nitrosodimethylamine	3.300	42	624	0.335	ng	100
6) bis(2-Chloroethyl)ether	6.766	93	1971	0.375	ng	100
9) Naphthalene	10.116	128	5810	0.418	ng	100
10) Hexachlorobutadiene	10.415	225	1664	0.408	ng	# 100
12) 2-Methylnaphthalene	11.742	142	3623	0.353	ng	100
16) Acenaphthylene	13.688	152	5556	0.441	ng	100
17) Acenaphthene	14.030	154	3465	0.420	ng	100
18) Fluorene	15.035	166	4816	0.397	ng	100
20) 4,6-Dinitro-2-methylph...	15.131	198	233	0.162	ng	100
21) 4-Bromophenyl-phenylether	15.940	248	1623	0.369	ng	100
22) Hexachlorobenzene	16.051	284	2429	0.533	ng	100
23) Atrazine	16.238	200	689	0.175	ng	100
24) Pentachlorophenol	16.399	266	544	0.255	ng	100
25) Phenanthrene	16.784	178	7798	0.430	ng	100
26) Anthracene	16.871	178	6280	0.377	ng	100
28) Fluoranthene	18.820	202	9240	0.370	ng	100
30) Pyrene	19.192	202	8957	4.590	ng	100
32) Benzo(a)anthracene	21.017	228	8100	3.970	ng	100
33) Chrysene	21.017	228	8100	4.007	ng	100
34) Bis(2-ethylhexyl)phtha...	21.017	149	533	0.498	ng	# 100
36) Indeno(1,2,3-cd)pyrene	25.122	276	7146	0.382	ng	100
37) Benzo(b)fluoranthene	22.464	252	6813	0.432	ng	100
38) Benzo(k)fluoranthene	22.505	252	7659	0.485	ng	100
39) Benzo(a)pyrene	22.990	252	5365	0.386	ng	100
40) Dibenzo(a,h)anthracene	25.139	278	5194	0.350	ng	100
41) Benzo(g,h,i)perylene	25.738	276	5823	0.369	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112524\
Data File : BN035279.D
Acq On : 25 Nov 2024 11:28
Operator : RC/JU
Sample : SSTDICC0.4
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.4

Quant Time: Nov 25 14:13:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 14:13:00 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.315 min Scan# 614
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4


Tgt Ion:152 Resp: 2200
 Ion Ratio Lower Upper
 152 100
 150 155.6 124.5 186.7
 115 59.8 47.8 71.8



#2
 1,4-Dioxane
 Concen: 0.316 ng
 RT: 3.003 min Scan# 17
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28

Tgt Ion: 88 Resp: 631
 Ion Ratio Lower Upper
 88 100
 43 60.1 48.1 72.1
 58 58.6 46.9 70.3









#3
n-Nitrosodimethylamine
Concen: 0.335 ng
RT: 3.300 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 624
Ion Ratio Lower Upper
42 100
74 128.4 102.7 154.1
44 42.6 34.1 51.1



#4
2-Fluorophenol
Concen: 0.311 ng
RT: 4.968 min Scan# 289
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion: 112 Resp: 1735
Ion Ratio Lower Upper
112 100
64 50.5 40.4 60.6
63 29.8 23.8 35.8









#5
Phenol-d6
Concen: 0.347 ng
RT: 6.521 min Scan# 504
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 2431
Ion Ratio Lower Upper
99 100
42 16.0 12.8 19.2
71 39.3 31.4 47.2



#6
bis(2-Chloroethyl)ether
Concen: 0.375 ng
RT: 6.766 min Scan# 538
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion: 93 Resp: 1971
Ion Ratio Lower Upper
93 100
63 65.4 52.3 78.5
95 31.9 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.063 min Scan# 916
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	5326
Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.7	14.5
54	6.6	5.3	7.9
68	7.4	5.9	8.9



#8
Nitrobenzene-d5
Concen: 0.398 ng
RT: 8.451 min Scan# 765
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:	82	Resp:	1846
Ion	Ratio	Lower	Upper
82	100		
128	41.2	33.0	49.4
54	41.6	33.3	49.9









#9
Naphthalene
Concen: 0.418 ng
RT: 10.116 min Scan# 921
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 5810
Ion Ratio Lower Upper
128 100
129 12.0 9.6 14.4
127 14.1 11.3 16.9



#10
Hexachlorobutadiene
Concen: 0.408 ng
RT: 10.415 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:225 Resp: 1664
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.3 76.9









#11
2-Methylnaphthalene-d10
Concen: 0.303 ng
RT: 11.667 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2873
Ion Ratio Lower Upper
152 100
151 21.5 17.2 25.8



#12
2-Methylnaphthalene
Concen: 0.353 ng
RT: 11.742 min Scan# 1114
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:142 Resp: 3623
Ion Ratio Lower Upper
142 100
141 89.8 71.8 107.8
115 40.8 32.6 49.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 13.976 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 2947
Ion Ratio Lower Upper
164 100
162 102.5 82.0 123.0
160 78.7 63.0 94.4



#14
2,4,6-Tribromophenol
Concen: 0.250 ng
RT: 15.484 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:330 Resp: 531
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.5 32.4 48.6









#15
2-Fluorobiphenyl
Concen: 0.060 ng
RT: 12.608 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	721
Ion Ratio	Lower	Upper	
172	100		
171	45.8	36.6	55.0
170	37.1	29.7	44.5



#16
Acenaphthylene
Concen: 0.441 ng
RT: 13.688 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:	152	Resp:	5556
Ion Ratio	Lower	Upper	
152	100		
151	19.9	15.9	23.9
153	12.8	10.2	15.4









#17
Acenaphthene
Concen: 0.420 ng
RT: 14.030 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 3465
Ion Ratio Lower Upper
154 100
153 120.1 96.1 144.1
152 90.2 72.2 108.4



#18
Fluorene
Concen: 0.397 ng
RT: 15.035 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:166 Resp: 4816
Ion Ratio Lower Upper
166 100
165 101.5 81.2 121.8
167 13.1 10.5 15.7





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.734 min Scan# 1713

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 6895

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 0.162 ng

RT: 15.131 min Scan# 1579

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28




Tgt Ion:198 Resp: 233


Ion Ratio Lower Upper

198 100

51 67.0 53.6 80.4

105 59.3 47.4 71.2









#21
4-Bromophenyl-phenylether
Concen: 0.369 ng
RT: 15.940 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	1623
Ion Ratio	Lower	Upper	
248	100		
250	93.5	74.8	112.2
141	73.7	59.0	88.4



#22
Hexachlorobenzene
Concen: 0.533 ng
RT: 16.051 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:	284	Resp:	2429
Ion Ratio	Lower	Upper	
284	100		
142	35.9	28.7	43.1
249	31.7	25.4	38.0









#23
 Atrazine
 Concen: 0.175 ng
 RT: 16.238 min Scan# 1673
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4


Tgt Ion:	200	Resp:	689
Ion Ratio	Lower	Upper	
200	100		
173	32.3	25.8	38.8
215	48.3	38.6	58.0



#24
 Pentachlorophenol
 Concen: 0.255 ng
 RT: 16.399 min Scan# 1686
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28

Tgt Ion:	266	Resp:	544
Ion Ratio	Lower	Upper	
266	100		
264	61.6	49.3	73.9
268	60.8	48.6	73.0









#25
Phenanthrene
Concen: 0.430 ng
RT: 16.784 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 7798
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.0
179 15.7 12.6 18.8



#26
Anthracene
Concen: 0.377 ng
RT: 16.871 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:178 Resp: 6280
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.5 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.316 ng

RT: 18.792 min Scan# 1917

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 6669

Ion Ratio Lower Upper

212 100

106 12.3 9.8 14.8

104 7.0 5.6 8.4



#28

Fluoranthene

Concen: 0.370 ng

RT: 18.820 min Scan# 1923

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28




Tgt Ion:202 Resp: 9240


Ion Ratio Lower Upper

202 100

101 9.5 7.6 11.4

203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.008 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 586
Ion Ratio Lower Upper
240 100
120 68.6 54.9 82.3
236 67.5 54.0 81.0



#30
Pyrene
Concen: 4.590 ng
RT: 19.192 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:202 Resp: 8957
Ion Ratio Lower Upper
202 100
200 21.0 16.8 25.2
203 18.0 14.4 21.6





#31

Terphenyl-d14

Concen: 3.460 ng

RT: 19.419 min Scan# 2052

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:244 Resp: 4254

Ion Ratio Lower Upper

244 100

212 11.7 9.4 14.0

122 16.2 13.0 19.4



#32

Benzo(a)anthracene

Concen: 3.970 ng

RT: 21.017 min Scan# 2395

Delta R.T. 0.000 min

Lab File: BN035279.D

Acq: 25 Nov 2024 11:28




Tgt Ion:228 Resp: 8100


Ion Ratio Lower Upper

228 100

226 30.9 24.7 37.1

229 20.2 16.2 24.2









#33
 Chrysene
 Concen: 4.007 ng
 RT: 21.017 min Scan# 2395
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28




Instrument : BNA_N
 ClientSampleId : SSTDICCC0.4


Tgt Ion:	228	Resp:	8100
Ion Ratio	Lower	Upper	
228	100		
226	30.9	24.7	37.1
229	20.2	16.2	24.2



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.498 ng
 RT: 21.017 min Scan# 2395
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28

Tgt Ion:	149	Resp:	533
Ion Ratio	Lower	Upper	
149	100		
167	0.0	0.0	0.0
279	16.3	13.0	19.6









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.078 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:264 Resp: 4692
 Ion Ratio Lower Upper
 264 100
 260 27.9 22.3 33.5
 265 85.8 68.6 103.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.382 ng
 RT: 25.122 min Scan# 3479
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28

Tgt Ion:276 Resp: 7146
 Ion Ratio Lower Upper
 276 100
 138 23.5 18.8 28.2
 277 22.8 18.2 27.4









#37
Benzo(b)fluoranthene
Concen: 0.432 ng
RT: 22.464 min Scan# 2570
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 6813
Ion Ratio Lower Upper
252 100
253 25.2 20.2 30.2
125 15.0 12.0 18.0



#38
Benzo(k)fluoranthene
Concen: 0.485 ng
RT: 22.505 min Scan# 2584
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:252 Resp: 7659
Ion Ratio Lower Upper
252 100
253 25.1 20.1 30.1
125 15.9 12.7 19.1









#39
Benzo(a)pyrene
Concen: 0.386 ng
RT: 22.990 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	252	Resp:	5365
Ion Ratio	Lower	Upper	
252	100		
253	27.9	22.3	33.5
125	19.1	15.3	22.9



#40
Dibenzo(a,h)anthracene
Concen: 0.350 ng
RT: 25.139 min Scan# 3485
Delta R.T. 0.000 min
Lab File: BN035279.D
Acq: 25 Nov 2024 11:28

Tgt Ion:	278	Resp:	5194
Ion Ratio	Lower	Upper	
278	100		
139	21.3	17.0	25.6
279	28.7	23.0	34.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.369 ng
 RT: 25.738 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BN035279.D
 Acq: 25 Nov 2024 11:28

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	5823
Ion	Ratio	Lower	Upper
276	100		
277	26.4	21.1	31.7
138	24.8	19.8	29.8

