

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4959 SAS No.: P4959 SDG NO.: P4959

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 11/25/2024

Lab File ID: BN035293.D Time Analyzed: 23:02

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3724	7.315	9273	10.06	5865	13.97
UPPER LIMIT	7448	7.815	18546	10.563	11730	14.473
LOWER LIMIT	1862	6.815	4636.5	9.563	2932.5	13.473
EPA SAMPLE NO.						
01 RW5-SP201-20241121	3353	7.32	8974	10.06	7827	13.97
02 RW5-SP303-20241121	4025	7.32	10571	10.06	6682	13.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.