



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**CHA Companies, Inc.**

**Project Name: OMH Tank Pull-011 - 078673-01 C-03**

**Project # N/A**

**Chemtech Project # P4960**

**Test Name: SVOCMS Group1**

### **A. Number of Samples and Date of Receipt:**

6 Solid samples were received on 11/22/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group1 and VOCMS Group1. This data package contains results for SVOCMS Group1.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD for {P4954-01MSD} with File ID: BF140626.D met criteria except for Benzo(k)fluoranthene[25%] . Due to Difference in MS and MSD concentrations.

The Blank Spike for {PB165185BS} with File ID: BF140602.D met requirements for all samples except for Acenaphthylene[106%], Anthracene[106%],

Benzo(a)anthracene[106%], Benzo(a)pyrene[112%] and Chrysene[106%] . But associated samples have not positive hit for these compounds therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

### **E. Additional Comments:**

The soil samples results are based on a dry weight basis.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_