

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CLOU03
 Lab Code: CHEM Case No.: P4960 SAS No.: P4960 SDG NO.: P4960
 Lab File ID: VX044024.D Date Analyzed: 11/27/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	130547	5.54	218884	6.75	184783	10.05
UPPER LIMIT	261094	6.037	437768	7.251	369566	10.549
LOWER LIMIT	65273.5	5.037	109442	6.251	92391.5	9.549
EPA SAMPLE NO.						
SW3	101230	5.54	203097	6.76	182737	10.06
VX1127MBL01	119396	5.54	228047	6.75	207542	10.05
VX1127MBS01	117049	5.54	213408	6.76	180534	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.