

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CLOU03
 Lab Code: CHEM Case No.: P4960 SAS No.: P4960 SDG NO.: P4960
 Lab File ID: VX044024.D Date Analyzed: 11/27/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	86408	12.018				
UPPER LIMIT	172816	12.518				
LOWER LIMIT	43204	11.518				
EPA SAMPLE NO.						
SW3	82804	12.02				
VX1127MBL01	87354	12.02				
VX1127MBS01	80899	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.