

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CLOU03
Lab Code: CHEM Case No.: P4960 SAS No.: P4960 SDG NO.: P4960
Lab File ID: VY020400.D BFB Injection Date: 11/22/2024
Instrument ID: MSVOA_Y BFB Injection Time: 08:56
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	55.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	73.8
175	5.0 - 9.0% of mass 174	5.7 (7.7) 1
176	95.0 - 101.0% of mass 174	72.5 (98.3) 1
177	5.0 - 9.0% of mass 176	5 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY020401.D	11/22/2024	09:40
VY1122SBL01	VY1122SBL01	VY020402.D	11/22/2024	10:15
VY1122SBS02	VY1122SBS02	VY020404.D	11/22/2024	11:04
B1	P4960-01	VY020414.D	11/22/2024	15:36
B2	P4960-02	VY020415.D	11/22/2024	16:00