

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112124

Review By	John Carlone	Review On	11/21/2024 9:44:50 AM
Supervise By	Maresh Dadoda	Supervise On	11/22/2024 3:34:00 PM
SubDirectory	VX112124	HP Acquire Method	HP Processing Method 82X112124W.M

STD. NAME	STD REF.#
Tune/Reschk	VP131686,VP131690
Initial Calibration Stds	VP131716,VP131717,VP131718,VP131719,VP131720,VP131721
CCC	VP131691
Internal Standard/PEM	
ICV/I.BLK	VP131722
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043915.D	20 Nov 2024 15:43		JC/MD	Not Ok
2	VSTDICC001	VSTDICC001	VX043916.D	20 Nov 2024 16:11	%D failed for Comp. #71,74,90,92 in 01 PPB	JC/MD	Not Ok
3	VSTDICC005	VSTDICC005	VX043917.D	20 Nov 2024 16:57	%D failed for Comp. #81 in 05 PPB	JC/MD	Not Ok
4	VSTDICC020	VSTDICC020	VX043918.D	20 Nov 2024 17:20		JC/MD	Not Ok
5	VSTDICCC050	VSTDICCC050	VX043919.D	20 Nov 2024 17:43		JC/MD	Not Ok
6	VSTDICC100	VSTDICC100	VX043920.D	20 Nov 2024 18:06		JC/MD	Not Ok
7	VSTDICC150	VSTDICC150	VX043921.D	20 Nov 2024 18:30		JC/MD	Not Ok
8	IBLK	IBLK	VX043922.D	20 Nov 2024 18:53		JC/MD	Not Ok
9	VSTDICV050	ICVVX112124	VX043923.D	20 Nov 2024 19:16	Failed	JC/MD	Not Ok
10	BFB	BFB	VX043924.D	21 Nov 2024 08:51		JC/MD	Ok
11	VSTDICC001	VSTDICC001	VX043925.D	21 Nov 2024 09:47	%D failed for Comp. #71 in 01 PPB	JC/MD	Ok,M
12	VSTDICC005	VSTDICC005	VX043926.D	21 Nov 2024 10:14	QR- 58,71	JC/MD	Ok,M
13	VSTDICC020	VSTDICC020	VX043927.D	21 Nov 2024 10:37		JC/MD	Ok,M
14	VSTDICCC050	VSTDICCC050	VX043928.D	21 Nov 2024 11:17		JC/MD	Ok,M
15	VSTDICC100	VSTDICC100	VX043929.D	21 Nov 2024 11:40		JC/MD	Ok,M
16	VSTDICC150	VSTDICC150	VX043930.D	21 Nov 2024 12:03		JC/MD	Ok,M
17	IBLK	IBLK	VX043931.D	21 Nov 2024 12:27		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112124

Review By	John Carlone	Review On	11/21/2024 9:44:50 AM
Supervise By	Maresh Dadoda	Supervise On	11/22/2024 3:34:00 PM
SubDirectory	VX112124	HP Acquire Method	HP Processing Method 82X112124W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP131686,VP131690		
Initial Calibration Stds	VP131716,VP131717,VP131718,VP131719,VP131720,VP131721		
CCC	VP131691		
Internal Standard/PEM			
ICV/I.BLK	VP131722		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	VSTDICV050	ICVVX112124	VX043932.D	21 Nov 2024 13:57		JC/MD	Ok,M
19	VX1121MBL01	VX1121MBL01	VX043933.D	21 Nov 2024 14:27		JC/MD	Ok
20	VX1121WBL01	VX1121WBL01	VX043934.D	21 Nov 2024 14:50		JC/MD	Ok
21	VX1121WBS01	VX1121WBS01	VX043935.D	21 Nov 2024 15:13		JC/MD	Ok,M
22	VX1121WBSD01	VX1121WBSD01	VX043936.D	21 Nov 2024 15:40		JC/MD	Ok,M
23	P4921-01	WC-11-A-202411	VX043937.D	21 Nov 2024 16:03	vial B pH#5.0	JC/MD	Ok
24	P4934-01	TB-01-20241118	VX043938.D	21 Nov 2024 16:27	vial A pH<2 TB;	JC/MD	Ok
25	P4934-09	FB-01-20241119	VX043939.D	21 Nov 2024 16:50	vial A pH<2 FB;	JC/MD	Ok
26	P4934-11	RB-01-20241119	VX043940.D	21 Nov 2024 17:13	vial A pH<2	JC/MD	Ok
27	P4934-02	BPOW6-11-20241118	VX043941.D	21 Nov 2024 17:36	vial A pH<2	JC/MD	Ok
28	P4934-03	BPOW6-7-20241118	VX043942.D	21 Nov 2024 17:59	vial A pH<2	JC/MD	Ok
29	P4934-08	BPOW6-9-20241119	VX043943.D	21 Nov 2024 18:23	vial A pH<2	JC/MD	Ok
30	P4934-10	BPOW6-8-20241119	VX043944.D	21 Nov 2024 18:46	vial A pH<2	JC/MD	Ok
31	P4934-04	DUP-01-20241118	VX043945.D	21 Nov 2024 19:09	vial A pH<2	JC/MD	Ok
32	P4934-12	DUP-02-20241119	VX043946.D	21 Nov 2024 19:32	vial A pH<2	JC/MD	Ok
33	P4934-05	BPOW6-10-20241119	VX043947.D	21 Nov 2024 19:55	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX043948.D	21 Nov 2024 20:18		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120324

Review By	John Carlone	Review On	12/4/2024 10:04:27 AM
Supervise By	Maresh Dadoda	Supervise On	12/4/2024 2:27:51 PM
SubDirectory	VX120324	HP Acquire Method	HP Processing Method 82X112124W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131874
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131875,VP131876

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044079.D	03 Dec 2024 09:09		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044080.D	03 Dec 2024 09:44	V13516	JC/MD	Ok,M
3	VX1203MBL01	VX1203MBL01	VX044081.D	03 Dec 2024 10:11		JC/MD	Ok
4	VX1203WBL01	VX1203WBL01	VX044082.D	03 Dec 2024 10:34		JC/MD	Ok
5	VX1203MBS01	VX1203MBS01	VX044083.D	03 Dec 2024 10:57		JC/MD	Ok,M
6	VX1203MBS02	VX1203MBS02	VX044084.D	03 Dec 2024 11:24		JC/MD	Ok,M
7	P5005-01	STOCK-PILE	VX044085.D	03 Dec 2024 11:47		JC/MD	Ok
8	P4960-05	SW3	VX044086.D	03 Dec 2024 12:10		JC/MD	Ok
9	VX1203WBS01	VX1203WBS01	VX044087.D	03 Dec 2024 13:01		JC/MD	Ok,M
10	P5006-01RE	SPLP-C4-2RE	VX044088.D	03 Dec 2024 13:24	vial B pH#6.0	JC/MD	Confirms
11	P5006-02	SPLP-C4-3	VX044089.D	03 Dec 2024 13:47	vial B pH#6.0	JC/MD	Ok,M
12	VX1203WBS02	VX1203WBS02	VX044090.D	03 Dec 2024 14:10	Recovery fail	JC/MD	Not Ok
13	P5006-03	SPLP-C5-1	VX044091.D	03 Dec 2024 14:33	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
14	VX1203WBS03	VX1203WBS03	VX044092.D	03 Dec 2024 14:56	Recovery fail	JC/MD	Not Ok
15	P5006-05	SPLP-C5-3	VX044093.D	03 Dec 2024 15:19	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
16	P5006-06	SPLP-C6-1	VX044094.D	03 Dec 2024 15:41	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
17	P5006-07	SPLP-C6-2	VX044095.D	03 Dec 2024 16:04	vial A pH#6.0	JC/MD	Ok
18	P5006-08	SPLP-C6-3	VX044096.D	03 Dec 2024 16:27	vial A pH#6.0 Surrogate fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120324

Review By	John Carlone	Review On	12/4/2024 10:04:27 AM
Supervise By	Maresh Dadoda	Supervise On	12/4/2024 2:27:51 PM
SubDirectory	VX120324	HP Acquire Method	HP Processing Method 82X112124W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131874		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131875,VP131876		

19	P5006-09	SPLP-C10-1	VX044097.D	03 Dec 2024 16:51	vial A pH#6.0 Surrogate fail; Need 5X	JC/MD	Dilution
20	P5006-10	SPLP-C10-2	VX044098.D	03 Dec 2024 17:14	vial A pH#6.0 Surrogate fail; Need 20X	JC/MD	Dilution
21	P5006-11	SPLP-C10-3	VX044099.D	03 Dec 2024 17:36	vial A pH#6.0 Need 20X	JC/MD	Dilution
22	P5006-12	SPLP-C11-1	VX044100.D	03 Dec 2024 18:00	vial A pH#6.0 E Flag in Previous Sample; Surrogate fail; Need 5X	JC/MD	Dilution
23	P5006-13	SPLP-C11-2	VX044101.D	03 Dec 2024 18:23	vial A pH#6.0 Surrogate fail; Concentration carryover	JC/MD	Not Ok
24	P5006-14	SPLP-C11-3	VX044102.D	03 Dec 2024 18:46	vial A pH#6.0 Surrogate fail; concentration carryover	JC/MD	Not Ok
25	P5006-15	SPLP-C12-1	VX044103.D	03 Dec 2024 19:09	vial A pH#6.0 Surrogate fail; Concentration carryover	JC/MD	Not Ok
26	P5006-16	SPLP-C12-2	VX044104.D	03 Dec 2024 19:32	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
27	P5006-17	SPLP-C12-3	VX044105.D	03 Dec 2024 19:55	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
28	VX1203WBSD01	VX1203WBSD01	VX044106.D	03 Dec 2024 20:18		JC/MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VX044107.D	03 Dec 2024 20:41		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY111924

Review By	Maresh Dadoda	Review On	11/20/2024 1:01:03 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/20/2024 1:03:18 PM
SubDirectory	VY111924	HP Acquire Method	HP Processing Method 82y111924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131638 VP131639,VP131640,VP131641,VP131642,VP131643,VP131644 VP131645

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020337.D	19 Nov 2024 14:40		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY020338.D	19 Nov 2024 15:49		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY020339.D	19 Nov 2024 16:12		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY020340.D	19 Nov 2024 16:35		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY020341.D	19 Nov 2024 16:57	Comp.#3 is on Linear Regression	SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY020342.D	19 Nov 2024 17:20	Comp.#11 is on Quadratic Regression	SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY020343.D	19 Nov 2024 17:42		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020344.D	19 Nov 2024 18:06		SY/MD	Ok
9	VSTDICV050	ICVVY111924	VY020345.D	19 Nov 2024 18:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112224

Review By	Maresh Dadoda	Review On	11/25/2024 2:01:45 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/25/2024 2:05:55 PM
SubDirectory	VY112224	HP Acquire Method	HP Processing Method

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131732
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131733,VP131734 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020400.D	22 Nov 2024 08:56		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020401.D	22 Nov 2024 09:40		SY/MD	Ok,M
3	VY1122SBL01	VY1122SBL01	VY020402.D	22 Nov 2024 10:15		SY/MD	Ok
4	VY1122SBS01	VY1122SBS01	VY020403.D	22 Nov 2024 10:42	BS failed low	SY/MD	Not Ok
5	VY1122SBS02	VY1122SBS02	VY020404.D	22 Nov 2024 11:04		SY/MD	Ok,M
6	P4936-01RE	PL-01-11202024RE	VY020405.D	22 Nov 2024 11:38	vial-B ISTD Fail	SY/MD	Confirms
7	P4892-02	WB-310-BOT	VY020406.D	22 Nov 2024 12:01	vial-B BS failed low for comp. #3,28	SY/MD	Ok
8	P4954-03	TR-06-112124	VY020407.D	22 Nov 2024 12:53	vial-A BS failed low for comp. #3,28; Internal Standard Fail	SY/MD	ReRun
9	P4948-03	72-11944	VY020408.D	22 Nov 2024 13:16	vial-A	SY/MD	ReRun
10	P4948-01	337	VY020409.D	22 Nov 2024 13:39	vial-A	SY/MD	Ok
11	P4951-02	AU-05-112124	VY020410.D	22 Nov 2024 14:03	vial-A BS failed low for comp. #3,28; Internal Standard Fail	SY/MD	ReRun
12	P4954-01	TR-05-112124	VY020411.D	22 Nov 2024 14:26	vial-A Internal Standard Fail	SY/MD	Ok,M
13	P4946-02	BW-2	VY020412.D	22 Nov 2024 14:50	vial-A Internal Standard Fail	SY/MD	ReRun
14	P4946-01	BW-1	VY020413.D	22 Nov 2024 15:13	vial-A	SY/MD	Ok
15	P4960-01	B1	VY020414.D	22 Nov 2024 15:36	vial-A	SY/MD	Ok
16	P4960-02	B2	VY020415.D	22 Nov 2024 16:00	vial-A	SY/MD	Ok
17	P4960-03	SW1	VY020416.D	22 Nov 2024 16:23	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112224

Review By	Maresh Dadoda	Review On	11/25/2024 2:01:45 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/25/2024 2:05:55 PM
SubDirectory	VY112224	HP Acquire Method	HP Processing Method
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131732		
CCC	VP131733,VP131734		
Internal Standard/PEM	VP128297		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	P4960-04	SW2	VY020417.D	22 Nov 2024 16:47	vial-A NOT PURGE	SY/MD	Not Ok
19	P4960-05	SW3	VY020418.D	22 Nov 2024 17:10	vial-A NOT PURGE	SY/MD	Not Ok
20	P4960-06	SW4	VY020419.D	22 Nov 2024 17:33	vial-A NOT PURGE	SY/MD	Not Ok
21	VIBLK	VIBLK	VY020420.D	22 Nov 2024 17:57		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY020421.D	22 Nov 2024 18:20		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112524

Review By	Maresh Dadoda	Review On	11/26/2024 9:43:29 AM
Supervise By	Semsettin Yesilyurt	Supervise On	11/26/2024 9:46:12 AM
SubDirectory	VY112524	HP Acquire Method	MSVOA_Y HP Processing Method 82y10311924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131777
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131785,VP131786 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020422.D	25 Nov 2024 09:01		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020423.D	25 Nov 2024 09:46		SY/MD	Ok,M
3	VY1125SBL01	VY1125SBL01	VY020424.D	25 Nov 2024 10:17		SY/MD	Ok
4	P4949-03	CONTAINMENT-STON	VY020425.D	25 Nov 2024 13:29	vial-A not purged	SY/MD	Not Ok
5	P4948-01RE	337RE	VY020426.D	25 Nov 2024 13:52	vial-B not purge	SY/MD	Not Ok
6	P4948-03	72-11944	VY020427.D	25 Nov 2024 14:15	vial-B	SY/MD	Ok
7	P4949-03	CONTAINMENT-STON	VY020428.D	25 Nov 2024 14:39	vial-B	SY/MD	Ok
8	VIBLK	VIBLK	VY020429.D	25 Nov 2024 15:02	NOT PURGE	SY/MD	Not Ok
9	VY1125SBS01	VY1125SBS01	VY020430.D	25 Nov 2024 15:24	Recovery failed	SY/MD	Not Ok
10	VY1125SBS02	VY1125SBS02	VY020431.D	25 Nov 2024 15:47		SY/MD	Ok,M
11	P4951-02RE	AU-05-112124RE	VY020432.D	25 Nov 2024 16:10	vial-B Internal Standard Fail	SY/MD	Confirms
12	P4960-03	SW1	VY020433.D	25 Nov 2024 16:34	vial-B	SY/MD	Ok
13	P4960-04	SW2	VY020434.D	25 Nov 2024 16:57	vial-B Internal Standard Fail	SY/MD	Ok
14	P4960-05	SW3	VY020435.D	25 Nov 2024 17:20	vial-B NOT PURGE	SY/MD	Not Ok
15	P4960-06	SW4	VY020436.D	25 Nov 2024 17:44	vial-B Internal Standard Fail; Surrogate fail	SY/MD	Ok
16	P4971-01	CL-02-112224	VY020437.D	25 Nov 2024 18:07	vial-A Internal Standard Fail	SY/MD	Ok
17	P4954-03RE	TR-06-112124RE	VY020438.D	25 Nov 2024 18:31	vial-B Internal Standard Fail	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112524

Review By	Mahesh Dadoda	Review On	11/26/2024 9:43:29 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	11/26/2024 9:46:12 AM		
SubDirectory	VY112524	HP Acquire Method	MSVOA_Y	HP Processing Method	82y10311924s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131777			
CCC		VP131785,VP131786			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4954-01	TR-05-112124	VY020439.D	25 Nov 2024 18:54	vial-B NOT PURGE	SY/MD	Not Ok
19	P4946-02	BW-2	VY020440.D	25 Nov 2024 19:17	vial-B	SY/MD	Ok
20	VY1125SBS02	VY1125SBS02	VY020441.D	25 Nov 2024 19:40		SY/MD	Ok,M
21	VSTDCCC050	VSTDCCC050EC	VY020442.D	25 Nov 2024 20:03		SY/MD	Ok,M

M : Manual Integration