

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4960
Client: CHA Companies, Inc.
Analytical Method: SW8260D **Datafile :** VX044083.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1203MBS01	Methyl tert-butyl Ether	2000	2100	ug/Kg	105			77	129	
	Benzene	2000	1800	ug/Kg	90			84	121	
	Toluene	2000	1900	ug/Kg	95			83	122	
	Ethyl Benzene	2000	1900	ug/Kg	95			82	124	
	m/p-Xylenes	4000	3800	ug/Kg	95			83	124	
	o-Xylene	2000	1900	ug/Kg	95			83	123	
	Isopropylbenzene	2000	2000	ug/Kg	100			82	124	
	N-propylbenzene	2000	2000	ug/Kg	100			81	123	
	1,3,5-Trimethylbenzene	2000	2000	ug/Kg	100			82	124	
	tert-Butylbenzene	2000	2000	ug/Kg	100			81	125	
	1,2,4-Trimethylbenzene	2000	2000	ug/Kg	100			81	125	
	Sec-butylbenzene	2000	2000	ug/Kg	100			80	124	
	p-Isopropyltoluene	2000	2000	ug/Kg	100			81	125	
	n-Butylbenzene	2000	1900	ug/Kg	95			78	126	
	Naphthalene	2000	2000	ug/Kg	100			70	134	