

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4960

Client: CHA Companies, Inc.

Analytical Method: SW8260D

Datafile : VY020441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1125SBSD02	Methyl tert-butyl Ether	20	24.2	ug/Kg	121	11		77	129	20
	Benzene	20	25.9	ug/Kg	130	8	*	84	121	20
	Toluene	20	26.1	ug/Kg	131	5	*	83	122	20
	Ethyl Benzene	20	25.2	ug/Kg	126	3	*	82	124	20
	m/p-Xylenes	40	52.2	ug/Kg	131	4	*	83	124	20
	o-Xylene	20	25.6	ug/Kg	128	2	*	83	123	20
	Isopropylbenzene	20	24.3	ug/Kg	121	1		82	124	20
	N-propylbenzene	20	24.1	ug/Kg	121	1		81	123	20
	1,3,5-Trimethylbenzene	20	24.4	ug/Kg	122	2		82	124	20
	tert-Butylbenzene	20	24.9	ug/Kg	125	1		81	125	20
	1,2,4-Trimethylbenzene	20	24.5	ug/Kg	123	2		81	125	20
	Sec-butylbenzene	20	22.9	ug/Kg	115	4		80	124	20
	p-Isopropyltoluene	20	24.4	ug/Kg	122	1		81	125	20
	n-Butylbenzene	20	23.7	ug/Kg	119	4		78	126	20
	Naphthalene	20	24.1	ug/Kg	121	9		70	134	20