

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CLOU03
 Lab Code: CHEM Case No.: P4960 SAS No.: P4960 SDG NO.: P4960
 Lab File ID: VX044080.D Date Analyzed: 12/03/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	80742	12.018				
UPPER LIMIT	161484	12.518				
LOWER LIMIT	40371	11.518				
EPA SAMPLE NO.						
SW3	93386	12.02				
VX1203MBL01	85483	12.02				
VX1203MBS01	88269	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.