

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CLOU03
 Lab Code: CHEM Case No.: P4960 SAS No.: P4960 SDG NO.: P4960
 Lab File ID: VY020423.D Date Analyzed: 11/25/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	107880	13.359				
UPPER LIMIT	215760	13.859				
LOWER LIMIT	53940	12.859				
EPA SAMPLE NO.						
SW1	63892	13.35				
SW2	10419 *	13.35				
SW4	2505 *	13.35				
VY1125SBL01	84429	13.35				
VY1125SBS02	96610	13.35				
VY1125SBSD02	104045	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.