

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P4967 SAS No.: P4967 SDG No.: P4967

Instrument ID: MSVOA_N Calibration Date/Time: 11/25/2024 10:59

Lab File ID: VN085022.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.494	1.470	0.5	-1.61	
Toluene	1.709	1.640	0.4	-4.04	
Ethyl Benzene	1.796	1.679	0.1	-6.51	
m/p-Xylenes	0.687	0.664	0.3	-3.35	
o-Xylene	0.663	0.616	0.3	-7.09	
1,2-Dichloroethane-d4	2.412	2.453	0.01	1.7	
Toluene-d8	1.420	1.409	0.01	-0.77	
4-Bromofluorobenzene	0.517	0.534	0.2	3.29	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.