

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4967

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN085024.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN1125WBSD01	Benzene	20	19.9	ug/L	100	8		65	135	20
	Toluene	20	19.6	ug/L	98	7		70	130	20
	Ethyl Benzene	20	19.4	ug/L	97	7		60	140	20
	m/p-Xylenes	40	39.5	ug/L	99	8		87	111	20
	o-Xylene	20	20.3	ug/L	102	14		87	111	20