

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: P4967 SAS No.: P4967 SDG NO.: P4967
Lab File ID: VN085022.D Date Analyzed: 11/25/2024
Instrument ID: MSVOA_N Time Analyzed: 10:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	28386	7.81	151585	9.09	140690	11.87
UPPER LIMIT	56772	8.312	303170	9.594	281380	12.365
LOWER LIMIT	14193	7.312	75792.5	8.594	70345	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD	32568	7.81	164674	9.09	145379	11.87
002-35TH-AVE	23934	7.81	127741	9.10	114318	11.87
VN1125WBL01	24457	7.81	127294	9.09	111508	11.87
VN1125WBS01	29961	7.81	160469	9.09	146596	11.87
VN1125WBSD01	28214	7.81	145410	9.10	133637	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.