

Surrogate Summary

SW-846

SDG No.: P4969

Client: Parkway Generation Operating LLC

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4969-02	TOWER-1	2-Methylnaphthalene-d10	0.4	0.41	101		20	139
		Fluoranthene-d10	0.4	0.47	116		30	150
		Nitrobenzene-d5	0.4	0.39	97		27	123
		2-Fluorobiphenyl	0.4	0.34	84		34	132
		Terphenyl-d14	0.4	0.51	127		35	157
P4969-04	TOWER-2	2-Methylnaphthalene-d10	0.4	0.34	85		20	139
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.35	86		27	123
		2-Fluorobiphenyl	0.4	0.29	73		34	132
		Terphenyl-d14	0.4	0.48	120		35	157
PB165224BL	PB165224BL	2-Methylnaphthalene-d10	0.4	0.42	106		20	139
		Fluoranthene-d10	0.4	0.42	106		30	150
		Nitrobenzene-d5	0.4	0.38	95		27	123
		2-Fluorobiphenyl	0.4	0.40	99		34	132
		Terphenyl-d14	0.4	0.48	121		35	157
PB165224BS	PB165224BS	2-Methylnaphthalene-d10	0.4	0.65	162	*	20	139
		Fluoranthene-d10	0.4	0.46	116		30	150
		Nitrobenzene-d5	0.4	0.45	113		27	123
		2-Fluorobiphenyl	0.4	0.48	120		34	132
		Terphenyl-d14	0.4	0.51	127		35	157
PB165224BSD	PB165224BSD	2-Methylnaphthalene-d10	0.4	0.51	127		20	139
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.39	97		27	123
		2-Fluorobiphenyl	0.4	0.42	105		34	132
		Terphenyl-d14	0.4	0.46	115		35	157