

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4969 SAS No.: P4969 SDG NO.: P4969

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 11/27/2024

Lab File ID: BN035352.D Time Analyzed: 16:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2048	7.308	5229	10.05	3799	13.97
UPPER LIMIT	4096	7.808	10458	10.552	7598	14.467
LOWER LIMIT	1024	6.808	2614.5	9.552	1899.5	13.467
EPA SAMPLE NO.						
01 PB165224BL	1607	7.31	3806	10.06	2822	13.97
02 TOWER-1	2270	7.31	6204	10.05	5199	13.97
03 PB165224BS	1695	7.31	3992	10.05	2589	13.97
04 TOWER-2	2260	7.31	6676	10.05	5269	13.96
05 PB165224BSD	2108	7.31	4934	10.05	3223	13.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.