

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No. : SDG No.: ME29D7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133883 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV014				ID: CCV089				ID: CCV090		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	410000	102	0	400000	100	0
Calcium	10000	10000	102	2	400000	410000	104	0	410000	101	0
Iron	10000	11000	106	1	400000	420000	104	0	410000	102	0
Magnesium	6000	6000	101	1	400000	420000	105	0	410000	102	0
Potassium	9900	9400	95	4	150000	160000	104	0	150000	102	0
Sodium	10000	10000	101	2	400000	410000	102	0	400000	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No. : SDG No.: ME29D7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133883 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV091				ID: CCV092		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	1
Calcium					400000	400000	100	1	410000	102	1
Iron					400000	400000	101	1	410000	102	1
Magnesium					400000	400000	99	1	400000	101	1
Potassium					150000	150000	100	1	150000	102	1
Sodium					400000	390000	99	1	400000	99	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No. : SDG No.: ME29D7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133883 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV093				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	0			
Calcium					400000	400000	101	0			
Iron					400000	400000	100	0			
Magnesium					400000	400000	100	0			
Potassium					150000	150000	101	0			
Sodium					400000	400000	99	0			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No. : SDG No.: ME29D7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133949 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV007				ID: CCV008		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	95	0	400000	380000	94	0	380000	94	0
Calcium	10000	10000	102	1	400000	390000	98	0	390000	97	0
Iron	10000	11000	105	1	400000	390000	98	0	390000	97	0
Magnesium	6000	6200	104	0	400000	380000	95	0	380000	94	0
Potassium	9900	9500	96	2	150000	140000	96	0	140000	96	0
Sodium	10000	11000	106	3	400000	380000	95	0	380000	95	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No. : SDG No.: ME29D7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133949 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV009				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	94	0			
Calcium					400000	390000	97	1			
Iron					400000	390000	97	0			
Magnesium					400000	380000	94	0			
Potassium					150000	140000	96	0			
Sodium					400000	380000	94	1			