

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51759 MA No.: SDG No.: ME29F6
 Instrument ID: P7 Start Date : 12/06/2024
 Run Batch: LB133800 End Date : 12/06/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		165Ho	Q	209Bi	Q		Q		Q		Q
S0	1134	100		100							
S01	1137	100		101							
S02	1141	100		100							
S03	1144	100		100							
S04	1147	102		103							
S05	1150	97		98							
S06	1152	98		99							
S07	1155	99		98							
S08	1158	93		85							
ICV001	1243	97		99							
ICB001	1303	97		100							
ICSA001	1306	100		97							
ICSAB001	1309	106		101							
CCV001	1319	102		98							
CCB001	1322	102		105							
PBS402	1325	105		109							
LCS402	1336	102		103							
ZZZZZZ	1339										
ZZZZZZ	1342										
ME29F6	1346	107		107							
ME29F7	1349	108		107							
ME29F8	1352	113		111							
ME29F8D	1355	111		109							
ME29F8L	1358	107		109							
ME29F8S	1401	110		109							
ME29F9	1404	111		109							
ME29G0	1408	110		109							
ME29G1	1411	113		110							
ME29G2	1414	112		110							
CCV002	1417	104		101							
CCB002	1420	101		105							
ZZZZZZ	1424										
ZZZZZZ	1427										
ZZZZZZ	1430										
ZZZZZZ	1433										
ZZZZZZ	1436										
ZZZZZZ	1440										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Instrument ID: P7	Start Date : 12/06/2024
Run Batch: LB133800	End Date : 12/06/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	1443										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1514										
ZZZZZZ	1517										
ZZZZZZ	1521										
ZZZZZZ	1524										
ZZZZZZ	1527										
ZZZZZZ	1530										
CCV003	1533	103		100							
CCB003	1536	102		108							
ZZZZZZ	1646										
CCV004	1703	98		94							
CCB004	1706	95		100							
ME29F8A	1721	101		102							
ZZZZZZ	1725										
ZZZZZZ	1735										
CCV005	1747	99		95							
CCB005	1749	98		104							

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

[illegible]

