

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/11/2024
 Analytical Method: ICP-AES Run Batch: LB133902
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	-0	0	200	220	10	38750	38000	-2
Antimony	0	0.020	0	60	54	-10	1250	1100	-12
Arsenic	0	0.0	0	10	8.8	-12	1250	1300	4
Barium	0	-0	0	200	200	0	2500	2500	0
Beryllium	0	-0	0	5	5.4	8	125	130	4
Cadmium	0	0.0	0	5	4.7	-6	625	630	1
Calcium	0	-0.44	0	5000	5200	4	62500	65000	4
Chromium	0	-0	0	10	11	10	3875	4000	3
Cobalt	0	-0	0	50	51	2	625	640	2
Copper	0	-0	0	25	29	16	3750	3800	1
Iron	0	-0.040	0	100	120	20	37500	40000	7
Lead	0	0.0	0	10	7.6	-24	6250	6200	-1
Magnesium	0	-0.78	0	5000	5200	4	62500	66000	6
Manganese	0	-0	0	15	14	-7	3750	3900	4
Nickel	0	-0	0	40	40	0	625	630	1
Potassium	0	0.55	0	5000	4700	-6	37500	36000	-4
Selenium	0	0.010	0	35	25	-29	1250	1200	-4
Silver	0	0.0	0	10	9.7	-3	312.5	310	-1
Sodium	0	-0.29	0	5000	5400	8	37500	38000	1
Thallium	0	0.0	0	25	24	-4	1250	1300	4
Vanadium	0	-0	0	50	52	4	625	630	1
Zinc	0	0.010	0	60	58	-3	3750	3300	-12

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/11/2024
 Analytical Method: ICP-AES Run Batch: LB133902
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	75000	-3	155000	150000	-3	310000	310000	0
Antimony	2500	2300	-8	5000	4800	-4	10000	11000	10
Arsenic	2500	2600	4	5000	5100	2	10000	9800	-2
Barium	5000	5000	0	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	500	0	1000	990	-1
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Calcium	125000	130000	4	250000	250000	0	500000	500000	0
Chromium	7750	7900	2	15500	16000	3	31000	31000	0
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7500	0	15000	15000	0	30000	30000	0
Iron	75000	78000	4	150000	150000	0	300000	300000	0
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Magnesium	125000	130000	4	250000	260000	4	500000	510000	2
Manganese	7500	7600	1	15000	15000	0	30000	30000	0
Nickel	1250	1300	4	2500	2500	0	5000	5000	0
Potassium	75000	73000	-3	150000	150000	0	300000	300000	0
Selenium	2500	2300	-8	5000	4800	-4	10000	10000	0
Silver	625	610	-2	1250	1200	-4	2500	2500	0
Sodium	75000	74000	-1	150000	150000	0	300000	300000	0
Thallium	2500	2500	0	5000	5000	0	10000	9900	-1
Vanadium	1250	1200	-4	2500	2500	0	5000	5000	0
Zinc	7500	6700	-11	15000	14000	-7	30000	32000	7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHJ56
Instrument ID: P5 Start Date: 12/11/2024
Analytical Method: ICP-AES Run Batch: LB133902
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	810000	1						
Calcium	800000	800000	0						
Iron	800000	790000	-1						
Magnesium	800000	770000	-4						
Sodium	800000	800000	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/12/2024
 Analytical Method: ICP-AES Run Batch: LB133932
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	-0.010	0	200	220	10	38750	38000	-2
Antimony	0	0.010	0	60	53	-12	1250	1200	-4
Arsenic	0	0.0	0	10	8.3	-17	1250	1300	4
Barium	0	0.0	0	200	200	0	2500	2500	0
Beryllium	0	-0	0	5	5.2	4	125	130	4
Cadmium	0	0.0	0	5	4.6	-8	625	630	1
Calcium	0	-0.25	0	5000	5000	0	62500	64000	2
Chromium	0	-0	0	10	9.9	-1	3875	4100	6
Cobalt	0	-0	0	50	51	2	625	640	2
Copper	0	0.0	0	25	24	-4	3750	3800	1
Iron	0	0.0	0	100	84	-16	37500	39000	4
Lead	0	0.0	0	10	7.9	-21	6250	6200	-1
Magnesium	0	-0.64	0	5000	5000	0	62500	66000	6
Manganese	0	-0	0	15	16	7	3750	3800	1
Nickel	0	0.0	0	40	38	-5	625	630	1
Potassium	0	0.13	0	5000	5000	0	37500	37000	-1
Selenium	0	0.010	0	35	26	-26	1250	1200	-4
Silver	0	-0	0	10	11	10	312.5	310	-1
Sodium	0	0.29	0	5000	4800	-4	37500	37000	-1
Thallium	0	0.0	0	25	23	-8	1250	1300	4
Vanadium	0	-0	0	50	51	2	625	630	1
Zinc	0	0.010	0	60	62	3	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/12/2024
 Analytical Method: ICP-AES Run Batch: LB133932
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	76000	-2	155000	150000	-3	310000	310000	0
Antimony	2500	2300	-8	5000	4900	-2	10000	10000	0
Arsenic	2500	2600	4	5000	5200	4	10000	9700	-3
Barium	5000	5000	0	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	500	0	1000	990	-1
Cadmium	1250	1300	4	2500	2500	0	5000	4900	-2
Calcium	125000	130000	4	250000	250000	0	500000	490000	-2
Chromium	7750	8000	3	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	4900	-2
Copper	7500	7600	1	15000	15000	0	30000	30000	0
Iron	75000	78000	4	150000	160000	7	300000	300000	0
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Magnesium	125000	130000	4	250000	260000	4	500000	510000	2
Manganese	7500	7600	1	15000	15000	0	30000	30000	0
Nickel	1250	1300	4	2500	2500	0	5000	5000	0
Potassium	75000	74000	-1	150000	150000	0	300000	300000	0
Selenium	2500	2400	-4	5000	4900	-2	10000	10000	0
Silver	625	630	1	1250	1300	4	2500	2500	0
Sodium	75000	75000	0	150000	150000	0	300000	300000	0
Thallium	2500	2600	4	5000	5100	2	10000	9800	-2
Vanadium	1250	1300	4	2500	2500	0	5000	5000	0
Zinc	7500	6900	-8	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHJ56
Instrument ID: P5 Start Date: 12/12/2024
Analytical Method: ICP-AES Run Batch: LB133932
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	810000	1						
Calcium	800000	800000	0						
Iron	800000	790000	-1						
Magnesium	800000	770000	-4						
Sodium	800000	800000	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/13/2024
 Analytical Method: ICP-AES Run Batch: LB133949
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.040	0	200	180	-10	38750	37000	-5
Antimony	0	0.020	0	60	52	-13	1250	1100	-12
Arsenic	0	-0	0	10	13	30	1250	1300	4
Barium	0	0.0	0	200	200	0	2500	2400	-4
Beryllium	0	-0	0	5	5.2	4	125	130	4
Cadmium	0	0.0	0	5	4.8	-4	625	610	-2
Calcium	0	0.32	0	5000	5000	0	62500	61000	-2
Chromium	0	-0	0	10	9.9	-1	3875	4000	3
Cobalt	0	-0	0	50	51	2	625	630	1
Copper	0	-0	0	25	29	16	3750	3700	-1
Iron	0	0.010	0	100	88	-12	37500	38000	1
Lead	0	0.0	0	10	8.3	-17	6250	6000	-4
Magnesium	0	-0.37	0	5000	5300	6	62500	63000	1
Manganese	0	-0	0	15	17	13	3750	3800	1
Nickel	0	0.0	0	40	39	-3	625	610	-2
Potassium	0	0.30	0	5000	4900	-2	37500	36000	-4
Selenium	0	0.010	0	35	26	-26	1250	1100	-12
Silver	0	0.0	0	10	9.6	-4	312.5	300	-4
Sodium	0	-1.0	0	5000	6400	28	37500	37000	-1
Thallium	0	-0	0	25	26	4	1250	1200	-4
Vanadium	0	-0	0	50	52	4	625	610	-2
Zinc	0	0.020	0	60	50	-17	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/13/2024
 Analytical Method: ICP-AES Run Batch: LB133949
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	75000	-3	155000	150000	-3	310000	300000	-3
Antimony	2500	2300	-8	5000	4800	-4	10000	11000	10
Arsenic	2500	2600	4	5000	5100	2	10000	9800	-2
Barium	5000	5000	0	10000	10000	0	20000	20000	0
Beryllium	250	260	4	500	510	2	1000	990	-1
Cadmium	1250	1200	-4	2500	2500	0	5000	5000	0
Calcium	125000	120000	-4	250000	250000	0	500000	490000	-2
Chromium	7750	7900	2	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7600	1	15000	15000	0	30000	30000	0
Iron	75000	77000	3	150000	150000	0	300000	300000	0
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Magnesium	125000	130000	4	250000	250000	0	500000	490000	-2
Manganese	7500	7700	3	15000	15000	0	30000	30000	0
Nickel	1250	1200	-4	2500	2500	0	5000	5000	0
Potassium	75000	74000	-1	150000	150000	0	300000	300000	0
Selenium	2500	2300	-8	5000	4800	-4	10000	11000	10
Silver	625	620	-1	1250	1200	-4	2500	2500	0
Sodium	75000	74000	-1	150000	150000	0	300000	290000	-3
Thallium	2500	2500	0	5000	5000	0	10000	10000	0
Vanadium	1250	1200	-4	2500	2500	0	5000	5000	0
Zinc	7500	7000	-7	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHJ56
Instrument ID: P5 Start Date: 12/13/2024
Analytical Method: ICP-AES Run Batch: LB133949
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	820000	3						
Calcium	800000	820000	3						
Iron	800000	800000	0						
Magnesium	800000	800000	0						
Sodium	800000	810000	1						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/15/2024
 Analytical Method: ICP-AES Run Batch: LB133959
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.020	0	200	180	-10	38750	39000	1
Antimony	0	0.020	0	60	48	-20	1250	1100	-12
Arsenic	0	-0	0	10	12	20	1250	1300	4
Barium	0	0.010	0	200	200	0	2500	2500	0
Beryllium	0	-0	0	5	5.0	0	125	130	4
Cadmium	0	0.0	0	5	4.5	-10	625	610	-2
Calcium	0	-0.12	0	5000	5000	0	62500	63000	1
Chromium	0	-0	0	10	9.9	-1	3875	3900	1
Cobalt	0	0.0	0	50	48	-4	625	620	-1
Copper	0	0.0	0	25	24	-4	3750	3800	1
Iron	0	0.010	0	100	78	-22	37500	39000	4
Lead	0	0.0	0	10	10	0	6250	6100	-2
Magnesium	0	-0.57	0	5000	5200	4	62500	65000	4
Manganese	0	-0	0	15	17	13	3750	3800	1
Nickel	0	0.0	0	40	38	-5	625	610	-2
Potassium	0	0.20	0	5000	5000	0	37500	37000	-1
Selenium	0	0.010	0	35	26	-26	1250	1100	-12
Silver	0	0.0	0	10	9.8	-2	312.5	300	-4
Sodium	0	-0.22	0	5000	4900	-2	37500	39000	4
Thallium	0	0.0	0	25	23	-8	1250	1300	4
Vanadium	0	0.010	0	50	37	-26	625	630	1
Zinc	0	0.020	0	60	50	-17	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No.: _____ SDG No.: MBHJ56
 Instrument ID: P5 Start Date: 12/15/2024
 Analytical Method: ICP-AES Run Batch: LB133959
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	77000	-1	155000	150000	-3	310000	310000	0
Antimony	2500	2300	-8	5000	4800	-4	10000	11000	10
Arsenic	2500	2600	4	5000	5100	2	10000	9800	-2
Barium	5000	5000	0	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	500	0	1000	990	-1
Cadmium	1250	1200	-4	2500	2500	0	5000	5000	0
Calcium	125000	130000	4	250000	250000	0	500000	500000	0
Chromium	7750	7800	1	15500	16000	3	31000	31000	0
Cobalt	1250	1200	-4	2500	2500	0	5000	5000	0
Copper	7500	7500	0	15000	15000	0	30000	30000	0
Iron	75000	78000	4	150000	160000	7	300000	310000	3
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Magnesium	125000	130000	4	250000	250000	0	500000	500000	0
Manganese	7500	7600	1	15000	15000	0	30000	30000	0
Nickel	1250	1200	-4	2500	2500	0	5000	5000	0
Potassium	75000	74000	-1	150000	150000	0	300000	300000	0
Selenium	2500	2300	-8	5000	4800	-4	10000	10000	0
Silver	625	610	-2	1250	1200	-4	2500	2500	0
Sodium	75000	76000	1	150000	150000	0	300000	300000	0
Thallium	2500	2500	0	5000	5000	0	10000	9900	-1
Vanadium	1250	1300	4	2500	2500	0	5000	5000	0
Zinc	7500	7000	-7	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHJ56
Instrument ID: P5 Start Date: 12/15/2024
Analytical Method: ICP-AES Run Batch: LB133959
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	800000	0						
Calcium	800000	790000	-1						
Iron	800000	780000	-3						
Magnesium	800000	790000	-1						
Sodium	800000	800000	0						