

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133902

Review By	Kareem Khairalla	Review On	12/13/2024 7:33:52 PM
Supervise By	Sarabjit Jaswal	Supervise On	12/17/2024 10:28:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83513,MP83514		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/11/24 13:26		Kareem	OK
2	S1	S01	CAL2	12/11/24 13:33		Kareem	OK
3	S2	S02	CAL3	12/11/24 13:37		Kareem	OK
4	S3	S03	CAL4	12/11/24 13:41		Kareem	OK
5	S4	S04	CAL5	12/11/24 13:45		Kareem	OK
6	S5	S05	CAL6	12/11/24 13:48		Kareem	OK
7	S6	S06	CAL7	12/11/24 13:53		Kareem	OK
8	ICV015	ICV015	ICV	12/11/24 15:11	Inst. paused for calibration review	Kareem	OK
9	ICB015	ICB015	ICB	12/11/24 15:20	Inst. paused for review	Kareem	OK
10	ICSA015	ICSA015	ICSA	12/11/24 15:24		Kareem	OK
11	ICSAB015	ICSAB015	ICSAB	12/11/24 15:28		Kareem	OK
12	CCV097	CCV097	CCV	12/11/24 15:33		Kareem	OK
13	CCB097	CCB097	CCB	12/11/24 15:37		Kareem	OK
14	PB165217BL	PBS217	MB	12/11/24 15:41		Kareem	OK
15	PB165217BS	LCS217	LCS	12/11/24 15:45		Kareem	OK
16	PB165218BL	PBS218	MB	12/11/24 15:49		Kareem	OK
17	PB165218BS	LCS218	LCS	12/11/24 15:53		Kareem	OK
18	PB165415BL	PBS415	MB	12/11/24 15:57		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133902

Review By	Kareem Khairalla	Review On	12/13/2024 7:33:52 PM
Supervise By	Sarabjit Jaswal	Supervise On	12/17/2024 10:28:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83513,MP83514

19	PB165415BS	LCS415	LCS	12/11/24 16:01		Kareem	OK
20	PB165216BL	PBS216	MB	12/11/24 16:05		Kareem	OK
21	PB165216BS	LCS216	LCS	12/11/24 16:09		Kareem	OK
22	PB165281BL	PBS281	MB	12/11/24 16:13		Kareem	OK
23	PB165281BS	LCS281	LCS	12/11/24 16:18		Kareem	OK
24	PB165291BL	PBS291	MB	12/11/24 16:22		Kareem	OK
25	PB165291BS	LCS291	LCS	12/11/24 16:26		Kareem	OK
26	P5067-19	ME28N8	SAM	12/11/24 16:30		Kareem	OK
27	P5067-20	ME28N8D	DUP	12/11/24 16:34		Kareem	OK
28	P5067-19L	ME28N8L	SD	12/11/24 16:38		Kareem	OK
29	P5067-21	ME28N8S	MS	12/11/24 16:42		Kareem	OK
30	P4940-16	MBHHX8	SAM	12/11/24 16:46	Ca high	Kareem	Dilution
31	P4940-17	MBHHX8D	DUP	12/11/24 16:50	Ca high	Kareem	Dilution
32	P4940-16L	MBHHX8L	SD	12/11/24 16:53	Ca high	Kareem	Dilution
33	P4940-18	MBHHX8S	MS	12/11/24 16:57	MS fail for Zn (789.88 ppb), Sb,Se,Ag (Below RL)	Kareem	OK
34	CCV098	CCV098	CCV	12/11/24 17:01		Kareem	OK
35	CCB098	CCB098	CCB	12/11/24 17:05		Kareem	OK
36	P4963-15	MBHJ13	SAM	12/11/24 17:09	Ca high	Kareem	Dilution
37	P4963-16	MBHJ13D	DUP	12/11/24 17:13	Ca high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133902

Review By	Kareem Khairalla	Review On	12/13/2024 7:33:52 PM
Supervise By	Sarabjit Jaswal	Supervise On	12/17/2024 10:28:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83513,MP83514

38	P4963-15L	MBHJ13L	SD	12/11/24 17:17	Ca high	Kareem	Dilution
39	P4963-17	MBHJ13S	MS	12/11/24 17:28	MS fail for Sb,Se,Ag (Below RL)	Kareem	OK
40	P4964-01	MBHJ01	SAM	12/11/24 17:32	Ca high	Kareem	Dilution
41	P4964-02	MBHJ01D	DUP	12/11/24 17:36	Ca high	Kareem	Dilution
42	P4964-01L	MBHJ01L	SD	12/11/24 17:39	Ca high	Kareem	Dilution
43	P4964-03	MBHJ01S	MS	12/11/24 17:43	Ca high, MS fail for Zn (1644.27 ppb), Sb,Se,Ag (Below RL)	Kareem	Dilution
44	P4981-18	MBHJC9	SAM	12/11/24 17:47	Ca high	Kareem	Dilution
45	P4981-19	MBHJC9D	DUP	12/11/24 17:51	Ca high	Kareem	Dilution
46	P4981-18L	MBHJC9L	SD	12/11/24 17:55	Ca high	Kareem	Dilution
47	P4981-20	MBHJC9S	MS	12/11/24 17:59	MS fail for Zn (784.24 ppb), Sb,Se,Ag (Below RL)	Kareem	OK
48	P4983-10	MBHJG2	SAM	12/11/24 18:03		Kareem	OK
49	P4983-11	MBHJG2D	DUP	12/11/24 18:06		Kareem	OK
50	P4983-10L	MBHJG2L	SD	12/11/24 18:10		Kareem	OK
51	P4983-12	MBHJG2S	MS	12/11/24 18:14	MS fail for Sb,Se,Ag (Below RL)	Kareem	OK
52	P5067-01	ME28M0	SAM	12/11/24 18:18	Ca,Fe high	Kareem	Dilution
53	P5067-02	ME28M1	SAM	12/11/24 18:25	Ca,Fe high	Kareem	Dilution
54	P5067-03	ME28M2	SAM	12/11/24 18:29	Ca high	Kareem	Dilution
55	P5067-04	ME28M3	SAM	12/11/24 18:32	Ca high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133902

Review By	Kareem Khairalla	Review On	12/13/2024 7:33:52 PM
Supervise By	Sarabjit Jaswal	Supervise On	12/17/2024 10:28:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83513,MP83514

56	CCV099	CCV099	CCV	12/11/24 18:36		Kareem	OK
57	CCB099	CCB099	CCB	12/11/24 18:40		Kareem	OK