


Instrument I			Analyst:		NF	Method: .00.0 / 9056A					Analyst
Ident	Con F-	Con CL-	Con NO2	Con BR-	Con NO3	Con HPO4	Con SO4	Method name	date time	initial wt/Final Vc	
STD1	0	0	0	0	0	0	0	IC1-111824	11/18/2024 11:24	10	NF/IZ
STD2	0.402	0.571	0.601	1.961	0.489	1.025	3.062	IC1-111824	11/18/2024 11:46	10	NF/IZ
STD3	0.794	1.185	1.175	3.939	0.997	1.928	6.062	IC1-111824	11/18/2024 12:07	10	NF/IZ
STD4	1.023	1.554	1.535	5.128	1.284	2.665	7.843	IC1-111824	11/18/2024 12:29	10	NF/IZ
STD5	1.997	3.009	2.993	9.997	2.488	4.96	14.679	IC1-111824	11/18/2024 12:50	10	NF/IZ
STD6	3.936	5.952	5.978	19.928	4.964	9.645	28.993	IC1-111824	11/18/2024 13:11	10	NF/IZ
STD7	5.049	7.529	7.517	25.045	6.279	12.776	37.863	IC1-111824	11/18/2024 13:33	10	NF/IZ
ICV	1.966	3.006	2.997	9.98	2.49	4.979	14.594	IC1-111824	11/18/2024 13:54	10	NF/IZ
ICB	0	0	0	0	0	0	0	IC1-111824	11/18/2024 14:16	10	NF/IZ
CCV	1.926	2.971	2.974	10.004	2.473	4.986	14.286	IC1-111824	11/25/2024 11:01	10	NF/IZ
CCB	0	0	0	0	0	0	0	IC1-111824	11/25/2024 11:22	10	NF/IZ
LB133607BLW	0	0	0	0	0	0	0	IC1-111824	11/25/2024 11:44	5.00g/100ML	NF/IZ
LB133607BSW	1.947	2.956	2.995	9.982	2.469	5.119	14.269	IC1-111824	11/25/2024 12:05	5.00g/100ML	NF/IZ
P4986-01	0.047	0.096	0	0	0	0	0.932	IC1-111824	11/25/2024 13:11	5.01g/100ML	NF/IZ
P4986-02	0.07	0.041	0	0	0	0	0.648	IC1-111824	11/25/2024 13:33	5.02g/100ML	NF/IZ
P4986-02MS	1.876	3.17	3.05	10.136	2.524	4.443	14.867	IC1-111824	11/25/2024 13:55	5.02g/100ML	NF/IZ
P4986-02MSD	1.864	3.163	3.057	10.149	2.524	4.605	15.021	IC1-111824	11/25/2024 14:16	5.02g/100ML	NF/IZ
CCV	1.979	2.983	3.002	10.045	2.487	5.108	14.335	IC1-111824	11/25/2024 14:38	10	NF/IZ
CCB	0	0	0	0	0	0	0	IC1-111824	11/25/2024 14:59	10	NF/IZ

Clear table

Instrument ID: IC-2												Analyst :		IZ		Method:		300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final	Analyst								
STD1	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-111824	11/18/2024 11:24	10	NF/IZ								
STD2	0.402	0.571	0.601	1.961	0.489	1.025	3.062	IC1-111824	11/18/2024 11:46	10	NF/IZ								
STD3	0.794	1.185	1.175	3.939	0.997	1.928	6.062	IC1-111824	11/18/2024 12:07	10	NF/IZ								
STD4	1.023	1.554	1.535	5.128	1.284	2.665	7.843	IC1-111824	11/18/2024 12:29	10	NF/IZ								
STD5	1.997	3.009	2.993	9.997	2.488	4.960	14.679	IC1-111824	11/18/2024 12:50	10	NF/IZ								
STD6	3.936	5.952	5.978	19.928	4.964	9.645	28.993	IC1-111824	11/18/2024 13:11	10	NF/IZ								
STD7	5.049	7.529	7.517	25.045	6.279	12.776	37.863	IC1-111824	11/18/2024 13:33	10	NF/IZ								
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4												
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000												
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000												
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000												
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000												
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000												
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000												
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000												
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4												
STD1																			
STD2	0.5000	-4.8333	0.1667	-1.9500	-2.2000	2.5000	2.0667												
STD3	-0.7500	-1.2500	-2.0833	-1.5250	-0.3000	-3.6000	1.0333												
STD4	2.3000	3.6000	2.3333	2.5600	2.7200	6.6000	4.5733												
STD5	-0.1500	0.3000	-0.2333	-0.0300	-0.4800	-0.8000	-2.1400												
STD6	-1.6000	-0.8000	-0.3667	-0.3600	-0.7200	-3.5500	-3.3567												
STD7	0.9800	0.3867	0.2267	0.1800	0.4640	2.2080	2.3324												

Fluoride (Anions)

($\mu\text{S}/\text{cm}$) x min



Function: $A = -1.48773E-3 + 0.0145652 \times Q$


Relative standard deviation: 1.923523 %

Correlation coefficient: 0.999800

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	0.400	10.0	1.0	1.0	0.057	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	0.800	10.0	1.0	1.0	0.114	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	1.000	10.0	1.0	1.0	0.147	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	2.000	10.0	1.0	1.0	0.289	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	4.000	10.0	1.0	1.0	0.572	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	5.000	10.0	1.0	1.0	0.734	STD7	2024-11-18 13:33:24 UTC-5	used

Chloride (Anions)

($\mu\text{S}/\text{cm}$) x min



Function: $A = 4.85841E-3 + 9.50243E-3 \times Q$


Relative standard deviation 1.265280 %

Correlation coefficient 0.999910

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	0.600	10.0	1.0	1.0	0.059	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	1.200	10.0	1.0	1.0	0.117	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	1.500	10.0	1.0	1.0	0.153	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	3.000	10.0	1.0	1.0	0.291	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	6.000	10.0	1.0	1.0	0.570	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	7.500	10.0	1.0	1.0	0.720	STD7	2024-11-18 13:33:24 UTC-5	used

Nitrite (Anions)

(μS/cm) x min

Function: $A = -5.27414E-3 + 0.0216316 \times Q$


Relative standard deviation 0.783603 %

Correlation coefficient 0.999967

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	0.600	10.0	1.0	1.0	0.125	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	1.200	10.0	1.0	1.0	0.249	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	1.500	10.0	1.0	1.0	0.327	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	3.000	10.0	1.0	1.0	0.642	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	6.000	10.0	1.0	1.0	1.288	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	7.500	10.0	1.0	1.0	1.621	STD7	2024-11-18 13:33:24 UTC-5	used

Bromide (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = 7.64331E-4 + 4.24604E-3 \times Q$


Relative standard deviation: 0.771268 %

Correlation coefficient: 0.999967

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	2.000	10.0	1.0	1.0	0.084	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	4.000	10.0	1.0	1.0	0.168	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	5.000	10.0	1.0	1.0	0.219	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	10.000	10.0	1.0	1.0	0.425	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	20.000	10.0	1.0	1.0	0.847	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	25.000	10.0	1.0	1.0	1.064	STD7	2024-11-18 13:33:24 UTC-5	used

Nitrate (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min




Function: $A = 1.14484E-3 + 0.0238833 \times Q$

Relative standard deviation: 1.085494 %

Correlation coefficient 0.999936

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	0.500	10.0	1.0	1.0	0.118	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	1.000	10.0	1.0	1.0	0.239	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	1.250	10.0	1.0	1.0	0.308	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	2.500	10.0	1.0	1.0	0.595	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	5.000	10.0	1.0	1.0	1.187	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	6.250	10.0	1.0	1.0	1.501	STD7	2024-11-18 13:33:24 UTC-5	used


Phosphate (Anions) $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -2.33576E-3 + 8.30878E-3 \times Q$

Relative standard deviation 4.444706 %

Correlation coefficient 0.998936

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	1.000	10.0	1.0	1.0	0.083	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	2.000	10.0	1.0	1.0	0.158	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	2.500	10.0	1.0	1.0	0.219	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	5.000	10.0	1.0	1.0	0.410	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	10.000	10.0	1.0	1.0	0.799	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	12.500	10.0	1.0	1.0	1.059	STD7	2024-11-18 13:33:24 UTC-5	used

Sulfate (Anions)

 $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -9.71222\text{E-}3 + 7.64397\text{E-}3 \times Q$

Relative standard deviation 4.327283 %

Correlation coefficient 0.998986

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-11-18 11:24:52 UTC-5	used
Standard 2	1	3.000	10.0	1.0	1.0	0.224	STD2	2024-11-18 11:46:16 UTC-5	used
Standard 3	1	6.000	10.0	1.0	1.0	0.454	STD3	2024-11-18 12:07:40 UTC-5	used
Standard 4	1	7.500	10.0	1.0	1.0	0.590	STD4	2024-11-18 12:29:05 UTC-5	used
Standard 5	1	15.000	10.0	1.0	1.0	1.112	STD5	2024-11-18 12:50:30 UTC-5	used
Standard 6	1	30.000	10.0	1.0	1.0	2.206	STD6	2024-11-18 13:11:57 UTC-5	used
Standard 7	1	37.000	10.0	1.0	1.0	2.884	STD7	2024-11-18 13:33:24 UTC-5	used


Sample data

Ident STD1
Sample type Standard 1
Determination start 2024-11-18 11:24:52 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.60 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident STD2
Sample type Standard 2
Determination start 2024-11-18 11:46:16 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.54 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.372	0.0570	0.347	0.402	Fluoride
2	6.477	0.0592	0.232	0.571	Chloride
3	8.515	0.1247	0.380	0.601	Nitrite
4	10.463	0.0840	0.209	1.961	Bromide
5	12.127	0.1179	0.258	0.489	Nitrate
6	14.418	0.0829	0.139	1.025	Phosphate
7	16.130	0.2243	0.381	3.062	Sulfate


Sample data

Ident STD3
Sample type Standard 3
Determination start 2024-11-18 12:07:40 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.37 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.378	0.1141	0.690	0.794	Fluoride
2	6.478	0.1175	0.465	1.185	Chloride
3	8.515	0.2490	0.763	1.175	Nitrite
4	10.453	0.1680	0.422	3.939	Bromide
5	12.117	0.2394	0.526	0.997	Nitrate
6	14.440	0.1579	0.257	1.928	Phosphate
7	16.173	0.4537	0.772	6.062	Sulfate


Sample data

Ident STD4
Sample type Standard 4
Determination start 2024-11-18 12:29:05 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.32 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.375	0.1475	0.909	1.023	Fluoride
2	6.465	0.1525	0.612	1.554	Chloride
3	8.493	0.3268	1.014	1.535	Nitrite
4	10.420	0.2185	0.557	5.128	Bromide
5	12.077	0.3078	0.689	1.284	Nitrate
6	14.427	0.2191	0.375	2.665	Phosphate
7	16.192	0.5898	1.014	7.843	Sulfate


Sample data

Ident STD5
Sample type Standard 5
Determination start 2024-11-18 12:50:30 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.26 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.378	0.2894	1.819	1.997	Fluoride
2	6.467	0.2908	1.180	3.009	Chloride
3	8.495	0.6422	2.018	2.993	Nitrite
4	10.422	0.4253	1.098	9.997	Bromide
5	12.082	0.5953	1.348	2.488	Nitrate
6	14.432	0.4098	0.722	4.960	Phosphate
7	16.203	1.1123	1.944	14.679	Sulfate


Sample data

Ident STD6
Sample type Standard 6
Determination start 2024-11-18 13:11:57 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.37 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.383	0.5717	3.596	3.936	Fluoride
2	6.488	0.5704	2.311	5.952	Chloride
3	8.533	1.2879	4.031	5.978	Nitrite
4	10.483	0.8469	2.176	19.928	Bromide
5	12.148	1.1866	2.673	4.964	Nitrate
6	14.438	0.7991	1.399	9.645	Phosphate
7	16.165	2.2065	3.859	28.993	Sulfate


Sample data

Ident STD7
Sample type Standard 7
Determination start 2024-11-18 13:33:24 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.54 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.383	0.7339	4.626	5.049	Fluoride
2	6.510	0.7203	2.901	7.529	Chloride
3	8.573	1.6208	5.052	7.517	Nitrite
4	10.542	1.0642	2.717	25.045	Bromide
5	12.213	1.5007	3.361	6.279	Nitrate
6	14.452	1.0592	1.891	12.776	Phosphate
7	16.138	2.8845	5.052	37.863	Sulfate


Sample data

Ident ICV
Sample type Check standard 1
Determination start 2024-11-18 13:54:44 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.66 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.375	0.2848	1.764	1.966	Fluoride
2	6.508	0.2905	1.150	3.006	Chloride
3	8.577	0.6430	1.968	2.997	Nitrite
4	10.557	0.4245	1.063	9.980	Bromide
5	12.230	0.5959	1.307	2.490	Nitrate
6	14.447	0.4114	0.721	4.979	Phosphate
7	16.107	1.1058	1.909	14.594	Sulfate


Sample data

Ident ICB
Sample type Sample
Determination start 2024-11-18 14:16:12 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.66 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident CCV
Sample type Check standard 1
Determination start 2024-11-25 11:01:03 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.373	0.2791	1.566	1.926	Fluoride
2	6.523	0.2872	1.046	2.971	Chloride
3	8.610	0.6380	1.801	2.974	Nitrite
4	10.610	0.4255	1.002	10.004	Bromide
5	12.283	0.5918	1.215	2.473	Nitrate
6	14.208	0.4120	0.666	4.986	Phosphate
7	16.038	1.0823	1.693	14.286	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2024-11-25 11:22:33 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident LB133607BLW
Sample type Sample
Determination start 2024-11-25 11:44:03 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.77 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident LB133607BSW
Sample type Check standard 1
Determination start 2024-11-25 12:05:33 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.71 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.373	0.2822	1.582	1.947	Fluoride
2	6.488	0.2858	1.044	2.956	Chloride
3	8.548	0.6425	1.816	2.995	Nitrite
4	10.510	0.4246	1.006	9.982	Bromide
5	12.172	0.5908	1.217	2.469	Nitrate
6	14.195	0.4230	0.698	5.119	Phosphate
7	16.087	1.0810	1.680	14.269	Sulfate


Sample data

Ident P4986-01
Sample type Sample
Determination start 2024-11-25 13:11:53 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.77 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	3.380	0.0053	0.028	0.047	Fluoride
2	6.512	0.0140	0.051	0.096	Chloride
3	16.093	0.0615	0.096	0.932	Sulfate


Sample data

Ident P4986-02
Sample type Sample
Determination start 2024-11-25 13:33:29 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.77 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.372	0.0088	0.048	0.070	Fluoride
2	6.498	0.0088	0.031	0.041	Chloride
3	16.082	0.0398	0.062	0.648	Sulfate


Sample data

Ident P4986-02MS
Sample type Sample
Determination start 2024-11-25 13:55:02 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.66 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.370	0.2717	1.558	1.876	Fluoride
2	6.482	0.3061	1.121	3.170	Chloride
3	8.540	0.6545	1.855	3.050	Nitrite
4	10.495	0.4312	1.018	10.136	Bromide
5	12.155	0.6040	1.241	2.524	Nitrate
6	14.195	0.3668	0.612	4.443	Phosphate
7	16.095	1.1267	1.768	14.867	Sulfate


Sample data

Ident P4986-02MSD
Sample type Sample
Determination start 2024-11-25 14:16:36 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.54 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.377	0.2700	1.531	1.864	Fluoride
2	6.482	0.3054	1.115	3.163	Chloride
3	8.530	0.6560	1.852	3.057	Nitrite
4	10.475	0.4317	1.019	10.149	Bromide
5	12.132	0.6038	1.240	2.524	Nitrate
6	14.205	0.3803	0.629	4.605	Phosphate
7	16.128	1.1385	1.766	15.021	Sulfate


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2024-11-25 14:38:10 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.54 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.373	0.2867	1.651	1.979	Fluoride
2	6.467	0.2883	1.060	2.983	Chloride
3	8.507	0.6440	1.834	3.002	Nitrite
4	10.437	0.4273	1.011	10.045	Bromide
5	12.090	0.5951	1.226	2.487	Nitrate
6	14.198	0.4221	0.699	5.108	Phosphate
7	16.142	1.0860	1.693	14.335	Sulfate


Sample data

Ident CCB
Sample type Sample
Determination start 2024-11-25 14:59:40 UTC-5
Method IC1-111824
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.43 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



WORKLIST(Hardcopy Internal Chain)

LB133607

WorkList Name : ANIONS S-11252024

WorkList ID : 185759

Department : Wet-Chemistry

Date : 11-25-2024 12:18:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4986-01	5.01g TAPIAL-SB04D-9-1122224-00-T1	Solid	Anions Group1	Cool 4 deg C	WEST04	L51	11/22/2024	9056A
P4986-02	5.02g TAPIAL1-SB04D-4-1122224-00-T1	Solid	Anions Group1	Cool 4 deg C	WEST04	L51	11/22/2024	9056A

P4986-02 MS 5.02g

P4986-02 MS 5.02g

BL 5.00g

BS 5.00g

Final volume 100ml with DI water

Balance ID: WC SC-7

NF

11.26.2024

Date/Time 11-25-2024 12:30
Raw Sample Received by: NFWC
Raw Sample Relinquished by: NFWC

Date/Time 11-25-2024 15:00
Raw Sample Received by: NFWC
Raw Sample Relinquished by: NFWC