

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133628

Slope : 97.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/26/2024	09:25
2	CAL2	1	Water	NA	NA	20.3	7.01	11/26/2024	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	11/26/2024	09:28
4	ICV	1	Water	NA	NA	20.3	7.00	11/26/2024	09:30
5	CCV1	1	Water	NA	NA	20.3	2.01	11/26/2024	09:33
6	P4986-01	1	Solid	20.02	20	23.7	6.56	11/26/2024	09:40
7	P4986-01DUP	1	Solid	20.03	20	23.8	6.57	11/26/2024	09:41
8	P4986-02	1	Solid	20.03	20	23.7	6.12	11/26/2024	09:45
9	CCV2	1	Water	NA	NA	20.3	12.02	11/26/2024	09:50

WORKLIST(Hardcopy Internal Chain)

WB 133628

WorkList Name : ph s p4986

WorkList ID : 185784

Department : Wet-Chemistry

Date : 11-26-2024 08:38:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4986-01	<i>A</i> TAPIAL 1-SB04D-9-112224-00-T	Solid	pH	Cool 4 deg C	WEST04	L51	11/22/2024	9045D
P4986-02	<i>A</i> TAPIAL 1-SB04D-4-112224-00-T	Solid	pH	Cool 4 deg C	WEST04	L51	11/22/2024	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11/26/24 08:45

to WCC

CSB

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11/26/24

CSB

CSB

12:00