

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	P4986
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB133607

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133607BSS							
Bromide	mg/Kg	200	200		100	1	90-110	11/25/2024
Chloride	mg/Kg	60	59.1		98	1	90-110	11/25/2024
Fluoride	mg/Kg	40	38.9		97	1	90-110	11/25/2024
Nitrite	mg/Kg	60	59.9		100	1	90-110	11/25/2024
Nitrate	mg/Kg	50	49.4		99	1	90-110	11/25/2024
Sulfate	mg/Kg	300	285		95	1	90-110	11/25/2024
Orthophosphate as P	mg/Kg	100	102		102	1	90-110	11/25/2024

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: P4986

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB133629

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133629BSS							
TOC	mg/Kg	1000	952		95	1	90-110	11/26/2024