

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions **SDG No.:** P4986
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P4986 **SAS No.:** P4986
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV104	Mercury	4.15	4.0	104	90 - 110	CV	11/27/2024	12:27	LB133657

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions **SDG No.:** P4986
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P4986 **SAS No.:** P4986
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV59	Mercury	4.88	5.0	98	90 - 110	CV	11/27/2024	12:31	LB133657
CCV60	Mercury	4.92	5.0	98	90 - 110	CV	11/27/2024	13:04	LB133657
CCV61	Mercury	5.04	5.0	101	90 - 110	CV	11/27/2024	13:27	LB133657