



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** Weston Solutions **SDG No.:** P4986  
**Contract:** WEST04 **Lab Code:** CHEM **Case No.:** P4986 **SAS No.:** P4986  
**Initial Calibration Source:** \_\_\_\_\_  
**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|---------------------------|----|------------------|------------------|---------------|
| CRA       | Mercury | 0.19           | 0.2                | 95            | 40 - 160                  | CV | 11/27/2024       | 12:39            | LB133657      |