



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4995

Project: West Lake

RunNo.: LB133616

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/25/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	11/25/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	11/25/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4995

Project: West Lake

RunNo.: LB133627

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	11/26/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/26/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/26/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4995

Project: West Lake

RunNo.: LB133685

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	12/02/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/02/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/02/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	12/02/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4995

Project: West Lake

RunNo.: LB133701

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	81.7	81	101	78-84	12/03/2024