

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	P4995
Project:	West Lake	Sample ID:	P4947-01
Client ID:	A3988DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/L	+/-20	0.48	J	0.48	J	1	0		11/26/2024

Duplicate Sample Summary

Client: ENTACT Project: West Lake Client ID: 001DUP	SDG No.: P4995 Sample ID: P4995-02 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		11/25/2024
pH	pH	+/-20	7.65		7.66		1	0.13		11/26/2024
Reactive Cyanide	mg/L	+/-20	0.00099	U	0.00099	U	1	0		12/02/2024
Flash Point	o F	+/-2	>212.0		>212.0		1	0		12/03/2024

Duplicate Sample Summary

Client: ENTACT Project: West Lake Client ID: 001MSD	SDG No.: P4995 Sample ID: P4995-02 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.98		0.98		2	0.41		11/25/2024