

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4995

Project: West Lake

Run No.: LB133616

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133616BS							
Hexavalent Chromium	mg/L	0.5	0.50		101	1	90-111	11/25/2024