



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P4995
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4995 **SAS No.:** P4995
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	96	40 - 160	CV	11/27/2024	16:11	LB133664
CRI01	Arsenic	18.2	20.0	91	40 - 160	P	12/02/2024	11:34	LB133689
	Barium	102	100	102	40 - 160	P	12/02/2024	11:34	LB133689
	Cadmium	5.93	6.0	99	40 - 160	P	12/02/2024	11:34	LB133689
	Chromium	10.2	10.0	102	40 - 160	P	12/02/2024	11:34	LB133689
	Lead	12.4	12.0	104	40 - 160	P	12/02/2024	11:34	LB133689
	Selenium	19.0	20.0	95	40 - 160	P	12/02/2024	11:34	LB133689
	Silver	10.3	10.0	103	40 - 160	P	12/02/2024	11:34	LB133689