



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	P4995				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	P4995	SAS No.:	P4995		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB15	Mercury	0.20	+/-0.20	U	0.20	CV	11/27/2024	16:01	LB133664	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: P4995

Contract: ENTA05

Lab Code: CHEM

Case No.: P4995

SAS No.: P4995

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	11/27/2024	16:06	LB133664
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	11/27/2024	16:38	LB133664
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	11/27/2024	17:05	LB133664
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	11/27/2024	17:26	LB133664

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: P4995

Contract: ENTA05

Lab Code: CHEM

Case No.: P4995

SAS No.: P4995

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	11:29	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	11:29	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	11:29	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	11:29	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	11:29	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	11:29	LB133689
	Silver	10.0	+/-10.0	U	10.0	P	12/02/2024	11:29	LB133689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: P4995							
Contract: ENTA05	Lab Code: CHEM	Case No.: P4995	SAS No.: P4995						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	11:51	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	11:51	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	11:51	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	11:51	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	11:51	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	11:51	LB133689
	Silver	10.0	+/-10.0	U	10.0	P	12/02/2024	11:51	LB133689
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	12:50	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	12:50	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	12:50	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	12:50	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	12:50	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	12:50	LB133689
	Silver	10.0	+/-10.0	U	10.0	P	12/02/2024	12:50	LB133689
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	13:46	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	13:46	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	13:46	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	13:46	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	13:46	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	13:46	LB133689
	Silver	10.0	+/-10.0	U	10.0	P	12/02/2024	13:46	LB133689
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	14:37	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	14:37	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	14:37	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	14:37	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	14:37	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	14:37	LB133689
	Silver	10.0	+/-10.0	U	10.0	P	12/02/2024	14:37	LB133689
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	15:29	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	15:29	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	15:29	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	15:29	LB133689
	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	15:29	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	15:29	LB133689
	Silver	1.61	+/-10.0	J	10.0	P	12/02/2024	15:29	LB133689
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	12/02/2024	16:15	LB133689
	Barium	100	+/-100	U	100	P	12/02/2024	16:15	LB133689
	Cadmium	6.00	+/-6.00	U	6.00	P	12/02/2024	16:15	LB133689
	Chromium	10.0	+/-10.0	U	10.0	P	12/02/2024	16:15	LB133689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	P4995			
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	P4995	SAS No.:	P4995	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	12/02/2024	16:15	LB133689
	Selenium	20.0	+/-20.0	U	20.0	P	12/02/2024	16:15	LB133689
	Silver	1.50	+/-10.0	J	10.0	P	12/02/2024	16:15	LB133689