

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	ENTACT	SDG No.:	P4995
Contract:	ENTA05	Lab Code:	CHEM
ICS Source:	EPA	Case No.:	P4995
		SAS No.:	P4995
		Instrument ID:	P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.50			-20	20	12/02/2024	11:38	LB133689
	Barium	2.81	6.0	47	-94	106	12/02/2024	11:38	LB133689
	Cadmium	6.75	1.0	675	-5	7	12/02/2024	11:38	LB133689
	Chromium	57.4	52.0	110	42	62	12/02/2024	11:38	LB133689
	Lead	7.95			-12	12	12/02/2024	11:38	LB133689
	Selenium	-19.8			-20	20	12/02/2024	11:38	LB133689
	Silver	-3.00			-10	10	12/02/2024	11:38	LB133689
ICSAB01	Arsenic	106	104	102	88.4	120	12/02/2024	11:42	LB133689
	Barium	515	537	96	437	637	12/02/2024	11:42	LB133689
	Cadmium	1010	972	104	826	1120	12/02/2024	11:42	LB133689
	Chromium	560	542	103	460	624	12/02/2024	11:42	LB133689
	Lead	55.3	49.0	113	37	61	12/02/2024	11:42	LB133689
	Selenium	31.5	46.0	68	26	66	12/02/2024	11:42	LB133689
	Silver	195	201	97	170	232	12/02/2024	11:42	LB133689