

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                 |                                                          |
|-------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>CHEMTECH</u>                       | Contract: <u>ENTA05</u>                                  |
| Lab Code: <u>CHEM</u> Case No.: <u>P4995</u>    | SAS No.: <u>P4995</u> SDG No.: <u>P4995</u>              |
| Instrument ID: <u>MSVOA_X</u>                   | Calibration Date(s): <u>11/21/2024</u> <u>11/21/2024</u> |
| Heated Purge: (Y/N) <u>N</u>                    | Calibration Time(s): <u>09:47</u> <u>12:03</u>           |
| GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm) |                                                          |

|                                |        |                     |        |                     |        |                     |       |       |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                   |        | RRF001 = VX043925.D |        | RRF005 = VX043926.D |        | RRF020 = VX043927.D |       |       |
|                                |        | RRF050 = VX043928.D |        | RRF100 = VX043929.D |        | RRF150 = VX043930.D |       |       |
| COMPOUND                       | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.634  | 0.633               | 0.641  | 0.684               | 0.601  | 0.608               | 0.633 | 4.6   |
| Chloromethane                  | 0.578  | 0.662               | 0.720  | 0.697               | 0.665  | 0.677               | 0.667 | 7.3   |
| Vinyl Chloride                 | 0.675  | 0.708               | 0.730  | 0.762               | 0.695  | 0.684               | 0.709 | 4.6   |
| Bromomethane                   |        | 0.425               | 0.438  | 0.445               | 0.429  | 0.442               | 0.436 | 2     |
| Chloroethane                   | 0.579  | 0.429               | 0.404  | 0.459               | 0.368  | 0.354               | 0.432 | 18.9  |
| Trichlorofluoromethane         | 1.038  | 1.267               | 1.335  | 1.357               | 1.281  | 1.328               | 1.268 | 9.3   |
| 1,1,2-Trichlorotrifluoroethane | 0.592  | 0.585               | 0.605  | 0.658               | 0.589  | 0.599               | 0.605 | 4.5   |
| 1,1-Dichloroethene             | 0.596  | 0.530               | 0.585  | 0.600               | 0.572  | 0.576               | 0.576 | 4.4   |
| Acetone                        | 0.401  | 0.396               | 0.401  | 0.425               | 0.370  | 0.370               | 0.394 | 5.4   |
| Carbon Disulfide               | 1.082  | 0.956               | 1.078  | 1.287               | 1.330  | 1.398               | 1.188 | 14.6  |
| Methyl tert-butyl Ether        | 1.732  | 1.991               | 2.212  | 2.264               | 2.160  | 2.192               | 2.092 | 9.5   |
| Methyl Acetate                 | 0.931  | 0.871               | 0.911  | 0.987               | 0.894  | 0.910               | 0.917 | 4.3   |
| Methylene Chloride             | 0.704  | 0.656               | 0.684  | 0.667               | 0.643  | 0.655               | 0.668 | 3.3   |
| trans-1,2-Dichloroethene       | 0.545  | 0.563               | 0.609  | 0.633               | 0.600  | 0.619               | 0.595 | 5.7   |
| 1,1-Dichloroethane             | 0.980  | 1.118               | 1.221  | 1.256               | 1.185  | 1.208               | 1.161 | 8.6   |
| Cyclohexane                    |        | 0.911               | 0.976  | 1.041               | 0.926  | 0.925               | 0.956 | 5.6   |
| 2-Butanone                     | 0.448  | 0.486               | 0.538  | 0.567               | 0.500  | 0.510               | 0.508 | 8.1   |
| Carbon Tetrachloride           | 0.439  | 0.466               | 0.491  | 0.558               | 0.520  | 0.528               | 0.500 | 8.7   |
| cis-1,2-Dichloroethene         | 0.674  | 0.673               | 0.761  | 0.792               | 0.745  | 0.767               | 0.735 | 6.8   |
| Bromochloromethane             | 0.501  | 0.611               | 0.482  | 0.450               | 0.500  | 0.523               | 0.511 | 10.7  |
| Chloroform                     | 1.070  | 1.238               | 1.309  | 1.316               | 1.269  | 1.293               | 1.249 | 7.4   |
| 1,1,1-Trichloroethane          | 0.873  | 1.005               | 1.133  | 1.184               | 1.119  | 1.149               | 1.077 | 10.9  |
| Methylcyclohexane              | 0.544  | 0.556               | 0.564  | 0.656               | 0.574  | 0.576               | 0.578 | 6.9   |
| Benzene                        | 1.211  | 1.369               | 1.457  | 1.494               | 1.388  | 1.402               | 1.387 | 7     |
| 1,2-Dichloroethane             | 0.499  | 0.546               | 0.585  | 0.603               | 0.551  | 0.558               | 0.557 | 6.4   |
| Trichloroethene                | 0.547  | 0.372               | 0.368  | 0.373               | 0.345  | 0.352               | 0.393 | 19.5  |
| 1,2-Dichloropropane            | 0.288  | 0.322               | 0.361  | 0.371               | 0.346  | 0.350               | 0.340 | 8.8   |
| Bromodichloromethane           | 0.349  | 0.425               | 0.501  | 0.544               | 0.528  | 0.545               | 0.482 | 16.4  |
| 4-Methyl-2-Pentanone           | 0.437  | 0.527               | 0.562  | 0.606               | 0.539  | 0.548               | 0.537 | 10.4  |
| Toluene                        | 0.636  | 0.839               | 0.883  | 0.921               | 0.850  | 0.852               | 0.830 | 12    |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05  
 Lab Code: CHEM Case No.: P4995 SAS No.: P4995 SDG No.: P4995  
 Instrument ID: MSVOA\_X Calibration Date(s): 11/21/2024 11/21/2024  
 Heated Purge: (Y/N) N Calibration Time(s): 09:47 12:03  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                | RRF001 = VX043925.D | RRF005 = VX043926.D | RRF020 = VX043927.D | RRF050 = VX043928.D | RRF100 = VX043929.D | RRF150 = VX043930.D |       |       |
|-----------------------------|---------------------|---------------------|---------------------|---------------------|---------------------|---------------------|-------|-------|
| COMPOUND                    | RRF001              | RRF005              | RRF020              | RRF050              | RRF100              | RRF150              | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.341               | 0.413               | 0.507               | 0.569               | 0.549               | 0.570               | 0.491 | 19.2  |
| cis-1,3-Dichloropropene     | 0.409               | 0.462               | 0.569               | 0.611               | 0.580               | 0.601               | 0.539 | 15.4  |
| 1,1,2-Trichloroethane       | 0.287               | 0.340               | 0.375               | 0.370               | 0.340               | 0.345               | 0.343 | 9.1   |
| 2-Hexanone                  | 0.297               | 0.393               | 0.431               | 0.471               | 0.410               | 0.419               | 0.403 | 14.5  |
| Dibromochloromethane        | 0.249               | 0.292               | 0.344               | 0.396               | 0.380               | 0.401               | 0.344 | 18    |
| 1,2-Dibromoethane           | 0.258               | 0.339               | 0.373               | 0.385               | 0.358               | 0.367               | 0.346 | 13.3  |
| Tetrachloroethene           | 0.304               | 0.342               | 0.336               | 0.349               | 0.315               | 0.332               | 0.330 | 5.1   |
| Chlorobenzene               | 1.050               | 1.109               | 1.107               | 1.145               | 1.069               | 1.108               | 1.098 | 3.1   |
| Ethyl Benzene               | 1.579               | 1.831               | 1.910               | 2.024               | 1.884               | 1.938               | 1.861 | 8.2   |
| m/p-Xylenes                 | 0.610               | 0.663               | 0.704               | 0.762               | 0.700               | 0.725               | 0.694 | 7.6   |
| o-Xylene                    | 0.543               | 0.665               | 0.691               | 0.757               | 0.693               | 0.716               | 0.678 | 10.7  |
| Styrene                     | 0.859               | 1.025               | 1.174               | 1.263               | 1.186               | 1.217               | 1.121 | 13.5  |
| Bromoform                   | 0.145               | 0.200               | 0.222               | 0.284               | 0.286               | 0.306               | 0.240 | 25.9  |
| Isopropylbenzene            | 3.435               | 3.927               | 3.996               | 4.097               | 3.679               | 3.927               | 3.843 | 6.3   |
| 1,1,2,2-Tetrachloroethane   | 1.198               | 1.356               | 1.375               | 1.393               | 1.260               | 1.314               | 1.316 | 5.7   |
| 1,3-Dichlorobenzene         | 1.580               | 1.656               | 1.701               | 1.764               | 1.623               | 1.696               | 1.670 | 3.9   |
| 1,4-Dichlorobenzene         | 1.730               | 1.638               | 1.706               | 1.782               | 1.650               | 1.710               | 1.702 | 3.1   |
| 1,2-Dichlorobenzene         | 1.506               | 1.733               | 1.723               | 1.768               | 1.642               | 1.740               | 1.685 | 5.8   |
| 1,2-Dibromo-3-Chloropropane | 0.213               | 0.254               | 0.252               | 0.284               | 0.275               | 0.304               | 0.264 | 12    |
| 1,2,4-Trichlorobenzene      | 0.849               | 0.905               | 0.987               | 1.084               | 1.012               | 1.125               | 0.994 | 10.5  |
| 1,2,3-Trichlorobenzene      | 0.753               | 0.960               | 1.055               | 1.118               | 1.018               | 1.130               | 1.006 | 13.8  |
| 1,2-Dichloroethane-d4       |                     | 0.926               | 0.876               | 0.751               | 0.855               | 0.880               | 0.858 | 7.6   |
| Dibromofluoromethane        |                     | 0.371               | 0.353               | 0.327               | 0.359               | 0.376               | 0.357 | 5.4   |
| Toluene-d8                  |                     | 1.327               | 1.237               | 1.104               | 1.232               | 1.257               | 1.232 | 6.6   |
| 4-Bromofluorobenzene        |                     | 0.401               | 0.414               | 0.387               | 0.438               | 0.455               | 0.419 | 6.6   |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.