

LAB CHRONICLE

OrderID:	P4997	OrderDate:	11/25/2024 1:27:00 PM
Client:	New York City DEP of Environmental Protection/BWS	Project:	Industrial Wastewater Discharge Permit - Fall 2024
Contact:	Nicholas Prokopowicz	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4997-01	14B-1	Water	VOCMS Group1	624.1	11/25/24		11/25/24	11/25/24
P4997-02	14B-2	Water	VOCMS Group1	624.1	11/25/24		11/25/24	11/25/24
P4997-03	14B-3	Water	VOCMS Group1	624.1	11/25/24		11/25/24	11/25/24
P4997-04	14B-4	Water	VOCMS Group1	624.1	11/25/24		11/25/24	11/25/24



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P4997

Client: New York City DEP of Environmental Protection/BW

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID: 0

Total Voc :
Total Concentration:

- A
- B
- C
- D
- E
- F
- G



SAMPLE DATA

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS		Date Collected:	11/25/24	
Project:	Industrial Wastewater Discharge Permit - Fall 2024		Date Received:	11/25/24	
Client Sample ID:	14B-1		SDG No.:	P4997	
Lab Sample ID:	P4997-01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085029.D	5		11/25/24 15:42	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	6.10	U	6.10	25.0	ug/L
74-87-3	Chloromethane	5.90	U	5.90	25.0	ug/L
75-01-4	Vinyl Chloride	6.10	U	6.10	25.0	ug/L
74-83-9	Bromomethane	6.90	U	6.90	25.0	ug/L
75-00-3	Chloroethane	14.6	U	14.6	25.0	ug/L
75-69-4	Trichlorofluoromethane	5.10	U	5.10	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.30	U	5.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	5.30	U	5.30	25.0	ug/L
67-64-1	Acetone	24.6	U	24.6	130	ug/L
75-15-0	Carbon Disulfide	4.70	U	4.70	25.0	ug/L
1634-04-4	Methyl tert-Butyl Ether	4.20	U	4.20	25.0	ug/L
79-20-9	Methyl Acetate	6.10	U	6.10	25.0	ug/L
75-09-2	Methylene Chloride	6.10	U	6.10	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	4.80	U	4.80	25.0	ug/L
75-34-3	1,1-Dichloroethane	4.10	U	4.10	25.0	ug/L
110-82-7	Cyclohexane	5.20	U	5.20	25.0	ug/L
78-93-3	2-Butanone	17.8	U	17.8	130	ug/L
56-23-5	Carbon Tetrachloride	4.60	U	4.60	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	4.50	U	4.50	25.0	ug/L
67-66-3	Chloroform	3.60	U	3.60	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	4.00	U	4.00	25.0	ug/L
108-87-2	Methylcyclohexane	4.50	U	4.50	25.0	ug/L
71-43-2	Benzene	3.50	U	3.50	25.0	ug/L
107-06-2	1,2-Dichloroethane	3.80	U	3.80	25.0	ug/L
79-01-6	Trichloroethene	3.90	U	3.90	25.0	ug/L
78-87-5	1,2-Dichloropropane	3.30	U	3.30	25.0	ug/L
75-27-4	Bromodichloromethane	4.10	U	4.10	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	20.8	U	20.8	130	ug/L
108-88-3	Toluene	3.60	U	3.60	25.0	ug/L
10061-02-6	t-1,3-Dichloropropene	4.00	U	4.00	25.0	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-1	SDG No.:	P4997
Lab Sample ID:	P4997-01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085029.D	5		11/25/24 15:42	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	4.20	U	4.20	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	3.40	U	3.40	25.0	ug/L
591-78-6	2-Hexanone	19.7	U	19.7	130	ug/L
124-48-1	Dibromochloromethane	3.60	U	3.60	25.0	ug/L
106-93-4	1,2-Dibromoethane	3.90	U	3.90	25.0	ug/L
127-18-4	Tetrachloroethene	4.70	U	4.70	25.0	ug/L
108-90-7	Chlorobenzene	3.40	U	3.40	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	U	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	U	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	U	4.10	25.0	ug/L
100-42-5	Styrene	4.00	U	4.00	25.0	ug/L
75-25-2	Bromoform	5.00	U	5.00	25.0	ug/L
98-82-8	Isopropylbenzene	4.30	U	4.30	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	3.00	U	3.00	25.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	3.70	U	3.70	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	4.80	U	4.80	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	U	4.60	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.50	U	5.50	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.10	U	7.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.9		91 - 110	103%	SPK: 30
2037-26-5	Toluene-d8	28.3		91 - 112	94%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.0		63 - 112	83%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	26200	7.806			
540-36-3	1,4-Difluorobenzene	134000	9.1			
3114-55-4	Chlorobenzene-d5	115000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-1	SDG No.:	P4997
Lab Sample ID:	P4997-01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085029.D	5		11/25/24 15:42	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-2	SDG No.:	P4997
Lab Sample ID:	P4997-02	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085030.D	5		11/25/24 16:06	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	6.10	U	6.10	25.0	ug/L
74-87-3	Chloromethane	5.90	U	5.90	25.0	ug/L
75-01-4	Vinyl Chloride	6.10	U	6.10	25.0	ug/L
74-83-9	Bromomethane	6.90	U	6.90	25.0	ug/L
75-00-3	Chloroethane	14.6	U	14.6	25.0	ug/L
75-69-4	Trichlorofluoromethane	5.10	U	5.10	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.30	U	5.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	5.30	U	5.30	25.0	ug/L
67-64-1	Acetone	24.6	U	24.6	130	ug/L
75-15-0	Carbon Disulfide	4.70	U	4.70	25.0	ug/L
1634-04-4	Methyl tert-Butyl Ether	4.20	U	4.20	25.0	ug/L
79-20-9	Methyl Acetate	6.10	U	6.10	25.0	ug/L
75-09-2	Methylene Chloride	6.10	U	6.10	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	4.80	U	4.80	25.0	ug/L
75-34-3	1,1-Dichloroethane	4.10	U	4.10	25.0	ug/L
110-82-7	Cyclohexane	5.20	U	5.20	25.0	ug/L
78-93-3	2-Butanone	17.8	U	17.8	130	ug/L
56-23-5	Carbon Tetrachloride	4.60	U	4.60	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	4.50	U	4.50	25.0	ug/L
67-66-3	Chloroform	3.60	U	3.60	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	4.00	U	4.00	25.0	ug/L
108-87-2	Methylcyclohexane	4.50	U	4.50	25.0	ug/L
71-43-2	Benzene	3.50	U	3.50	25.0	ug/L
107-06-2	1,2-Dichloroethane	3.80	U	3.80	25.0	ug/L
79-01-6	Trichloroethene	3.90	U	3.90	25.0	ug/L
78-87-5	1,2-Dichloropropane	3.30	U	3.30	25.0	ug/L
75-27-4	Bromodichloromethane	4.10	U	4.10	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	20.8	U	20.8	130	ug/L
108-88-3	Toluene	3.60	U	3.60	25.0	ug/L
10061-02-6	t-1,3-Dichloropropene	4.00	U	4.00	25.0	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-2	SDG No.:	P4997
Lab Sample ID:	P4997-02	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085030.D	5		11/25/24 16:06	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	4.20	U	4.20	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	3.40	U	3.40	25.0	ug/L
591-78-6	2-Hexanone	19.7	U	19.7	130	ug/L
124-48-1	Dibromochloromethane	3.60	U	3.60	25.0	ug/L
106-93-4	1,2-Dibromoethane	3.90	U	3.90	25.0	ug/L
127-18-4	Tetrachloroethene	4.70	U	4.70	25.0	ug/L
108-90-7	Chlorobenzene	3.40	U	3.40	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	U	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	U	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	U	4.10	25.0	ug/L
100-42-5	Styrene	4.00	U	4.00	25.0	ug/L
75-25-2	Bromoform	5.00	U	5.00	25.0	ug/L
98-82-8	Isopropylbenzene	4.30	U	4.30	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	3.00	U	3.00	25.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	3.70	U	3.70	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	4.80	U	4.80	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	U	4.60	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.50	U	5.50	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.10	U	7.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.1		91 - 110	100%	SPK: 30
2037-26-5	Toluene-d8	28.6		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	24.9		63 - 112	83%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	23400	7.812			
540-36-3	1,4-Difluorobenzene	118000	9.094			
3114-55-4	Chlorobenzene-d5	101000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-2	SDG No.:	P4997
Lab Sample ID:	P4997-02	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085030.D	5		11/25/24 16:06	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-3	SDG No.:	P4997
Lab Sample ID:	P4997-03	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085031.D	5		11/25/24 16:30	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	6.10	U	6.10	25.0	ug/L
74-87-3	Chloromethane	5.90	U	5.90	25.0	ug/L
75-01-4	Vinyl Chloride	6.10	U	6.10	25.0	ug/L
74-83-9	Bromomethane	6.90	U	6.90	25.0	ug/L
75-00-3	Chloroethane	14.6	U	14.6	25.0	ug/L
75-69-4	Trichlorofluoromethane	5.10	U	5.10	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.30	U	5.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	5.30	U	5.30	25.0	ug/L
67-64-1	Acetone	24.6	U	24.6	130	ug/L
75-15-0	Carbon Disulfide	4.70	U	4.70	25.0	ug/L
1634-04-4	Methyl tert-Butyl Ether	4.20	U	4.20	25.0	ug/L
79-20-9	Methyl Acetate	6.10	U	6.10	25.0	ug/L
75-09-2	Methylene Chloride	6.10	U	6.10	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	4.80	U	4.80	25.0	ug/L
75-34-3	1,1-Dichloroethane	4.10	U	4.10	25.0	ug/L
110-82-7	Cyclohexane	5.20	U	5.20	25.0	ug/L
78-93-3	2-Butanone	17.8	U	17.8	130	ug/L
56-23-5	Carbon Tetrachloride	4.60	U	4.60	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	4.50	U	4.50	25.0	ug/L
67-66-3	Chloroform	3.60	U	3.60	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	4.00	U	4.00	25.0	ug/L
108-87-2	Methylcyclohexane	4.50	U	4.50	25.0	ug/L
71-43-2	Benzene	3.50	U	3.50	25.0	ug/L
107-06-2	1,2-Dichloroethane	3.80	U	3.80	25.0	ug/L
79-01-6	Trichloroethene	3.90	U	3.90	25.0	ug/L
78-87-5	1,2-Dichloropropane	3.30	U	3.30	25.0	ug/L
75-27-4	Bromodichloromethane	4.10	U	4.10	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	20.8	U	20.8	130	ug/L
108-88-3	Toluene	3.60	U	3.60	25.0	ug/L
10061-02-6	t-1,3-Dichloropropene	4.00	U	4.00	25.0	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-3	SDG No.:	P4997
Lab Sample ID:	P4997-03	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085031.D	5		11/25/24 16:30	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	4.20	U	4.20	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	3.40	U	3.40	25.0	ug/L
591-78-6	2-Hexanone	19.7	U	19.7	130	ug/L
124-48-1	Dibromochloromethane	3.60	U	3.60	25.0	ug/L
106-93-4	1,2-Dibromoethane	3.90	U	3.90	25.0	ug/L
127-18-4	Tetrachloroethene	4.70	U	4.70	25.0	ug/L
108-90-7	Chlorobenzene	3.40	U	3.40	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	U	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	U	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	U	4.10	25.0	ug/L
100-42-5	Styrene	4.00	U	4.00	25.0	ug/L
75-25-2	Bromoform	5.00	U	5.00	25.0	ug/L
98-82-8	Isopropylbenzene	4.30	U	4.30	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	3.00	U	3.00	25.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	3.70	U	3.70	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	4.80	U	4.80	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	U	4.60	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.50	U	5.50	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.10	U	7.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.3		91 - 110	101%	SPK: 30
2037-26-5	Toluene-d8	28.4		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	23.8		63 - 112	79%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	25300	7.812			
540-36-3	1,4-Difluorobenzene	124000	9.094			
3114-55-4	Chlorobenzene-d5	109000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-3	SDG No.:	P4997
Lab Sample ID:	P4997-03	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085031.D	5		11/25/24 16:30	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-4	SDG No.:	P4997
Lab Sample ID:	P4997-04	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085032.D	5		11/25/24 16:54	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	6.10	U	6.10	25.0	ug/L
74-87-3	Chloromethane	5.90	U	5.90	25.0	ug/L
75-01-4	Vinyl Chloride	6.10	U	6.10	25.0	ug/L
74-83-9	Bromomethane	6.90	U	6.90	25.0	ug/L
75-00-3	Chloroethane	14.6	U	14.6	25.0	ug/L
75-69-4	Trichlorofluoromethane	5.10	U	5.10	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.30	U	5.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	5.30	U	5.30	25.0	ug/L
67-64-1	Acetone	24.6	U	24.6	130	ug/L
75-15-0	Carbon Disulfide	4.70	U	4.70	25.0	ug/L
1634-04-4	Methyl tert-Butyl Ether	4.20	U	4.20	25.0	ug/L
79-20-9	Methyl Acetate	6.10	U	6.10	25.0	ug/L
75-09-2	Methylene Chloride	6.10	U	6.10	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	4.80	U	4.80	25.0	ug/L
75-34-3	1,1-Dichloroethane	4.10	U	4.10	25.0	ug/L
110-82-7	Cyclohexane	5.20	U	5.20	25.0	ug/L
78-93-3	2-Butanone	17.8	U	17.8	130	ug/L
56-23-5	Carbon Tetrachloride	4.60	U	4.60	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	4.50	U	4.50	25.0	ug/L
67-66-3	Chloroform	3.60	U	3.60	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	4.00	U	4.00	25.0	ug/L
108-87-2	Methylcyclohexane	4.50	U	4.50	25.0	ug/L
71-43-2	Benzene	3.50	U	3.50	25.0	ug/L
107-06-2	1,2-Dichloroethane	3.80	U	3.80	25.0	ug/L
79-01-6	Trichloroethene	3.90	U	3.90	25.0	ug/L
78-87-5	1,2-Dichloropropane	3.30	U	3.30	25.0	ug/L
75-27-4	Bromodichloromethane	4.10	U	4.10	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	20.8	U	20.8	130	ug/L
108-88-3	Toluene	3.60	U	3.60	25.0	ug/L
10061-02-6	t-1,3-Dichloropropene	4.00	U	4.00	25.0	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-4	SDG No.:	P4997
Lab Sample ID:	P4997-04	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085032.D	5		11/25/24 16:54	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	4.20	U	4.20	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	3.40	U	3.40	25.0	ug/L
591-78-6	2-Hexanone	19.7	U	19.7	130	ug/L
124-48-1	Dibromochloromethane	3.60	U	3.60	25.0	ug/L
106-93-4	1,2-Dibromoethane	3.90	U	3.90	25.0	ug/L
127-18-4	Tetrachloroethene	4.70	U	4.70	25.0	ug/L
108-90-7	Chlorobenzene	3.40	U	3.40	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	U	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	U	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	U	4.10	25.0	ug/L
100-42-5	Styrene	4.00	U	4.00	25.0	ug/L
75-25-2	Bromoform	5.00	U	5.00	25.0	ug/L
98-82-8	Isopropylbenzene	4.30	U	4.30	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	3.00	U	3.00	25.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	3.70	U	3.70	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	4.80	U	4.80	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	4.40	U	4.40	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	U	4.60	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.50	U	5.50	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.10	U	7.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.3		91 - 110	101%	SPK: 30
2037-26-5	Toluene-d8	28.1		91 - 112	94%	SPK: 30
460-00-4	4-Bromofluorobenzene	23.9		63 - 112	80%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	23800	7.806			
540-36-3	1,4-Difluorobenzene	119000	9.094			
3114-55-4	Chlorobenzene-d5	106000	11.865			



QC SUMMARY

Surrogate Summary

SDG No.: P4997

Client: New York City DEP of Environmental Protection/BWS

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4997-01	14B-1	1,2-Dichloroethane-d4	30	30.9	103	91	110
		Toluene-d8	30	28.3	94	91	112
		4-Bromofluorobenzene	30	25.0	83	63	112
P4997-02	14B-2	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	28.6	95	91	112
		4-Bromofluorobenzene	30	24.9	83	63	112
P4997-03	14B-3	1,2-Dichloroethane-d4	30	30.3	101	91	110
		Toluene-d8	30	28.4	95	91	112
		4-Bromofluorobenzene	30	23.8	79	63	112
P4997-04	14B-4	1,2-Dichloroethane-d4	30	30.3	101	91	110
		Toluene-d8	30	28.1	94	91	112
		4-Bromofluorobenzene	30	23.9	80	63	112
VN1125WBL01	VN1125WBL01	1,2-Dichloroethane-d4	30	30.7	102	91	110
		Toluene-d8	30	28.6	95	91	112
		4-Bromofluorobenzene	30	25.2	84	63	112
VN1125WBS01	VN1125WBS01	1,2-Dichloroethane-d4	30	29.6	99	91	110
		Toluene-d8	30	29.5	98	91	112
		4-Bromofluorobenzene	30	29.6	99	63	112
VN1125WBSD0	VN1125WBSD01	1,2-Dichloroethane-d4	30	29.2	97	91	110
		Toluene-d8	30	29.4	98	91	112
		4-Bromofluorobenzene	30	30.5	102	63	112

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4997

Client: New York City DEP of Environmental Prot

Analytical Method: SW624.1

Datafile : VN085023.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1125WBS01	Dichlorodifluoromethane	20	19.6	ug/L	98			72	118	
	Chloromethane	20	16.8	ug/L	84			1	205	
	Vinyl Chloride	20	18.1	ug/L	91			5	195	
	Bromomethane	20	18.8	ug/L	94			15	185	
	Chloroethane	20	18.8	ug/L	94			40	160	
	Trichlorofluoromethane	20	19.4	ug/L	97			50	150	
	1,1,2-Trichlorotrifluoroethane	20	19.1	ug/L	96			64	127	
	1,1-Dichloroethene	20	18.5	ug/L	93			50	150	
	Acetone	100	93.4	ug/L	93			41	148	
	Carbon disulfide	20	17.3	ug/L	86			76	107	
	Methyl tert-butyl Ether	20	18.1	ug/L	91			82	114	
	Methyl Acetate	20	15.3	ug/L	77			63	139	
	Methylene Chloride	20	17.6	ug/L	88			60	140	
	trans-1,2-Dichloroethene	20	18.0	ug/L	90			70	130	
	1,1-Dichloroethane	20	17.8	ug/L	89			70	130	
	Cyclohexane	20	18.5	ug/L	93			79	113	
	2-Butanone	100	91.1	ug/L	91			69	129	
	Carbon Tetrachloride	20	19.4	ug/L	97			70	130	
	cis-1,2-Dichloroethene	20	17.9	ug/L	90			81	112	
	Chloroform	20	18.4	ug/L	92			70	135	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			70	130	
	Methylcyclohexane	20	18.5	ug/L	93			79	112	
	Benzene	20	18.4	ug/L	92			65	135	
	1,2-Dichloroethane	20	18.4	ug/L	92			70	130	
	Trichloroethene	20	18.8	ug/L	94			65	135	
	1,2-Dichloropropane	20	18.7	ug/L	94			35	165	
	Bromodichloromethane	20	18.2	ug/L	91			65	135	
	4-Methyl-2-Pentanone	100	85.6	ug/L	86			73	131	
	Toluene	20	18.2	ug/L	91			70	130	
	t-1,3-Dichloropropene	20	17.5	ug/L	88			50	150	
	cis-1,3-Dichloropropene	20	18.2	ug/L	91			25	175	
	1,1,2-Trichloroethane	20	18.7	ug/L	94			70	130	
	2-Hexanone	100	88.0	ug/L	88			72	128	
	Dibromochloromethane	20	18.2	ug/L	91			70	135	
	1,2-Dibromoethane	20	17.6	ug/L	88			86	114	
	Tetrachloroethene	20	19.2	ug/L	96			70	130	
	Chlorobenzene	20	18.4	ug/L	92			65	135	
	Ethyl Benzene	20	17.9	ug/L	90			60	140	
	m/p-Xylenes	40	36.4	ug/L	91			87	111	
	o-Xylene	20	17.8	ug/L	89			87	111	
	Styrene	20	17.7	ug/L	89			85	106	
	Bromoform	20	17.8	ug/L	89			70	130	
	Isopropylbenzene	20	18.2	ug/L	91			86	112	
	1,1,2,2-Tetrachloroethane	20	18.5	ug/L	93			60	140	
	1,3,5-Trimethylbenzene	20	18.7	ug/L	94			66	139	
	1,3-Dichlorobenzene	20	18.7	ug/L	94			70	130	
	1,4-Dichlorobenzene	20	18.0	ug/L	90			65	135	
	1,2-Dichlorobenzene	20	18.1	ug/L	91			65	135	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4997

Client: New York City DEP of Environmental Prot

Analytical Method: SW624.1 **Datafile :** VN085023.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1125WBS01	1,2-Dibromo-3-Chloropropane	20	18.2	ug/L	91			69	122	
	1,2,4-Trichlorobenzene	20	18.1	ug/L	91			61	118	
	1,2,3-Trichlorobenzene	20	18.1	ug/L	91			38	159	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4997

Client: New York City DEP of Environmental Prot

Analytical Method: SW624.1

Datafile : VN085024.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1125WBSD01	Dichlorodifluoromethane	20	19.8	ug/L	99	1		72	118	20
	Chloromethane	20	18.2	ug/L	91	8		1	205	20
	Vinyl Chloride	20	18.2	ug/L	91	0		5	195	20
	Bromomethane	20	20.6	ug/L	103	9		15	185	20
	Chloroethane	20	19.8	ug/L	99	5		40	160	20
	Trichlorofluoromethane	20	19.5	ug/L	98	1		50	150	20
	1,1,2-Trichlorotrifluoroethane	20	20.1	ug/L	101	5		64	127	20
	1,1-Dichloroethene	20	18.7	ug/L	94	1		50	150	20
	Acetone	100	110	ug/L	110	17		41	148	20
	Carbon disulfide	20	17.7	ug/L	89	3		76	107	20
	Methyl tert-butyl Ether	20	20.7	ug/L	104	13		82	114	20
	Methyl Acetate	20	17.5	ug/L	88	13		63	139	20
	Methylene Chloride	20	19.0	ug/L	95	8		60	140	20
	trans-1,2-Dichloroethene	20	18.8	ug/L	94	4		70	130	20
	1,1-Dichloroethane	20	19.1	ug/L	96	8		70	130	20
	Cyclohexane	20	19.6	ug/L	98	5		79	113	20
	2-Butanone	100	110	ug/L	110	19		69	129	20
	Carbon Tetrachloride	20	20.3	ug/L	102	5		70	130	20
	cis-1,2-Dichloroethene	20	19.8	ug/L	99	10		81	112	20
	Chloroform	20	19.4	ug/L	97	5		70	135	20
	1,1,1-Trichloroethane	20	20.0	ug/L	100	5		70	130	20
	Methylcyclohexane	20	19.2	ug/L	96	3		79	112	20
	Benzene	20	19.9	ug/L	100	8		65	135	20
	1,2-Dichloroethane	20	21.2	ug/L	106	14		70	130	20
	Trichloroethene	20	20.9	ug/L	104	10		65	135	20
	1,2-Dichloropropane	20	20.3	ug/L	102	8		35	165	20
	Bromodichloromethane	20	20.5	ug/L	103	12		65	135	20
	4-Methyl-2-Pentanone	100	100	ug/L	100	15		73	131	20
	Toluene	20	19.6	ug/L	98	7		70	130	20
	t-1,3-Dichloropropene	20	21.1	ug/L	106	19		50	150	20
	cis-1,3-Dichloropropene	20	21.2	ug/L	106	15		25	175	20
	1,1,2-Trichloroethane	20	21.5	ug/L	108	14		70	130	20
	2-Hexanone	100	110	ug/L	110	22	*	72	128	20
	Dibromochloromethane	20	21.4	ug/L	107	16		70	135	20
	1,2-Dibromoethane	20	21.3	ug/L	106	19		86	114	20
	Tetrachloroethene	20	20.3	ug/L	102	6		70	130	20
	Chlorobenzene	20	20.3	ug/L	102	10		65	135	20
	Ethyl Benzene	20	19.4	ug/L	97	7		60	140	20
	m/p-Xylenes	40	39.5	ug/L	99	8		87	111	20
	o-Xylene	20	20.3	ug/L	102	14		87	111	20
	Styrene	20	20.5	ug/L	103	15		85	106	20
	Bromoform	20	21.1	ug/L	106	17		70	130	20
	Isopropylbenzene	20	19.8	ug/L	99	8		86	112	20
	1,1,2,2-Tetrachloroethane	20	20.9	ug/L	104	11		60	140	20
	1,3,5-Trimethylbenzene	20	21.4	ug/L	107	13		66	139	20
	1,3-Dichlorobenzene	20	20.1	ug/L	101	7		70	130	20
	1,4-Dichlorobenzene	20	20.2	ug/L	101	12		65	135	20
	1,2-Dichlorobenzene	20	20.2	ug/L	101	10		65	135	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4997

Client: New York City DEP of Environmental Prot

Analytical Method: SW624.1

Datafile : VN085024.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1125WBSD01	1,2-Dibromo-3-Chloropropane	20	20.6	ug/L	103	12		69	122	20
	1,2,4-Trichlorobenzene	20	20.9	ug/L	104	13		61	118	20
	1,2,3-Trichlorobenzene	20	20.9	ug/L	104	13		38	159	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN1125WBL01

Lab Name: CHEMTECH

Contract: NEWY17

Lab Code: CHEM Case No.: P4997

SAS No.: P4997 SDG NO.: P4997

Lab File ID: VN085025.D

Lab Sample ID: VN1125WBL01

Date Analyzed: 11/25/2024

Time Analyzed: 12:39

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN1125WBS01	VN1125WBS01	VN085023.D	11/25/2024
VN1125WBSD01	VN1125WBSD01	VN085024.D	11/25/2024
14B-1	P4997-01	VN085029.D	11/25/2024
14B-2	P4997-02	VN085030.D	11/25/2024
14B-3	P4997-03	VN085031.D	11/25/2024
14B-4	P4997-04	VN085032.D	11/25/2024

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: NEWY17
Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG NO.: P4997
Lab File ID: VN084612.D BFB Injection Date: 10/31/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:56
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.5
75	30.0 - 60.0% of mass 95	52.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	73.9
175	5.0 - 9.0% of mass 174	5.6 (7.5) 1
176	95.0 - 101.0% of mass 174	72.1 (97.5) 1
177	5.0 - 9.0% of mass 176	4.7 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VN084613.D	10/31/2024	12:08
VSTDIC020	VSTDIC020	VN084614.D	10/31/2024	12:32
VSTDIC050	VSTDIC050	VN084615.D	10/31/2024	12:56
VSTDIC100	VSTDIC100	VN084616.D	10/31/2024	13:20
VSTDIC150	VSTDIC150	VN084617.D	10/31/2024	13:44

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: NEWY17
Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG NO.: P4997
Lab File ID: VN085021.D BFB Injection Date: 11/25/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:26
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	52.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.6 (7.4) 1
176	95.0 - 101.0% of mass 174	73.2 (96.4) 1
177	5.0 - 9.0% of mass 176	4.5 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC020	VSTDCCC020	VN085022.D	11/25/2024	10:59
VN1125WBS01	VN1125WBS01	VN085023.D	11/25/2024	11:35
VN1125WBSD01	VN1125WBSD01	VN085024.D	11/25/2024	12:15
VN1125WBL01	VN1125WBL01	VN085025.D	11/25/2024	12:39
14B-1	P4997-01	VN085029.D	11/25/2024	15:42
14B-2	P4997-02	VN085030.D	11/25/2024	16:06
14B-3	P4997-03	VN085031.D	11/25/2024	16:30
14B-4	P4997-04	VN085032.D	11/25/2024	16:54

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NEWY17
Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG NO.: P4997
Lab File ID: VN085022.D Date Analyzed: 11/25/2024
Instrument ID: MSVOA_N Time Analyzed: 10:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	28386	7.81	151585	9.09	140690	11.87
UPPER LIMIT	56772	8.312	303170	9.594	281380	12.365
LOWER LIMIT	14193	7.312	75792.5	8.594	70345	11.365
EPA SAMPLE NO.						
14B-1	26188	7.81	133606	9.10	114616	11.87
14B-2	23416	7.81	118468	9.09	101396	11.87
14B-3	25312	7.81	123987	9.09	109090	11.87
14B-4	23776	7.81	118523	9.09	105687	11.87
VN1125WBL01	24457	7.81	127294	9.09	111508	11.87
VN1125WBS01	29961	7.81	160469	9.09	146596	11.87
VN1125WBSD01	28214	7.81	145410	9.10	133637	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NEWY17
 Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG NO.: P4997
 Lab File ID: VN085022.D Date Analyzed: 11/25/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
14B-1	0	0.00				
14B-2	0	0.00				
14B-3	0	0.00				
14B-4	0	0.00				
VN1125WBL01	0	0.00				
VN1125WBS01	0	0.00				
VN1125WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS		Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024		Date Received:	
Client Sample ID:	VN1125WBL01		SDG No.:	P4997
Lab Sample ID:	VN1125WBL01		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085025.D	1		11/25/24 12:39	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	1.20	U	1.20	5.00	ug/L
74-87-3	Chloromethane	1.20	U	1.20	5.00	ug/L
75-01-4	Vinyl Chloride	1.20	U	1.20	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	2.90	U	2.90	5.00	ug/L
75-69-4	Trichlorofluoromethane	1.00	U	1.00	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/L
75-35-4	1,1-Dichloroethene	1.10	U	1.10	5.00	ug/L
67-64-1	Acetone	4.90	U	4.90	25.0	ug/L
75-15-0	Carbon Disulfide	0.93	U	0.93	5.00	ug/L
1634-04-4	Methyl tert-Butyl Ether	0.83	U	0.83	5.00	ug/L
79-20-9	Methyl Acetate	1.20	U	1.20	5.00	ug/L
75-09-2	Methylene Chloride	1.20	U	1.20	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.95	U	0.95	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.81	U	0.81	5.00	ug/L
110-82-7	Cyclohexane	1.00	U	1.00	5.00	ug/L
78-93-3	2-Butanone	3.60	U	3.60	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.91	U	0.91	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.90	U	0.90	5.00	ug/L
67-66-3	Chloroform	0.72	U	0.72	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.79	U	0.79	5.00	ug/L
108-87-2	Methylcyclohexane	0.89	U	0.89	5.00	ug/L
71-43-2	Benzene	0.69	U	0.69	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.75	U	0.75	5.00	ug/L
79-01-6	Trichloroethene	0.77	U	0.77	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.65	U	0.65	5.00	ug/L
75-27-4	Bromodichloromethane	0.81	U	0.81	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	4.20	U	4.20	25.0	ug/L
108-88-3	Toluene	0.72	U	0.72	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.79	U	0.79	5.00	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	
Client Sample ID:	VN1125WBL01	SDG No.:	P4997
Lab Sample ID:	VN1125WBL01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085025.D	1		11/25/24 12:39	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	0.83	U	0.83	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.68	U	0.68	5.00	ug/L
591-78-6	2-Hexanone	3.90	U	3.90	25.0	ug/L
124-48-1	Dibromochloromethane	0.72	U	0.72	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.78	U	0.78	5.00	ug/L
127-18-4	Tetrachloroethene	0.94	U	0.94	5.00	ug/L
108-90-7	Chlorobenzene	0.67	U	0.67	5.00	ug/L
100-41-4	Ethyl Benzene	0.73	U	0.73	5.00	ug/L
179601-23-1	m/p-Xylenes	1.70	U	1.70	10.0	ug/L
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/L
100-42-5	Styrene	0.80	U	0.80	5.00	ug/L
75-25-2	Bromoform	1.00	U	1.00	5.00	ug/L
98-82-8	Isopropylbenzene	0.85	U	0.85	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.60	U	0.60	5.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.74	U	0.74	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.88	U	0.88	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.95	U	0.95	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.88	U	0.88	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.91	U	0.91	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.10	U	1.10	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.40	U	1.40	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.7		91 - 110	102%	SPK: 30
2037-26-5	Toluene-d8	28.6		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.2		63 - 112	84%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	24500	7.806			
540-36-3	1,4-Difluorobenzene	127000	9.094			
3114-55-4	Chlorobenzene-d5	112000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	
Client Sample ID:	VN1125WBL01	SDG No.:	P4997
Lab Sample ID:	VN1125WBL01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085025.D	1		11/25/24 12:39	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	
Client Sample ID:	VN1125WBS01	SDG No.:	P4997
Lab Sample ID:	VN1125WBS01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085023.D	1		11/25/24 11:35	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.6		1.20	5.00	ug/L
74-87-3	Chloromethane	16.8		1.20	5.00	ug/L
75-01-4	Vinyl Chloride	18.1		1.20	5.00	ug/L
74-83-9	Bromomethane	18.8		1.40	5.00	ug/L
75-00-3	Chloroethane	18.8		2.90	5.00	ug/L
75-69-4	Trichlorofluoromethane	19.4		1.00	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.1		1.10	5.00	ug/L
75-35-4	1,1-Dichloroethene	18.5		1.10	5.00	ug/L
67-64-1	Acetone	93.4		4.90	25.0	ug/L
75-15-0	Carbon Disulfide	17.3		0.93	5.00	ug/L
1634-04-4	Methyl tert-Butyl Ether	18.1		0.83	5.00	ug/L
79-20-9	Methyl Acetate	15.3		1.20	5.00	ug/L
75-09-2	Methylene Chloride	17.6		1.20	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.0		0.95	5.00	ug/L
75-34-3	1,1-Dichloroethane	17.8		0.81	5.00	ug/L
110-82-7	Cyclohexane	18.5		1.00	5.00	ug/L
78-93-3	2-Butanone	91.1		3.60	25.0	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.91	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	17.9		0.90	5.00	ug/L
67-66-3	Chloroform	18.4		0.72	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.79	5.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.89	5.00	ug/L
71-43-2	Benzene	18.4		0.69	5.00	ug/L
107-06-2	1,2-Dichloroethane	18.4		0.75	5.00	ug/L
79-01-6	Trichloroethene	18.8		0.77	5.00	ug/L
78-87-5	1,2-Dichloropropane	18.7		0.65	5.00	ug/L
75-27-4	Bromodichloromethane	18.2		0.81	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	85.6		4.20	25.0	ug/L
108-88-3	Toluene	18.2		0.72	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	17.5		0.79	5.00	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	
Client Sample ID:	VN1125WBS01	SDG No.:	P4997
Lab Sample ID:	VN1125WBS01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085023.D	1		11/25/24 11:35	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	18.2		0.83	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.7		0.68	5.00	ug/L
591-78-6	2-Hexanone	88.0		3.90	25.0	ug/L
124-48-1	Dibromochloromethane	18.2		0.72	5.00	ug/L
106-93-4	1,2-Dibromoethane	17.6		0.78	5.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.94	5.00	ug/L
108-90-7	Chlorobenzene	18.4		0.67	5.00	ug/L
100-41-4	Ethyl Benzene	17.9		0.73	5.00	ug/L
179601-23-1	m/p-Xylenes	36.4		1.70	10.0	ug/L
95-47-6	o-Xylene	17.8		0.82	5.00	ug/L
100-42-5	Styrene	17.7		0.80	5.00	ug/L
75-25-2	Bromoform	17.8		1.00	5.00	ug/L
98-82-8	Isopropylbenzene	18.2		0.85	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.5		0.60	5.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	18.7		0.74	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.7		0.88	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.0		0.95	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.1		0.88	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	18.2		0.91	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.1		1.10	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.1		1.40	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.6		91 - 110	99%	SPK: 30
2037-26-5	Toluene-d8	29.5		91 - 112	98%	SPK: 30
460-00-4	4-Bromofluorobenzene	29.6		63 - 112	99%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	30000	7.806			
540-36-3	1,4-Difluorobenzene	160000	9.094			
3114-55-4	Chlorobenzene-d5	147000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	
Client Sample ID:	VN1125WBS01	SDG No.:	P4997
Lab Sample ID:	VN1125WBS01	Matrix:	Water
Analytical Method:	E624.1	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085023.D	1		11/25/24 11:35	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS		Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024		Date Received:	
Client Sample ID:	VN1125WBSD01		SDG No.:	P4997
Lab Sample ID:	VN1125WBSD01		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085024.D	1		11/25/24 12:15	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.8		1.20	5.00	ug/L
74-87-3	Chloromethane	18.2		1.20	5.00	ug/L
75-01-4	Vinyl Chloride	18.2		1.20	5.00	ug/L
74-83-9	Bromomethane	20.6		1.40	5.00	ug/L
75-00-3	Chloroethane	19.8		2.90	5.00	ug/L
75-69-4	Trichlorofluoromethane	19.5		1.00	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.1		1.10	5.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		1.10	5.00	ug/L
67-64-1	Acetone	110		4.90	25.0	ug/L
75-15-0	Carbon Disulfide	17.7		0.93	5.00	ug/L
1634-04-4	Methyl tert-Butyl Ether	20.7		0.83	5.00	ug/L
79-20-9	Methyl Acetate	17.5		1.20	5.00	ug/L
75-09-2	Methylene Chloride	19.0		1.20	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.8		0.95	5.00	ug/L
75-34-3	1,1-Dichloroethane	19.1		0.81	5.00	ug/L
110-82-7	Cyclohexane	19.6		1.00	5.00	ug/L
78-93-3	2-Butanone	110		3.60	25.0	ug/L
56-23-5	Carbon Tetrachloride	20.3		0.91	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.8		0.90	5.00	ug/L
67-66-3	Chloroform	19.4		0.72	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.0		0.79	5.00	ug/L
108-87-2	Methylcyclohexane	19.2		0.89	5.00	ug/L
71-43-2	Benzene	19.9		0.69	5.00	ug/L
107-06-2	1,2-Dichloroethane	21.2		0.75	5.00	ug/L
79-01-6	Trichloroethene	20.9		0.77	5.00	ug/L
78-87-5	1,2-Dichloropropane	20.3		0.65	5.00	ug/L
75-27-4	Bromodichloromethane	20.5		0.81	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		4.20	25.0	ug/L
108-88-3	Toluene	19.6		0.72	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	21.1		0.79	5.00	ug/L

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS			Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024			Date Received:	
Client Sample ID:	VN1125WBSD01			SDG No.:	P4997
Lab Sample ID:	VN1125WBSD01			Matrix:	Water
Analytical Method:	E624.1			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085024.D	1		11/25/24 12:15	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	21.2		0.83	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.5		0.68	5.00	ug/L
591-78-6	2-Hexanone	110		3.90	25.0	ug/L
124-48-1	Dibromochloromethane	21.4		0.72	5.00	ug/L
106-93-4	1,2-Dibromoethane	21.3		0.78	5.00	ug/L
127-18-4	Tetrachloroethene	20.3		0.94	5.00	ug/L
108-90-7	Chlorobenzene	20.3		0.67	5.00	ug/L
100-41-4	Ethyl Benzene	19.4		0.73	5.00	ug/L
179601-23-1	m/p-Xylenes	39.5		1.70	10.0	ug/L
95-47-6	o-Xylene	20.3		0.82	5.00	ug/L
100-42-5	Styrene	20.5		0.80	5.00	ug/L
75-25-2	Bromoform	21.1		1.00	5.00	ug/L
98-82-8	Isopropylbenzene	19.8		0.85	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.9		0.60	5.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	21.4		0.74	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.88	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.2		0.95	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.2		0.88	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.6		0.91	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.9		1.10	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.9		1.40	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.2		91 - 110	97%	SPK: 30
2037-26-5	Toluene-d8	29.4		91 - 112	98%	SPK: 30
460-00-4	4-Bromofluorobenzene	30.5		63 - 112	102%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	28200	7.806			
540-36-3	1,4-Difluorobenzene	145000	9.1			
3114-55-4	Chlorobenzene-d5	134000	11.865			

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS		Date Collected:	
Project:	Industrial Wastewater Discharge Permit - Fall 2024		Date Received:	
Client Sample ID:	VN1125WBSD01		SDG No.:	P4997
Lab Sample ID:	VN1125WBSD01		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085024.D	1		11/25/24 12:15	VN112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>NEWY17</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4997</u>
Instrument ID:	<u>MSVOA_N</u>	SAS No.:	<u>P4997</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P4997</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>10/31/2024</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>12:08</u>
			<u>18:13</u>

LAB FILE ID:	RRF005 = VN084613.D		RRF020 = VN084614.D		RRF050 = VN084615.D			
	RRF100 = VN084616.D		RRF150 = VN084617.D		RRFCAL = VN084626.D			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRFCAL	RRF	% RSD
Dichlorodifluoromethane	1.762	1.842	1.727	1.647	1.746		1.745	4
Chloromethane	2.294	2.180	1.987	1.942	2.100		2.101	6.8
Vinyl Chloride	1.987	2.059	1.937	1.841	1.929		1.951	4.1
Bromomethane	1.026	1.087	0.943	0.927	0.969		0.990	6.7
Chloroethane	1.361	1.345	1.305	1.162	1.231		1.281	6.5
Trichlorofluoromethane	3.268	3.375	3.091	2.971	3.135		3.168	5
1,1,2-Trichlorotrifluoroethane	1.776	1.938	1.786	1.675	1.777		1.790	5.3
1,1-Dichloroethene	1.685	1.840	1.711	1.699	1.807		1.748	4
Acetone	0.231	0.235	0.223	0.223	0.233		0.229	2.3
Carbon Disulfide	5.140	5.375	5.016	4.956	5.235		5.144	3.3
Methyl tert-Butyl Ether	4.972	5.809	5.618	5.772	6.099		5.654	7.4
Methyl Acetate	3.900	2.690	2.326	2.302	2.408		2.725	24.8
Methylene Chloride	2.119	2.210	1.980	1.930	2.004		2.049	5.6
trans-1,2-Dichloroethene	1.819	1.905	1.810	1.776	1.854		1.833	2.7
1,1-Dichloroethane	3.711	3.802	3.469	3.417	3.534		3.587	4.6
Cyclohexane	0.444	0.511	0.520	0.517	0.539		0.506	7.2
2-Butanone	0.206	0.207	0.207	0.210	0.213		0.208	1.3
Carbon Tetrachloride	0.545	0.536	0.517	0.504	0.517		0.524	3.2
cis-1,2-Dichloroethene	2.121	2.237	2.172	2.141	2.269		2.188	2.9
Chloroform	3.764	3.915	3.598	3.469	3.627		3.675	4.6
1,1,1-Trichloroethane	0.631	0.645	0.604	0.593	0.595		0.614	3.8
Methylcyclohexane	0.429	0.479	0.495	0.504	0.526		0.486	7.5
Benzene	1.549	1.525	1.476	1.460	1.458		1.494	2.8
1,2-Dichloroethane	0.521	0.521	0.488	0.469	0.484		0.497	4.7
Trichloroethene	0.340	0.350	0.331	0.325	0.338		0.337	2.8
1,2-Dichloropropane	0.371	0.372	0.351	0.349	0.355		0.360	3.1
Bromodichloromethane	0.566	0.547	0.526	0.526	0.531		0.539	3.2
4-Methyl-2-Pentanone	0.453	0.490	0.472	0.475	0.478		0.473	2.8
Toluene	1.732	1.781	1.675	1.658	1.698		1.709	2.9
t-1,3-Dichloropropene	0.509	0.529	0.535	0.538	0.551		0.532	2.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>NEWY17</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4997</u>
Instrument ID:	<u>MSVOA_N</u>	SAS No.:	<u>P4997</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P4997</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>10/31/2024</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>12:08</u>
			<u>18:13</u>

LAB FILE ID:		RRF005 = VN084613.D		RRF020 = VN084614.D		RRF050 = VN084615.D		
		RRF100 = VN084616.D		RRF150 = VN084617.D		RRFCAL = VN084626.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRFCAL	RRF	% RSD
cis-1,3-Dichloropropene	0.535	0.580	0.567	0.580	0.592		0.571	3.8
1,1,2-Trichloroethane	0.340	0.353	0.336	0.330	0.328		0.337	3
2-Hexanone	0.325	0.355	0.349	0.354	0.351		0.347	3.6
Dibromochloromethane	0.378	0.400	0.401	0.397	0.398		0.395	2.5
1,2-Dibromoethane	0.334	0.357	0.337	0.339	0.339		0.341	2.6
Tetrachloroethene	0.313	0.331	0.310	0.305	0.311		0.314	3.1
Chlorobenzene	1.030	1.071	1.018	1.011	1.028		1.032	2.3
Ethyl Benzene	1.596	1.824	1.829	1.836	1.894		1.796	6.4
m/p-Xylenes	0.615	0.710	0.706	0.695	0.712		0.687	6
o-Xylene	0.569	0.690	0.681	0.683	0.692		0.663	7.9
Styrene	0.911	1.136	1.179	1.154	1.180		1.112	10.2
Bromoform	0.254	0.262	0.269	0.270	0.268		0.265	2.6
Isopropylbenzene	1.407	1.671	1.715	1.692	1.752		1.647	8.3
1,1,2,2-Tetrachloroethane	0.555	0.568	0.529	0.510	0.509		0.534	4.9
1,3,5-Trimethylbenzene	1.193	1.438	1.480	1.442	1.485		1.408	8.6
1,3-Dichlorobenzene	0.736	0.812	0.793	0.787	0.801		0.786	3.7
1,4-Dichlorobenzene	0.729	0.784	0.790	0.779	0.801		0.776	3.6
1,2-Dichlorobenzene	0.724	0.821	0.787	0.760	0.780		0.774	4.6
1,2-Dibromo-3-Chloropropane	0.099	0.108	0.106	0.098	0.105		0.103	4.3
1,2,4-Trichlorobenzene	0.308	0.383	0.381	0.390	0.424		0.377	11.2
1,2,3-Trichlorobenzene	0.308	0.383	0.381	0.390	0.424		0.377	11.2
1,2-Dichloroethane-d4	2.444	2.473	2.377	2.353	2.413		2.412	2
Toluene-d8	1.451	1.442	1.416	1.393	1.398		1.420	1.8
4-Bromofluorobenzene	0.482	0.506	0.531	0.538	0.526		0.517	4.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: NEWY17

Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG No.: P4997

Instrument ID: MSVOA_N Calibration Date/Time: 11/25/2024 10:59

Lab File ID: VN085022.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.745	1.865		6.88	
Chloromethane	2.101	1.994		-5.09	
Vinyl Chloride	1.951	1.907	0.1	-2.26	
Bromomethane	0.990	1.011	0.1	2.12	
Chloroethane	1.281	1.395		8.9	
Trichlorofluoromethane	3.168	3.320		4.8	
1,1,2-Trichlorotrifluoroethane	1.790	1.908		6.59	
1,1-Dichloroethene	1.748	1.720	0.1	-1.6	
Acetone	0.229	0.227		-0.87	
Carbon Disulfide	5.144	4.833		-6.05	
Methyl tert-Butyl Ether	5.654	5.397		-4.55	
Methyl Acetate	2.725	2.223		-18.42	
Methylene Chloride	2.049	2.022		-1.32	
trans-1,2-Dichloroethene	1.833	1.726		-5.84	
1,1-Dichloroethane	3.587	3.541	0.2	-1.28	
Cyclohexane	0.506	0.513		1.38	
2-Butanone	0.208	0.198		-4.81	
Carbon Tetrachloride	0.524	0.544	0.1	3.82	
cis-1,2-Dichloroethene	2.188	2.063		-5.71	
Chloroform	3.675	3.651	0.2	-0.65	
1,1,1-Trichloroethane	0.614	0.627	0.1	2.12	
Methylcyclohexane	0.486	0.476		-2.06	
Benzene	1.494	1.470	0.5	-1.61	
1,2-Dichloroethane	0.497	0.504	0.1	1.41	
Trichloroethene	0.337	0.335	0.3	-0.59	
1,2-Dichloropropane	0.360	0.358		-0.56	
Bromodichloromethane	0.539	0.534	0.2	-0.93	
4-Methyl-2-Pentanone	0.473	0.432		-8.67	
Toluene	1.709	1.640	0.4	-4.04	
t-1,3-Dichloropropene	0.532	0.499	0.1	-6.2	
cis-1,3-Dichloropropene	0.571	0.561	0.2	-1.75	
1,1,2-Trichloroethane	0.337	0.339	0.1	0.59	
2-Hexanone	0.347	0.321		-7.49	
Dibromochloromethane	0.395	0.392	0.1	-0.76	
1,2-Dibromoethane	0.341	0.337		-1.17	
Tetrachloroethene	0.314	0.316	0.2	0.64	
Chlorobenzene	1.032	0.997	0.5	-3.39	
Ethyl Benzene	1.796	1.679	0.1	-6.51	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: NEWY17

Lab Code: CHEM Case No.: P4997 SAS No.: P4997 SDG No.: P4997

Instrument ID: MSVOA_N Calibration Date/Time: 11/25/2024 10:59

Lab File ID: VN085022.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
m/p-Xylenes	0.687	0.664	0.3	-3.35	
o-Xylene	0.663	0.616	0.3	-7.09	
Styrene	1.112	1.046	0.3	-5.93	
Bromoform	0.265	0.258	0.1	-2.64	
Isopropylbenzene	1.647	1.585		-3.76	
1,1,2,2-Tetrachloroethane	0.534	0.511	0.3	-4.31	
1,3,5-Trimethylbenzene	1.408	1.376		-2.27	
1,3-Dichlorobenzene	0.786	0.751	0.2	-4.45	
1,4-Dichlorobenzene	0.776	0.721	0.2	-7.09	
1,2-Dichlorobenzene	0.774	0.734	0.2	-5.17	
1,2-Dibromo-3-Chloropropane	0.103	0.097		-5.82	
1,2,4-Trichlorobenzene	0.377	0.338	0.2	-10.35	
1,2,3-Trichlorobenzene	0.377	0.338		-10.35	
1,2-Dichloroethane-d4	2.412	2.453	0.01	1.7	
Toluene-d8	1.420	1.409	0.01	-0.77	
4-Bromofluorobenzene	0.517	0.534	0.2	3.29	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	P4997	OrderDate:	11/25/2024 1:27:00 PM
Client:	New York City DEP of Environmental Protection/BWS	Project:	Industrial Wastewater Discharge Permit - Fall 2024
Contact:	Nicholas Prokopowicz	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4997-07	14B-(1-4)-COMP	Water			11/25/24			11/25/24
			Mercury	245.1		12/02/24	12/02/24	
			Metals ICP-Group1	200.7		11/27/24	12/06/24	

Hit Summary Sheet SW-846

SDG No.: P4997 **Order ID:** P4997
Client: New York City DEP of Environmental Protection/BW **Project ID:** Industrial Wastewater Discharge Permit - F

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : 14B-(1-4)-COMP								
P4997-07	14B-(1-4)-COMP	Water	Aluminum	772000		10.6	50.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Arsenic	78.5		2.84	10.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Barium	605		11.3	50.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Cadmium	3.19		0.21	3.00	ug/L
P4997-07	14B-(1-4)-COMP	Water	Calcium	35600		51.8	1000	ug/L
P4997-07	14B-(1-4)-COMP	Water	Chromium	22.3		0.52	5.00	ug/L
P4997-07	14B-(1-4)-COMP	Water	Cobalt	9.42	J	0.88	15.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Copper	591		1.52	10.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Iron	37500		12.4	50.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Lead	27.2		1.57	6.00	ug/L
P4997-07	14B-(1-4)-COMP	Water	Magnesium	8640		61.2	1000	ug/L
P4997-07	14B-(1-4)-COMP	Water	Manganese	22100		0.87	10.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Mercury	1.25		0.022	0.20	ug/L
P4997-07	14B-(1-4)-COMP	Water	Nickel	22.9		1.28	20.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Potassium	3340		281	1000	ug/L
P4997-07	14B-(1-4)-COMP	Water	Selenium	7.02	J	3.07	10.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Silver	2.68	J	0.83	5.00	ug/L
P4997-07	14B-(1-4)-COMP	Water	Sodium	35700		448	1000	ug/L
P4997-07	14B-(1-4)-COMP	Water	Vanadium	36.8		2.40	20.0	ug/L
P4997-07	14B-(1-4)-COMP	Water	Zinc	148		1.44	20.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-(1-4)-COMP	SDG No.:	P4997
Lab Sample ID:	P4997-07	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	772000		1	10.6	50.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-36-0	Antimony	2.96	U	1	2.96	25.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-38-2	Arsenic	78.5		1	2.84	10.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-39-3	Barium	605		1	11.3	50.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-41-7	Beryllium	0.19	U	1	0.19	3.00	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-43-9	Cadmium	3.19		1	0.21	3.00	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-70-2	Calcium	35600		1	51.8	1000	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-47-3	Chromium	22.3		1	0.52	5.00	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-48-4	Cobalt	9.42	J	1	0.88	15.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-50-8	Copper	591	N	1	1.52	10.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7439-89-6	Iron	37500		1	12.4	50.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7439-92-1	Lead	27.2		1	1.57	6.00	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7439-95-4	Magnesium	8640		1	61.2	1000	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7439-96-5	Manganese	22100		1	0.87	10.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7439-97-6	Mercury	1.25	N	1	0.022	0.20	ug/L	12/02/24 11:18	12/02/24 14:07	E245.1	
7440-02-0	Nickel	22.9		1	1.28	20.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-09-7	Potassium	3340		1	281	1000	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7782-49-2	Selenium	7.02	J	1	3.07	10.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-22-4	Silver	2.68	J	1	0.83	5.00	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-23-5	Sodium	35700		1	448	1000	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-28-0	Thallium	2.21	U	1	2.21	20.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-62-2	Vanadium	36.8		1	2.40	20.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	
7440-66-6	Zinc	148		1	1.44	20.0	ug/L	11/27/24 08:40	12/06/24 13:47	EPA 200.7	

Color Before:	Black	Clarity Before:	Clear	Texture:
Color After:	Yellow	Clarity After:	Clear	Artifacts:
Comments:	Metals Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV107	Mercury	3.89	4.0	97	95 - 105	CV	12/02/2024	13:43	LB133684

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV69	Mercury	5.09	5.0	102	90 - 110	CV	12/02/2024	13:48	LB133684
CCV70	Mercury	4.79	5.0	96	90 - 110	CV	12/02/2024	14:18	LB133684
CCV71	Mercury	4.88	5.0	98	90 - 110	CV	12/02/2024	14:28	LB133684

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ **SDG No.:** P4997

Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2570	2500	103	95 - 105	P	12/06/2024	12:14	LB133790
	Antimony	1020	1000	102	95 - 105	P	12/06/2024	12:14	LB133790
	Arsenic	995	1000	100	95 - 105	P	12/06/2024	12:14	LB133790
	Barium	516	520	99	95 - 105	P	12/06/2024	12:14	LB133790
	Beryllium	518	510	102	95 - 105	P	12/06/2024	12:14	LB133790
	Cadmium	503	510	99	95 - 105	P	12/06/2024	12:14	LB133790
	Calcium	10100	10000	101	95 - 105	P	12/06/2024	12:14	LB133790
	Chromium	526	520	101	95 - 105	P	12/06/2024	12:14	LB133790
	Cobalt	512	520	98	95 - 105	P	12/06/2024	12:14	LB133790
	Copper	528	510	104	95 - 105	P	12/06/2024	12:14	LB133790
	Iron	9870	10000	99	95 - 105	P	12/06/2024	12:14	LB133790
	Lead	1000	1000	100	95 - 105	P	12/06/2024	12:14	LB133790
	Magnesium	5930	6000	99	95 - 105	P	12/06/2024	12:14	LB133790
	Manganese	521	520	100	95 - 105	P	12/06/2024	12:14	LB133790
	Nickel	514	530	97	95 - 105	P	12/06/2024	12:14	LB133790
	Potassium	9770	9900	99	95 - 105	P	12/06/2024	12:14	LB133790
	Selenium	1030	1000	103	95 - 105	P	12/06/2024	12:14	LB133790
	Silver	253	250	101	95 - 105	P	12/06/2024	12:14	LB133790
	Sodium	9550	10000	96	95 - 105	P	12/06/2024	12:14	LB133790
	Thallium	1040	1000	104	95 - 105	P	12/06/2024	12:14	LB133790
	Vanadium	513	500	102	95 - 105	P	12/06/2024	12:14	LB133790
	Zinc	1030	1000	103	95 - 105	P	12/06/2024	12:14	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	105	100	105	80 - 120	P	12/06/2024	12:20	LB133790
	Antimony	51.1	50.0	102	80 - 120	P	12/06/2024	12:20	LB133790
	Arsenic	19.8	20.0	99	80 - 120	P	12/06/2024	12:20	LB133790
	Barium	96.9	100	97	80 - 120	P	12/06/2024	12:20	LB133790
	Beryllium	6.20	6.0	103	80 - 120	P	12/06/2024	12:20	LB133790
	Cadmium	5.98	6.0	100	80 - 120	P	12/06/2024	12:20	LB133790
	Calcium	2050	2000	102	80 - 120	P	12/06/2024	12:20	LB133790
	Chromium	10.0	10.0	100	80 - 120	P	12/06/2024	12:20	LB133790
	Cobalt	29.5	30.0	98	80 - 120	P	12/06/2024	12:20	LB133790
	Copper	22.1	20.0	111	80 - 120	P	12/06/2024	12:20	LB133790
	Iron	98.1	100	98	80 - 120	P	12/06/2024	12:20	LB133790
	Lead	10.5	12.0	88	80 - 120	P	12/06/2024	12:20	LB133790
	Magnesium	2140	2000	107	80 - 120	P	12/06/2024	12:20	LB133790
	Manganese	20.6	20.0	103	80 - 120	P	12/06/2024	12:20	LB133790
	Nickel	39.5	40.0	99	80 - 120	P	12/06/2024	12:20	LB133790
	Potassium	1890	2000	95	80 - 120	P	12/06/2024	12:20	LB133790
	Selenium	18.0	20.0	90	80 - 120	P	12/06/2024	12:20	LB133790
	Silver	10.6	10.0	106	80 - 120	P	12/06/2024	12:20	LB133790
	Sodium	1810	2000	90	80 - 120	P	12/06/2024	12:20	LB133790
	Thallium	40.1	40.0	100	80 - 120	P	12/06/2024	12:20	LB133790
	Vanadium	39.8	40.0	100	80 - 120	P	12/06/2024	12:20	LB133790
	Zinc	42.2	40.0	105	80 - 120	P	12/06/2024	12:20	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10000	10000	100	90 - 110	P	12/06/2024	13:09	LB133790
	Antimony	4950	5000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Arsenic	5010	5000	100	90 - 110	P	12/06/2024	13:09	LB133790
	Barium	9800	10000	98	90 - 110	P	12/06/2024	13:09	LB133790
	Beryllium	249	250	100	90 - 110	P	12/06/2024	13:09	LB133790
	Cadmium	2520	2500	101	90 - 110	P	12/06/2024	13:09	LB133790
	Calcium	24700	25000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Chromium	1020	1000	102	90 - 110	P	12/06/2024	13:09	LB133790
	Cobalt	2510	2500	100	90 - 110	P	12/06/2024	13:09	LB133790
	Copper	1280	1250	102	90 - 110	P	12/06/2024	13:09	LB133790
	Iron	4940	5000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Lead	5020	5000	100	90 - 110	P	12/06/2024	13:09	LB133790
	Magnesium	24800	25000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Manganese	2440	2500	98	90 - 110	P	12/06/2024	13:09	LB133790
	Nickel	2510	2500	100	90 - 110	P	12/06/2024	13:09	LB133790
	Potassium	24700	25000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Selenium	5060	5000	101	90 - 110	P	12/06/2024	13:09	LB133790
	Silver	1250	1250	100	90 - 110	P	12/06/2024	13:09	LB133790
	Sodium	24300	25000	97	90 - 110	P	12/06/2024	13:09	LB133790
	Thallium	5050	5000	101	90 - 110	P	12/06/2024	13:09	LB133790
	Vanadium	2500	2500	100	90 - 110	P	12/06/2024	13:09	LB133790
	Zinc	2550	2500	102	90 - 110	P	12/06/2024	13:09	LB133790
CCV02	Aluminum	9810	10000	98	90 - 110	P	12/06/2024	13:38	LB133790
	Antimony	4850	5000	97	90 - 110	P	12/06/2024	13:38	LB133790
	Arsenic	4880	5000	98	90 - 110	P	12/06/2024	13:38	LB133790
	Barium	9400	10000	94	90 - 110	P	12/06/2024	13:38	LB133790
	Beryllium	244	250	98	90 - 110	P	12/06/2024	13:38	LB133790
	Cadmium	2460	2500	98	90 - 110	P	12/06/2024	13:38	LB133790
	Calcium	24200	25000	97	90 - 110	P	12/06/2024	13:38	LB133790
	Chromium	1010	1000	101	90 - 110	P	12/06/2024	13:38	LB133790
	Cobalt	2450	2500	98	90 - 110	P	12/06/2024	13:38	LB133790
	Copper	1250	1250	100	90 - 110	P	12/06/2024	13:38	LB133790
	Iron	4810	5000	96	90 - 110	P	12/06/2024	13:38	LB133790
	Lead	4900	5000	98	90 - 110	P	12/06/2024	13:38	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24400	25000	98	90 - 110	P	12/06/2024	13:38	LB133790
	Manganese	2360	2500	95	90 - 110	P	12/06/2024	13:38	LB133790
	Nickel	2450	2500	98	90 - 110	P	12/06/2024	13:38	LB133790
	Potassium	23900	25000	96	90 - 110	P	12/06/2024	13:38	LB133790
	Selenium	4940	5000	99	90 - 110	P	12/06/2024	13:38	LB133790
	Silver	1230	1250	98	90 - 110	P	12/06/2024	13:38	LB133790
	Sodium	23500	25000	94	90 - 110	P	12/06/2024	13:38	LB133790
	Thallium	4910	5000	98	90 - 110	P	12/06/2024	13:38	LB133790
	Vanadium	2440	2500	98	90 - 110	P	12/06/2024	13:38	LB133790
	Zinc	2500	2500	100	90 - 110	P	12/06/2024	13:38	LB133790
CCV03	Aluminum	9820	10000	98	90 - 110	P	12/06/2024	14:42	LB133790
	Antimony	4780	5000	96	90 - 110	P	12/06/2024	14:42	LB133790
	Arsenic	4820	5000	96	90 - 110	P	12/06/2024	14:42	LB133790
	Barium	9370	10000	94	90 - 110	P	12/06/2024	14:42	LB133790
	Beryllium	245	250	98	90 - 110	P	12/06/2024	14:42	LB133790
	Cadmium	2450	2500	98	90 - 110	P	12/06/2024	14:42	LB133790
	Calcium	24200	25000	97	90 - 110	P	12/06/2024	14:42	LB133790
	Chromium	1000	1000	100	90 - 110	P	12/06/2024	14:42	LB133790
	Cobalt	2430	2500	97	90 - 110	P	12/06/2024	14:42	LB133790
	Copper	1230	1250	99	90 - 110	P	12/06/2024	14:42	LB133790
	Iron	4770	5000	95	90 - 110	P	12/06/2024	14:42	LB133790
	Lead	4880	5000	98	90 - 110	P	12/06/2024	14:42	LB133790
	Magnesium	24300	25000	97	90 - 110	P	12/06/2024	14:42	LB133790
	Manganese	2360	2500	94	90 - 110	P	12/06/2024	14:42	LB133790
	Nickel	2440	2500	98	90 - 110	P	12/06/2024	14:42	LB133790
	Potassium	23300	25000	93	90 - 110	P	12/06/2024	14:42	LB133790
	Selenium	4870	5000	97	90 - 110	P	12/06/2024	14:42	LB133790
	Silver	1210	1250	97	90 - 110	P	12/06/2024	14:42	LB133790
	Sodium	24900	25000	100	90 - 110	P	12/06/2024	14:42	LB133790
	Thallium	4850	5000	97	90 - 110	P	12/06/2024	14:42	LB133790
CCV04	Vanadium	2450	2500	98	90 - 110	P	12/06/2024	14:42	LB133790
	Zinc	2470	2500	99	90 - 110	P	12/06/2024	14:42	LB133790
	Aluminum	9820	10000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Antimony	4780	5000	96	90 - 110	P	12/06/2024	15:40	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
 Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4870	5000	97	90 - 110	P	12/06/2024	15:40	LB133790
	Barium	9560	10000	96	90 - 110	P	12/06/2024	15:40	LB133790
	Beryllium	250	250	100	90 - 110	P	12/06/2024	15:40	LB133790
	Cadmium	2500	2500	100	90 - 110	P	12/06/2024	15:40	LB133790
	Calcium	24600	25000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Chromium	1020	1000	102	90 - 110	P	12/06/2024	15:40	LB133790
	Cobalt	2480	2500	99	90 - 110	P	12/06/2024	15:40	LB133790
	Copper	1240	1250	100	90 - 110	P	12/06/2024	15:40	LB133790
	Iron	4880	5000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Lead	4960	5000	99	90 - 110	P	12/06/2024	15:40	LB133790
	Magnesium	24800	25000	99	90 - 110	P	12/06/2024	15:40	LB133790
	Manganese	2420	2500	97	90 - 110	P	12/06/2024	15:40	LB133790
	Nickel	2480	2500	99	90 - 110	P	12/06/2024	15:40	LB133790
	Potassium	23500	25000	94	90 - 110	P	12/06/2024	15:40	LB133790
	Selenium	4910	5000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Silver	1230	1250	98	90 - 110	P	12/06/2024	15:40	LB133790
	Sodium	25600	25000	102	90 - 110	P	12/06/2024	15:40	LB133790
	Thallium	4910	5000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Vanadium	2480	2500	99	90 - 110	P	12/06/2024	15:40	LB133790
	Zinc	2430	2500	97	90 - 110	P	12/06/2024	15:40	LB133790
CCV05	Aluminum	9800	10000	98	90 - 110	P	12/06/2024	16:35	LB133790
	Antimony	4730	5000	95	90 - 110	P	12/06/2024	16:35	LB133790
	Arsenic	4810	5000	96	90 - 110	P	12/06/2024	16:35	LB133790
	Barium	9510	10000	95	90 - 110	P	12/06/2024	16:35	LB133790
	Beryllium	250	250	100	90 - 110	P	12/06/2024	16:35	LB133790
	Cadmium	2460	2500	98	90 - 110	P	12/06/2024	16:35	LB133790
	Calcium	24500	25000	98	90 - 110	P	12/06/2024	16:35	LB133790
	Chromium	1010	1000	100	90 - 110	P	12/06/2024	16:35	LB133790
	Cobalt	2440	2500	98	90 - 110	P	12/06/2024	16:35	LB133790
	Copper	1230	1250	99	90 - 110	P	12/06/2024	16:35	LB133790
	Iron	4850	5000	97	90 - 110	P	12/06/2024	16:35	LB133790
	Lead	4890	5000	98	90 - 110	P	12/06/2024	16:35	LB133790
	Magnesium	24500	25000	98	90 - 110	P	12/06/2024	16:35	LB133790
	Manganese	2420	2500	97	90 - 110	P	12/06/2024	16:35	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	12/06/2024	16:35	LB133790
	Potassium	23200	25000	93	90 - 110	P	12/06/2024	16:35	LB133790
	Selenium	4850	5000	97	90 - 110	P	12/06/2024	16:35	LB133790
	Silver	1220	1250	98	90 - 110	P	12/06/2024	16:35	LB133790
	Sodium	25300	25000	101	90 - 110	P	12/06/2024	16:35	LB133790
	Thallium	4820	5000	96	90 - 110	P	12/06/2024	16:35	LB133790
	Vanadium	2470	2500	99	90 - 110	P	12/06/2024	16:35	LB133790
	Zinc	2380	2500	95	90 - 110	P	12/06/2024	16:35	LB133790
CCV06	Aluminum	9980	10000	100	90 - 110	P	12/06/2024	17:27	LB133790
	Antimony	4840	5000	97	90 - 110	P	12/06/2024	17:27	LB133790
	Arsenic	4950	5000	99	90 - 110	P	12/06/2024	17:27	LB133790
	Barium	9680	10000	97	90 - 110	P	12/06/2024	17:27	LB133790
	Beryllium	253	250	101	90 - 110	P	12/06/2024	17:27	LB133790
	Cadmium	2540	2500	102	90 - 110	P	12/06/2024	17:27	LB133790
	Calcium	24900	25000	100	90 - 110	P	12/06/2024	17:27	LB133790
	Chromium	1030	1000	103	90 - 110	P	12/06/2024	17:27	LB133790
	Cobalt	2510	2500	100	90 - 110	P	12/06/2024	17:27	LB133790
	Copper	1260	1250	101	90 - 110	P	12/06/2024	17:27	LB133790
	Iron	4910	5000	98	90 - 110	P	12/06/2024	17:27	LB133790
	Lead	5020	5000	100	90 - 110	P	12/06/2024	17:27	LB133790
	Magnesium	25000	25000	100	90 - 110	P	12/06/2024	17:27	LB133790
	Manganese	2440	2500	98	90 - 110	P	12/06/2024	17:27	LB133790
	Nickel	2520	2500	101	90 - 110	P	12/06/2024	17:27	LB133790
	Potassium	23700	25000	95	90 - 110	P	12/06/2024	17:27	LB133790
	Selenium	5000	5000	100	90 - 110	P	12/06/2024	17:27	LB133790
	Silver	1250	1250	100	90 - 110	P	12/06/2024	17:27	LB133790
	Sodium	22500	25000	90	90 - 110	P	12/06/2024	17:27	LB133790
	Thallium	4930	5000	99	90 - 110	P	12/06/2024	17:27	LB133790
	Vanadium	2520	2500	101	90 - 110	P	12/06/2024	17:27	LB133790
	Zinc	2480	2500	99	90 - 110	P	12/06/2024	17:27	LB133790
CCV07	Aluminum	9730	10000	97	90 - 110	P	12/06/2024	18:21	LB133790
	Antimony	4710	5000	94	90 - 110	P	12/06/2024	18:21	LB133790
	Arsenic	4810	5000	96	90 - 110	P	12/06/2024	18:21	LB133790
	Barium	9300	10000	93	90 - 110	P	12/06/2024	18:21	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ **SDG No.:** P4997

Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	251	250	101	90 - 110	P	12/06/2024	18:21	LB133790
	Cadmium	2490	2500	100	90 - 110	P	12/06/2024	18:21	LB133790
	Calcium	24400	25000	98	90 - 110	P	12/06/2024	18:21	LB133790
	Chromium	1010	1000	101	90 - 110	P	12/06/2024	18:21	LB133790
	Cobalt	2470	2500	99	90 - 110	P	12/06/2024	18:21	LB133790
	Copper	1230	1250	99	90 - 110	P	12/06/2024	18:21	LB133790
	Iron	4760	5000	95	90 - 110	P	12/06/2024	18:21	LB133790
	Lead	4930	5000	99	90 - 110	P	12/06/2024	18:21	LB133790
	Magnesium	24500	25000	98	90 - 110	P	12/06/2024	18:21	LB133790
	Manganese	2400	2500	96	90 - 110	P	12/06/2024	18:21	LB133790
	Nickel	2470	2500	99	90 - 110	P	12/06/2024	18:21	LB133790
	Potassium	22800	25000	91	90 - 110	P	12/06/2024	18:21	LB133790
	Selenium	4860	5000	97	90 - 110	P	12/06/2024	18:21	LB133790
	Silver	1220	1250	98	90 - 110	P	12/06/2024	18:21	LB133790
	Sodium	26000	25000	104	90 - 110	P	12/06/2024	18:21	LB133790
	Thallium	4870	5000	97	90 - 110	P	12/06/2024	18:21	LB133790
	Vanadium	2460	2500	98	90 - 110	P	12/06/2024	18:21	LB133790
	Zinc	2440	2500	98	90 - 110	P	12/06/2024	18:21	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
 Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2490	2500	100	95 - 105	P	12/09/2024	11:35	LB133834
	Antimony	980	1000	98	95 - 105	P	12/09/2024	11:35	LB133834
	Arsenic	955	1000	96	95 - 105	P	12/09/2024	11:35	LB133834
	Barium	508	520	98	95 - 105	P	12/09/2024	11:35	LB133834
	Beryllium	525	510	103	95 - 105	P	12/09/2024	11:35	LB133834
	Cadmium	494	510	97	95 - 105	P	12/09/2024	11:35	LB133834
	Calcium	10100	10000	101	95 - 105	P	12/09/2024	11:35	LB133834
	Chromium	517	520	99	95 - 105	P	12/09/2024	11:35	LB133834
	Cobalt	500	520	96	95 - 105	P	12/09/2024	11:35	LB133834
	Copper	508	510	100	95 - 105	P	12/09/2024	11:35	LB133834
	Iron	9630	10000	96	95 - 105	P	12/09/2024	11:35	LB133834
	Lead	971	1000	97	95 - 105	P	12/09/2024	11:35	LB133834
	Magnesium	5940	6000	99	95 - 105	P	12/09/2024	11:35	LB133834
	Manganese	520	520	100	95 - 105	P	12/09/2024	11:35	LB133834
	Nickel	550	530	104	95 - 105	P	12/09/2024	11:35	LB133834
	Potassium	10200	9900	103	95 - 105	P	12/09/2024	11:35	LB133834
	Selenium	975	1000	98	95 - 105	P	12/09/2024	11:35	LB133834
	Silver	248	250	99	95 - 105	P	12/09/2024	11:35	LB133834
	Sodium	10400	10000	104	95 - 105	P	12/09/2024	11:35	LB133834
	Thallium	1010	1000	101	95 - 105	P	12/09/2024	11:35	LB133834
	Vanadium	506	500	101	95 - 105	P	12/09/2024	11:35	LB133834
	Zinc	999	1000	100	95 - 105	P	12/09/2024	11:35	LB133834

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
 Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	107	100	107	80 - 120	P	12/09/2024	11:49	LB133834
	Antimony	50.9	50.0	102	80 - 120	P	12/09/2024	11:49	LB133834
	Arsenic	18.7	20.0	94	80 - 120	P	12/09/2024	11:49	LB133834
	Barium	99.3	100	99	80 - 120	P	12/09/2024	11:49	LB133834
	Beryllium	6.16	6.0	103	80 - 120	P	12/09/2024	11:49	LB133834
	Cadmium	5.91	6.0	98	80 - 120	P	12/09/2024	11:49	LB133834
	Calcium	2130	2000	106	80 - 120	P	12/09/2024	11:49	LB133834
	Chromium	11.3	10.0	113	80 - 120	P	12/09/2024	11:49	LB133834
	Cobalt	29.4	30.0	98	80 - 120	P	12/09/2024	11:49	LB133834
	Copper	22.3	20.0	111	80 - 120	P	12/09/2024	11:49	LB133834
	Iron	93.5	100	94	80 - 120	P	12/09/2024	11:49	LB133834
	Lead	12.3	12.0	102	80 - 120	P	12/09/2024	11:49	LB133834
	Magnesium	2250	2000	112	80 - 120	P	12/09/2024	11:49	LB133834
	Manganese	21.5	20.0	107	80 - 120	P	12/09/2024	11:49	LB133834
	Nickel	39.5	40.0	99	80 - 120	P	12/09/2024	11:49	LB133834
	Potassium	1830	2000	92	80 - 120	P	12/09/2024	11:49	LB133834
	Selenium	21.1	20.0	106	80 - 120	P	12/09/2024	11:49	LB133834
	Silver	9.94	10.0	99	80 - 120	P	12/09/2024	11:49	LB133834
	Sodium	1770	2000	89	80 - 120	P	12/09/2024	11:49	LB133834
	Thallium	39.5	40.0	99	80 - 120	P	12/09/2024	11:49	LB133834
	Vanadium	42.1	40.0	105	80 - 120	P	12/09/2024	11:49	LB133834
	Zinc	42.5	40.0	106	80 - 120	P	12/09/2024	11:49	LB133834

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
 Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9760	10000	98	90 - 110	P	12/09/2024	12:29	LB133834
	Antimony	4890	5000	98	90 - 110	P	12/09/2024	12:29	LB133834
	Arsenic	4840	5000	97	90 - 110	P	12/09/2024	12:29	LB133834
	Barium	9640	10000	96	90 - 110	P	12/09/2024	12:29	LB133834
	Beryllium	249	250	100	90 - 110	P	12/09/2024	12:29	LB133834
	Cadmium	2430	2500	97	90 - 110	P	12/09/2024	12:29	LB133834
	Calcium	24300	25000	97	90 - 110	P	12/09/2024	12:29	LB133834
	Chromium	993	1000	99	90 - 110	P	12/09/2024	12:29	LB133834
	Cobalt	2420	2500	97	90 - 110	P	12/09/2024	12:29	LB133834
	Copper	1220	1250	98	90 - 110	P	12/09/2024	12:29	LB133834
	Iron	4760	5000	95	90 - 110	P	12/09/2024	12:29	LB133834
	Lead	4850	5000	97	90 - 110	P	12/09/2024	12:29	LB133834
	Magnesium	24200	25000	97	90 - 110	P	12/09/2024	12:29	LB133834
	Manganese	2390	2500	96	90 - 110	P	12/09/2024	12:29	LB133834
	Nickel	2420	2500	97	90 - 110	P	12/09/2024	12:29	LB133834
	Potassium	24000	25000	96	90 - 110	P	12/09/2024	12:29	LB133834
	Selenium	4880	5000	98	90 - 110	P	12/09/2024	12:29	LB133834
	Silver	1230	1250	98	90 - 110	P	12/09/2024	12:29	LB133834
	Sodium	24000	25000	96	90 - 110	P	12/09/2024	12:29	LB133834
	Thallium	5080	5000	102	90 - 110	P	12/09/2024	12:29	LB133834
	Vanadium	2450	2500	98	90 - 110	P	12/09/2024	12:29	LB133834
	Zinc	2480	2500	99	90 - 110	P	12/09/2024	12:29	LB133834
CCV02	Aluminum	9930	10000	99	90 - 110	P	12/09/2024	13:30	LB133834
	Antimony	5210	5000	104	90 - 110	P	12/09/2024	13:30	LB133834
	Arsenic	5090	5000	102	90 - 110	P	12/09/2024	13:30	LB133834
	Barium	9940	10000	99	90 - 110	P	12/09/2024	13:30	LB133834
	Beryllium	234	250	93	90 - 110	P	12/09/2024	13:30	LB133834
	Cadmium	2430	2500	97	90 - 110	P	12/09/2024	13:30	LB133834
	Calcium	23600	25000	94	90 - 110	P	12/09/2024	13:30	LB133834
	Chromium	984	1000	98	90 - 110	P	12/09/2024	13:30	LB133834
	Cobalt	2430	2500	97	90 - 110	P	12/09/2024	13:30	LB133834
	Copper	1280	1250	103	90 - 110	P	12/09/2024	13:30	LB133834
	Iron	4890	5000	98	90 - 110	P	12/09/2024	13:30	LB133834
	Lead	4850	5000	97	90 - 110	P	12/09/2024	13:30	LB133834

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23400	25000	94	90 - 110	P	12/09/2024	13:30	LB133834
	Manganese	2350	2500	94	90 - 110	P	12/09/2024	13:30	LB133834
	Nickel	2430	2500	97	90 - 110	P	12/09/2024	13:30	LB133834
	Potassium	25800	25000	103	90 - 110	P	12/09/2024	13:30	LB133834
	Selenium	5240	5000	105	90 - 110	P	12/09/2024	13:30	LB133834
	Silver	1240	1250	100	90 - 110	P	12/09/2024	13:30	LB133834
	Sodium	25600	25000	102	90 - 110	P	12/09/2024	13:30	LB133834
	Thallium	5210	5000	104	90 - 110	P	12/09/2024	13:30	LB133834
	Vanadium	2430	2500	97	90 - 110	P	12/09/2024	13:30	LB133834
	Zinc	2540	2500	102	90 - 110	P	12/09/2024	13:30	LB133834
CCV03	Aluminum	9740	10000	97	90 - 110	P	12/09/2024	14:23	LB133834
	Antimony	5060	5000	101	90 - 110	P	12/09/2024	14:23	LB133834
	Arsenic	4930	5000	99	90 - 110	P	12/09/2024	14:23	LB133834
	Barium	9920	10000	99	90 - 110	P	12/09/2024	14:23	LB133834
	Beryllium	233	250	93	90 - 110	P	12/09/2024	14:23	LB133834
	Cadmium	2420	2500	97	90 - 110	P	12/09/2024	14:23	LB133834
	Calcium	23500	25000	94	90 - 110	P	12/09/2024	14:23	LB133834
	Chromium	967	1000	97	90 - 110	P	12/09/2024	14:23	LB133834
	Cobalt	2410	2500	96	90 - 110	P	12/09/2024	14:23	LB133834
	Copper	1260	1250	101	90 - 110	P	12/09/2024	14:23	LB133834
	Iron	4870	5000	97	90 - 110	P	12/09/2024	14:23	LB133834
	Lead	4790	5000	96	90 - 110	P	12/09/2024	14:23	LB133834
	Magnesium	23200	25000	93	90 - 110	P	12/09/2024	14:23	LB133834
	Manganese	2360	2500	94	90 - 110	P	12/09/2024	14:23	LB133834
	Nickel	2410	2500	97	90 - 110	P	12/09/2024	14:23	LB133834
	Potassium	25400	25000	102	90 - 110	P	12/09/2024	14:23	LB133834
	Selenium	5010	5000	100	90 - 110	P	12/09/2024	14:23	LB133834
	Silver	1220	1250	98	90 - 110	P	12/09/2024	14:23	LB133834
	Sodium	25900	25000	104	90 - 110	P	12/09/2024	14:23	LB133834
	Thallium	4900	5000	98	90 - 110	P	12/09/2024	14:23	LB133834
	Vanadium	2420	2500	97	90 - 110	P	12/09/2024	14:23	LB133834
	Zinc	2500	2500	100	90 - 110	P	12/09/2024	14:23	LB133834
CCV04	Aluminum	9580	10000	96	90 - 110	P	12/09/2024	15:13	LB133834
	Antimony	4940	5000	99	90 - 110	P	12/09/2024	15:13	LB133834

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ SDG No.: P4997
Contract: NEWY17 Lab Code: CHEM Case No.: P4997 SAS No.: P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4850	5000	97	90 - 110	P	12/09/2024	15:13	LB133834
	Barium	9630	10000	96	90 - 110	P	12/09/2024	15:13	LB133834
	Beryllium	234	250	94	90 - 110	P	12/09/2024	15:13	LB133834
	Cadmium	2370	2500	95	90 - 110	P	12/09/2024	15:13	LB133834
	Calcium	23200	25000	93	90 - 110	P	12/09/2024	15:13	LB133834
	Chromium	952	1000	95	90 - 110	P	12/09/2024	15:13	LB133834
	Cobalt	2360	2500	94	90 - 110	P	12/09/2024	15:13	LB133834
	Copper	1220	1250	98	90 - 110	P	12/09/2024	15:13	LB133834
	Iron	4720	5000	94	90 - 110	P	12/09/2024	15:13	LB133834
	Lead	4720	5000	94	90 - 110	P	12/09/2024	15:13	LB133834
	Magnesium	23000	25000	92	90 - 110	P	12/09/2024	15:13	LB133834
	Manganese	2320	2500	93	90 - 110	P	12/09/2024	15:13	LB133834
	Nickel	2370	2500	95	90 - 110	P	12/09/2024	15:13	LB133834
	Potassium	24300	25000	97	90 - 110	P	12/09/2024	15:13	LB133834
	Selenium	4940	5000	99	90 - 110	P	12/09/2024	15:13	LB133834
	Silver	1200	1250	96	90 - 110	P	12/09/2024	15:13	LB133834
	Sodium	23800	25000	95	90 - 110	P	12/09/2024	15:13	LB133834
	Thallium	5090	5000	102	90 - 110	P	12/09/2024	15:13	LB133834
	Vanadium	2360	2500	95	90 - 110	P	12/09/2024	15:13	LB133834
	Zinc	2410	2500	97	90 - 110	P	12/09/2024	15:13	LB133834
CCV05	Aluminum	9430	10000	94	90 - 110	P	12/09/2024	15:52	LB133834
	Antimony	4930	5000	98	90 - 110	P	12/09/2024	15:52	LB133834
	Arsenic	4810	5000	96	90 - 110	P	12/09/2024	15:52	LB133834
	Barium	9700	10000	97	90 - 110	P	12/09/2024	15:52	LB133834
	Beryllium	229	250	92	90 - 110	P	12/09/2024	15:52	LB133834
	Cadmium	2320	2500	93	90 - 110	P	12/09/2024	15:52	LB133834
	Calcium	22700	25000	91	90 - 110	P	12/09/2024	15:52	LB133834
	Chromium	925	1000	92	90 - 110	P	12/09/2024	15:52	LB133834
	Cobalt	2320	2500	93	90 - 110	P	12/09/2024	15:52	LB133834
	Copper	1220	1250	98	90 - 110	P	12/09/2024	15:52	LB133834
	Iron	4670	5000	93	90 - 110	P	12/09/2024	15:52	LB133834
	Lead	4610	5000	92	90 - 110	P	12/09/2024	15:52	LB133834
	Magnesium	24400	25000	98	90 - 110	P	12/09/2024	15:52	LB133834
	Manganese	2300	2500	92	90 - 110	P	12/09/2024	15:52	LB133834

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: New York City DEP of Environmental Protection/ **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2320	2500	93	90 - 110	P	12/09/2024	15:52	LB133834
	Potassium	24400	25000	98	90 - 110	P	12/09/2024	15:52	LB133834
	Selenium	4900	5000	98	90 - 110	P	12/09/2024	15:52	LB133834
	Silver	1180	1250	94	90 - 110	P	12/09/2024	15:52	LB133834
	Sodium	23800	25000	95	90 - 110	P	12/09/2024	15:52	LB133834
	Thallium	4870	5000	97	90 - 110	P	12/09/2024	15:52	LB133834
	Vanadium	2330	2500	93	90 - 110	P	12/09/2024	15:52	LB133834
	Zinc	2350	2500	94	90 - 110	P	12/09/2024	15:52	LB133834

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.045	0.05	90	40 - 160	CV	12/02/2024	13:56	LB133684
CRI01	Aluminum	105	100	104	40 - 160	P	12/06/2024	12:28	LB133790
	Antimony	50.1	50.0	100	40 - 160	P	12/06/2024	12:28	LB133790
	Arsenic	20.2	20.0	101	40 - 160	P	12/06/2024	12:28	LB133790
	Barium	97.5	100	98	40 - 160	P	12/06/2024	12:28	LB133790
	Beryllium	6.20	6.0	103	40 - 160	P	12/06/2024	12:28	LB133790
	Cadmium	5.89	6.0	98	40 - 160	P	12/06/2024	12:28	LB133790
	Calcium	2060	2000	103	40 - 160	P	12/06/2024	12:28	LB133790
	Chromium	10.0	10.0	100	40 - 160	P	12/06/2024	12:28	LB133790
	Cobalt	29.5	30.0	98	40 - 160	P	12/06/2024	12:28	LB133790
	Copper	22.0	20.0	110	40 - 160	P	12/06/2024	12:28	LB133790
	Iron	96.6	100	97	40 - 160	P	12/06/2024	12:28	LB133790
	Lead	11.1	12.0	93	40 - 160	P	12/06/2024	12:28	LB133790
	Magnesium	2160	2000	108	40 - 160	P	12/06/2024	12:28	LB133790
	Manganese	21.0	20.0	105	40 - 160	P	12/06/2024	12:28	LB133790
	Nickel	40.0	40.0	100	40 - 160	P	12/06/2024	12:28	LB133790
	Potassium	1910	2000	96	40 - 160	P	12/06/2024	12:28	LB133790
	Selenium	19.8	20.0	99	40 - 160	P	12/06/2024	12:28	LB133790
	Silver	10.3	10.0	103	40 - 160	P	12/06/2024	12:28	LB133790
	Sodium	1810	2000	90	40 - 160	P	12/06/2024	12:28	LB133790
	Thallium	41.4	40.0	104	40 - 160	P	12/06/2024	12:28	LB133790
	Vanadium	40.3	40.0	101	40 - 160	P	12/06/2024	12:28	LB133790
	Zinc	42.4	40.0	106	40 - 160	P	12/06/2024	12:28	LB133790
CRI01	Aluminum	105	100	104	40 - 160	P	12/09/2024	11:59	LB133834
	Antimony	50.2	50.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Arsenic	20.0	20.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Barium	96.4	100	96	40 - 160	P	12/09/2024	11:59	LB133834
	Beryllium	6.07	6.0	101	40 - 160	P	12/09/2024	11:59	LB133834
	Cadmium	5.65	6.0	94	40 - 160	P	12/09/2024	11:59	LB133834
	Calcium	2060	2000	103	40 - 160	P	12/09/2024	11:59	LB133834
	Chromium	9.89	10.0	99	40 - 160	P	12/09/2024	11:59	LB133834
	Cobalt	28.6	30.0	95	40 - 160	P	12/09/2024	11:59	LB133834
	Copper	21.8	20.0	109	40 - 160	P	12/09/2024	11:59	LB133834
	Iron	93.6	100	94	40 - 160	P	12/09/2024	11:59	LB133834
	Lead	11.0	12.0	91	40 - 160	P	12/09/2024	11:59	LB133834

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2160	2000	108	40 - 160	P	12/09/2024	11:59	LB133834
	Manganese	20.5	20.0	103	40 - 160	P	12/09/2024	11:59	LB133834
	Nickel	38.6	40.0	96	40 - 160	P	12/09/2024	11:59	LB133834
	Potassium	1910	2000	96	40 - 160	P	12/09/2024	11:59	LB133834
	Selenium	18.4	20.0	92	40 - 160	P	12/09/2024	11:59	LB133834
	Silver	9.73	10.0	97	40 - 160	P	12/09/2024	11:59	LB133834
	Sodium	1800	2000	90	40 - 160	P	12/09/2024	11:59	LB133834
	Thallium	40.0	40.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Vanadium	40.0	40.0	100	40 - 160	P	12/09/2024	11:59	LB133834
	Zinc	42.2	40.0	106	40 - 160	P	12/09/2024	11:59	LB133834



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:

New York City DEP of Environmental Protection/B

SDG No.:

P4997

Contract:

NEWY17

Lab Code:

CHEM

Case No.:

P4997

SAS No.:

P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB107	Mercury	0.20	+/-0.20	U	0.20	CV	12/02/2024	13:45	LB133684

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB69	Mercury	0.20	+/-0.20	U	0.20	CV	12/02/2024	13:50	LB133684
CCB70	Mercury	0.20	+/-0.20	U	0.20	CV	12/02/2024	14:21	LB133684
CCB71	Mercury	0.20	+/-0.20	U	0.20	CV	12/02/2024	14:30	LB133684

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	12/06/2024	12:24	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	12:24	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	12:24	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	12:24	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	12:24	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	12:24	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	12:24	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	12:24	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	12:24	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	12:24	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	12:24	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	12:24	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	12:24	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	12:24	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	12:24	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	12:24	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	12:24	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	12:24	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	12:24	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	12:24	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	12:24	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	12:24	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	12/06/2024	13:13	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	13:13	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	13:13	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	13:13	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	13:13	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	13:13	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	13:13	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	13:13	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	13:13	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	13:13	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	13:13	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	13:13	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	13:13	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	13:13	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	13:13	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	13:13	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	13:13	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	13:13	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	13:13	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	13:13	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	13:13	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	13:13	LB133790
CCB02	Aluminum	100	+/-100	U	100	P	12/06/2024	13:42	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	13:42	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	13:42	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	13:42	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	13:42	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	13:42	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	13:42	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	13:42	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	13:42	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	13:42	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	13:42	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	13:42	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	13:42	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	13:42	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	13:42	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	13:42	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	13:42	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	13:42	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	13:42	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	13:42	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	13:42	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	13:42	LB133790
CCB03	Aluminum	100	+/-100	U	100	P	12/06/2024	14:46	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	14:46	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	14:46	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	14:46	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	14:46	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	14:46	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	14:46	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	14:46	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	14:46	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	14:46	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	14:46	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	14:46	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	14:46	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	14:46	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	14:46	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	14:46	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	14:46	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	14:46	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	14:46	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	14:46	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	14:46	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	14:46	LB133790
CCB04	Aluminum	100	+/-100	U	100	P	12/06/2024	15:44	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	15:44	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	15:44	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	15:44	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	15:44	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	15:44	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	15:44	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	15:44	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	15:44	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	15:44	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	15:44	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	15:44	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	15:44	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	15:44	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	15:44	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	15:44	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	15:44	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	15:44	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	15:44	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	15:44	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	15:44	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	15:44	LB133790
CCB05	Aluminum	100	+/-100	U	100	P	12/06/2024	16:39	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	16:39	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	16:39	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	16:39	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	16:39	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	16:39	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	16:39	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	16:39	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	16:39	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	16:39	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	16:39	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	16:39	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	16:39	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	16:39	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	16:39	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	16:39	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	16:39	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	16:39	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	16:39	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	16:39	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	16:39	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	16:39	LB133790
CCB06	Aluminum	100	+/-100	U	100	P	12/06/2024	17:31	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	17:31	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	17:31	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	17:31	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	17:31	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	17:31	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	17:31	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	17:31	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	17:31	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	17:31	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	17:31	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	17:31	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	17:31	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	17:31	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	17:31	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	17:31	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	17:31	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	17:31	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	17:31	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	17:31	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	17:31	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	17:31	LB133790
CCB07	Aluminum	100	+/-100	U	100	P	12/06/2024	18:27	LB133790
	Antimony	50.0	+/-50.0	U	50.0	P	12/06/2024	18:27	LB133790
	Arsenic	20.0	+/-20.0	U	20.0	P	12/06/2024	18:27	LB133790
	Barium	100	+/-100	U	100	P	12/06/2024	18:27	LB133790
	Beryllium	6.00	+/-6.00	U	6.00	P	12/06/2024	18:27	LB133790
	Cadmium	6.00	+/-6.00	U	6.00	P	12/06/2024	18:27	LB133790
	Calcium	2000	+/-2000	U	2000	P	12/06/2024	18:27	LB133790
	Chromium	10.0	+/-10.0	U	10.0	P	12/06/2024	18:27	LB133790
	Cobalt	30.0	+/-30.0	U	30.0	P	12/06/2024	18:27	LB133790
	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	18:27	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	18:27	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	18:27	LB133790
	Magnesium	2000	+/-2000	U	2000	P	12/06/2024	18:27	LB133790
	Manganese	20.0	+/-20.0	U	20.0	P	12/06/2024	18:27	LB133790
	Nickel	40.0	+/-40.0	U	40.0	P	12/06/2024	18:27	LB133790
	Potassium	2000	+/-2000	U	2000	P	12/06/2024	18:27	LB133790
	Selenium	20.0	+/-20.0	U	20.0	P	12/06/2024	18:27	LB133790
	Silver	10.0	+/-10.0	U	10.0	P	12/06/2024	18:27	LB133790
	Sodium	2000	+/-2000	U	2000	P	12/06/2024	18:27	LB133790
	Thallium	40.0	+/-40.0	U	40.0	P	12/06/2024	18:27	LB133790
	Vanadium	40.0	+/-40.0	U	40.0	P	12/06/2024	18:27	LB133790
	Zinc	40.0	+/-40.0	U	40.0	P	12/06/2024	18:27	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	12/09/2024	11:54	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	11:54	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	11:54	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	11:54	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	11:54	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	11:54	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	11:54	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	11:54	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	11:54	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	11:54	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	11:54	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	11:54	LB133834
	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	11:54	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	11:54	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	11:54	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	11:54	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	11:54	LB133834
	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	11:54	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	11:54	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	11:54	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	11:54	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	11:54	LB133834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	12/09/2024	12:34	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	12:34	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	12:34	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	12:34	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	12:34	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	12:34	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	12:34	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	12:34	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	12:34	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	12:34	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	12:34	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	12:34	LB133834
	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	12:34	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	12:34	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	12:34	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	12:34	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	12:34	LB133834
	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	12:34	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	12:34	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	12:34	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	12:34	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	12:34	LB133834
CCB02	Aluminum	100	+/-100	U	100	P	12/09/2024	13:35	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	13:35	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	13:35	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	13:35	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	13:35	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	13:35	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	13:35	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	13:35	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	13:35	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	13:35	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	13:35	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	13:35	LB133834
	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	13:35	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	13:35	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	13:35	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	13:35	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	13:35	LB133834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	13:35	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	13:35	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	13:35	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	13:35	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	13:35	LB133834
CCB03	Aluminum	100	+/-100	U	100	P	12/09/2024	14:27	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	14:27	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	14:27	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	14:27	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	14:27	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	14:27	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	14:27	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	14:27	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	14:27	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	14:27	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	14:27	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	14:27	LB133834
	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	14:27	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	14:27	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	14:27	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	14:27	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	14:27	LB133834
	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	14:27	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	14:27	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	14:27	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	14:27	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	14:27	LB133834
CCB04	Aluminum	100	+/-100	U	100	P	12/09/2024	15:17	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	15:17	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	15:17	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	15:17	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	15:17	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	15:17	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	15:17	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	15:17	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	15:17	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	15:17	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	15:17	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	15:17	LB133834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: New York City DEP of Environmental Protection/B

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	15:17	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	15:17	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	15:17	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	15:17	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	15:17	LB133834
	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	15:17	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	15:17	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	15:17	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	15:17	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	15:17	LB133834
CCB05	Aluminum	100	+/-100	U	100	P	12/09/2024	15:56	LB133834
	Antimony	50.0	+/-50.0	U	50.0	P	12/09/2024	15:56	LB133834
	Arsenic	20.0	+/-20.0	U	20.0	P	12/09/2024	15:56	LB133834
	Barium	100	+/-100	U	100	P	12/09/2024	15:56	LB133834
	Beryllium	6.00	+/-6.00	U	6.00	P	12/09/2024	15:56	LB133834
	Cadmium	6.00	+/-6.00	U	6.00	P	12/09/2024	15:56	LB133834
	Calcium	2000	+/-2000	U	2000	P	12/09/2024	15:56	LB133834
	Chromium	10.0	+/-10.0	U	10.0	P	12/09/2024	15:56	LB133834
	Cobalt	30.0	+/-30.0	U	30.0	P	12/09/2024	15:56	LB133834
	Copper	20.0	+/-20.0	U	20.0	P	12/09/2024	15:56	LB133834
	Iron	100	+/-100	U	100	P	12/09/2024	15:56	LB133834
	Lead	12.0	+/-12.0	U	12.0	P	12/09/2024	15:56	LB133834
	Magnesium	2000	+/-2000	U	2000	P	12/09/2024	15:56	LB133834
	Manganese	20.0	+/-20.0	U	20.0	P	12/09/2024	15:56	LB133834
	Nickel	40.0	+/-40.0	U	40.0	P	12/09/2024	15:56	LB133834
	Potassium	2000	+/-2000	U	2000	P	12/09/2024	15:56	LB133834
	Selenium	20.0	+/-20.0	U	20.0	P	12/09/2024	15:56	LB133834
	Silver	10.0	+/-10.0	U	10.0	P	12/09/2024	15:56	LB133834
	Sodium	2000	+/-2000	U	2000	P	12/09/2024	15:56	LB133834
	Thallium	40.0	+/-40.0	U	40.0	P	12/09/2024	15:56	LB133834
	Vanadium	40.0	+/-40.0	U	40.0	P	12/09/2024	15:56	LB133834
	Zinc	40.0	+/-40.0	U	40.0	P	12/09/2024	15:56	LB133834

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: New York City DEP of Environmental Protecti **SDG No.:** P4997

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165329BL		WATER		Batch Number:	PB165329		Prep Date:	12/02/2024	
	Mercury	0.20	<0.20	U	0.20	CV	12/02/2024	14:02	LB133684

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: New York City DEP of Environmental Protecti

SDG No.: P4997

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165287BL	WATER			Batch Number:	PB165287		Prep Date:	11/27/2024	
	Aluminum	50.0	<50.0	U	50.0	P	12/09/2024	12:46	LB133834
	Antimony	25.0	<25.0	U	25.0	P	12/09/2024	12:46	LB133834
	Arsenic	10.0	<10.0	U	10.0	P	12/09/2024	12:46	LB133834
	Barium	50.0	<50.0	U	50.0	P	12/09/2024	12:46	LB133834
	Beryllium	3.00	<3.00	U	3.00	P	12/09/2024	12:46	LB133834
	Cadmium	3.00	<3.00	U	3.00	P	12/09/2024	12:46	LB133834
	Calcium	1000	<1000	U	1000	P	12/09/2024	12:46	LB133834
	Chromium	5.00	<5.00	U	5.00	P	12/09/2024	12:46	LB133834
	Cobalt	15.0	<15.0	U	15.0	P	12/09/2024	12:46	LB133834
	Copper	10.0	<10.0	U	10.0	P	12/09/2024	12:46	LB133834
	Iron	50.0	<50.0	U	50.0	P	12/09/2024	12:46	LB133834
	Lead	6.00	<6.00	U	6.00	P	12/09/2024	12:46	LB133834
	Magnesium	1000	<1000	U	1000	P	12/09/2024	12:46	LB133834
	Manganese	10.0	<10.0	U	10.0	P	12/09/2024	12:46	LB133834
	Nickel	20.0	<20.0	U	20.0	P	12/09/2024	12:46	LB133834
	Potassium	1000	<1000	U	1000	P	12/09/2024	12:46	LB133834
	Selenium	10.0	<10.0	U	10.0	P	12/09/2024	12:46	LB133834
	Silver	5.00	<5.00	U	5.00	P	12/09/2024	12:46	LB133834
	Sodium	1000	<1000	U	1000	P	12/09/2024	12:46	LB133834
	Thallium	20.0	<20.0	U	20.0	P	12/09/2024	12:46	LB133834
	Vanadium	20.0	<20.0	U	20.0	P	12/09/2024	12:46	LB133834
	Zinc	20.0	<20.0	U	20.0	P	12/09/2024	12:46	LB133834

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: New York City DEP of Environmental Protection/B **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	267000	255000	105	216000	294000	12/06/2024	12:50	LB133790
	Antimony	-2.00			-50	50	12/06/2024	12:50	LB133790
	Arsenic	6.39			-20	20	12/06/2024	12:50	LB133790
	Barium	1.82	6.0	30	-94	106	12/06/2024	12:50	LB133790
	Beryllium	1.48			-6	6	12/06/2024	12:50	LB133790
	Cadmium	4.33	1.0	433	-5	7	12/06/2024	12:50	LB133790
	Calcium	250000	245000	102	208000	282000	12/06/2024	12:50	LB133790
	Chromium	58.3	52.0	112	42	62	12/06/2024	12:50	LB133790
	Cobalt	2.43			-30	30	12/06/2024	12:50	LB133790
	Copper	-0.59	2.0	30	-18	22	12/06/2024	12:50	LB133790
	Iron	101000	101000	100	85600	116500	12/06/2024	12:50	LB133790
	Lead	11.9			-12	12	12/06/2024	12:50	LB133790
	Magnesium	273000	255000	107	216000	294000	12/06/2024	12:50	LB133790
	Manganese	4.87	7.0	70	-13	27	12/06/2024	12:50	LB133790
	Nickel	32.9	2.0	1645	-38	42	12/06/2024	12:50	LB133790
	Potassium	14.1			0	0	12/06/2024	12:50	LB133790
	Selenium	-18.6			-20	20	12/06/2024	12:50	LB133790
	Silver	-1.23			-10	10	12/06/2024	12:50	LB133790
	Sodium	120			0	0	12/06/2024	12:50	LB133790
	Thallium	-2.74			-40	40	12/06/2024	12:50	LB133790
	Vanadium	8.34			-40	40	12/06/2024	12:50	LB133790
	Zinc	-6.25			-40	40	12/06/2024	12:50	LB133790
ICSAB01	Aluminum	264000	247000	107	209000	285000	12/06/2024	12:55	LB133790
	Antimony	608	618	98	525	711	12/06/2024	12:55	LB133790
	Arsenic	106	104	102	88.4	120	12/06/2024	12:55	LB133790
	Barium	501	537	93	437	637	12/06/2024	12:55	LB133790
	Beryllium	507	495	102	420	570	12/06/2024	12:55	LB133790
	Cadmium	1010	972	104	826	1120	12/06/2024	12:55	LB133790
	Calcium	246000	235000	105	199000	271000	12/06/2024	12:55	LB133790
	Chromium	561	542	104	460	624	12/06/2024	12:55	LB133790
	Cobalt	508	476	107	404	548	12/06/2024	12:55	LB133790
	Copper	486	511	95	434	588	12/06/2024	12:55	LB133790
	Iron	99200	99300	100	84400	114500	12/06/2024	12:55	LB133790
	Lead	60.5	49.0	124	37	61	12/06/2024	12:55	LB133790
	Magnesium	268000	248000	108	210000	286000	12/06/2024	12:55	LB133790
	Manganese	495	507	98	430	584	12/06/2024	12:55	LB133790
	Nickel	1030	954	108	810	1100	12/06/2024	12:55	LB133790
	Potassium	-35.5			0	0	12/06/2024	12:55	LB133790
	Selenium	33.1	46.0	72	26	66	12/06/2024	12:55	LB133790
	Silver	196	201	98	170	232	12/06/2024	12:55	LB133790
	Sodium	104			0	0	12/06/2024	12:55	LB133790
	Thallium	81.4	108	75	68	148	12/06/2024	12:55	LB133790

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: New York City DEP of Environmental Protection/B **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	501	491	102	417	565	12/06/2024	12:55	LB133790
	Zinc	942	952	99	809	1095	12/06/2024	12:55	LB133790
ICSA	Aluminum	249000	255000	98	216000	294000	12/06/2024	13:00	LB133790
	Antimony	-7.98			-50	50	12/06/2024	13:00	LB133790
	Arsenic	-32.4			-20	20	12/06/2024	13:00	LB133790
	Barium	-61.5	6.0	1025	-94	106	12/06/2024	13:00	LB133790
	Beryllium	3.42			-6	6	12/06/2024	13:00	LB133790
	Cadmium	2.74	1.0	274	-5	7	12/06/2024	13:00	LB133790
	Calcium	246000	245000	100	208000	282000	12/06/2024	13:00	LB133790
	Chromium	58.9	52.0	113	42	62	12/06/2024	13:00	LB133790
	Cobalt	3.44			-30	30	12/06/2024	13:00	LB133790
	Copper	0.47	2.0	23	-18	22	12/06/2024	13:00	LB133790
	Iron	96500	101000	96	85600	116500	12/06/2024	13:00	LB133790
	Lead	-15.2			-12	12	12/06/2024	13:00	LB133790
	Magnesium	252000	255000	99	216000	294000	12/06/2024	13:00	LB133790
	Manganese	2.53	7.0	36	-13	27	12/06/2024	13:00	LB133790
	Nickel	31.5	2.0	1575	-38	42	12/06/2024	13:00	LB133790
	Potassium	-1660			0	0	12/06/2024	13:00	LB133790
	Selenium	-11.4			-20	20	12/06/2024	13:00	LB133790
	Silver	9.42			-10	10	12/06/2024	13:00	LB133790
	Sodium	475			0	0	12/06/2024	13:00	LB133790
	Thallium	-11.5			-40	40	12/06/2024	13:00	LB133790
	Vanadium	3.28			-40	40	12/06/2024	13:00	LB133790
	Zinc	132			-40	40	12/06/2024	13:00	LB133790
ICSAB	Aluminum	254000	247000	103	209000	285000	12/06/2024	13:04	LB133790
	Antimony	579	618	94	525	711	12/06/2024	13:04	LB133790
	Arsenic	107	104	103	88.4	120	12/06/2024	13:04	LB133790
	Barium	440	537	82	437	637	12/06/2024	13:04	LB133790
	Beryllium	526	495	106	420	570	12/06/2024	13:04	LB133790
	Cadmium	1010	972	104	826	1120	12/06/2024	13:04	LB133790
	Calcium	250000	235000	106	199000	271000	12/06/2024	13:04	LB133790
	Chromium	585	542	108	460	624	12/06/2024	13:04	LB133790
	Cobalt	504	476	106	404	548	12/06/2024	13:04	LB133790
	Copper	522	511	102	434	588	12/06/2024	13:04	LB133790
	Iron	97400	99300	98	84400	114500	12/06/2024	13:04	LB133790
	Lead	50.0	49.0	102	37	61	12/06/2024	13:04	LB133790
	Magnesium	257000	248000	104	210000	286000	12/06/2024	13:04	LB133790
	Manganese	511	507	101	430	584	12/06/2024	13:04	LB133790
	Nickel	1050	954	110	810	1100	12/06/2024	13:04	LB133790
	Potassium	-1380			0	0	12/06/2024	13:04	LB133790
	Selenium	59.2	46.0	129	26	66	12/06/2024	13:04	LB133790
	Silver	193	201	96	170	232	12/06/2024	13:04	LB133790

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: New York City DEP of Environmental Protection/B **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Sodium	520			0	0	12/06/2024	13:04	LB133790
	Thallium	80.1	108	74	68	148	12/06/2024	13:04	LB133790
	Vanadium	512	491	104	417	565	12/06/2024	13:04	LB133790
	Zinc	1200	952	126	809	1095	12/06/2024	13:04	LB133790
ICSA01	Aluminum	256000	255000	100	216000	294000	12/09/2024	12:08	LB133834
	Antimony	-0.86			-50	50	12/09/2024	12:08	LB133834
	Arsenic	3.35			-20	20	12/09/2024	12:08	LB133834
	Barium	5.62	6.0	94	-94	106	12/09/2024	12:08	LB133834
	Beryllium	1.49			-6	6	12/09/2024	12:08	LB133834
	Cadmium	3.33	1.0	333	-5	7	12/09/2024	12:08	LB133834
	Calcium	243000	245000	99	208000	282000	12/09/2024	12:08	LB133834
	Chromium	56.5	52.0	109	42	62	12/09/2024	12:08	LB133834
	Cobalt	2.04			-30	30	12/09/2024	12:08	LB133834
	Copper	18.7	2.0	935	-18	22	12/09/2024	12:08	LB133834
	Iron	95800	101000	95	85600	116500	12/09/2024	12:08	LB133834
	Lead	11.9			-12	12	12/09/2024	12:08	LB133834
	Magnesium	265000	255000	104	216000	294000	12/09/2024	12:08	LB133834
	Manganese	4.34	7.0	62	-13	27	12/09/2024	12:08	LB133834
	Nickel	31.5	2.0	1575	-38	42	12/09/2024	12:08	LB133834
	Potassium	18.3			0	0	12/09/2024	12:08	LB133834
	Selenium	-19.6			-20	20	12/09/2024	12:08	LB133834
	Silver	-4.87			-10	10	12/09/2024	12:08	LB133834
	Sodium	54.5			0	0	12/09/2024	12:08	LB133834
	Thallium	-3.39			-40	40	12/09/2024	12:08	LB133834
	Vanadium	9.50			-40	40	12/09/2024	12:08	LB133834
	Zinc	31.5			-40	40	12/09/2024	12:08	LB133834
ICSAB01	Aluminum	253000	247000	102	209000	285000	12/09/2024	12:17	LB133834
	Antimony	622	618	101	525	711	12/09/2024	12:17	LB133834
	Arsenic	105	104	101	88.4	120	12/09/2024	12:17	LB133834
	Barium	508	537	95	437	637	12/09/2024	12:17	LB133834
	Beryllium	522	495	106	420	570	12/09/2024	12:17	LB133834
	Cadmium	1000	972	103	826	1120	12/09/2024	12:17	LB133834
	Calcium	239000	235000	102	199000	271000	12/09/2024	12:17	LB133834
	Chromium	563	542	104	460	624	12/09/2024	12:17	LB133834
	Cobalt	507	476	106	404	548	12/09/2024	12:17	LB133834
	Copper	498	511	98	434	588	12/09/2024	12:17	LB133834
	Iron	95800	99300	96	84400	114500	12/09/2024	12:17	LB133834
	Lead	52.4	49.0	107	37	61	12/09/2024	12:17	LB133834
	Magnesium	259000	248000	104	210000	286000	12/09/2024	12:17	LB133834
	Manganese	494	507	97	430	584	12/09/2024	12:17	LB133834
	Nickel	1020	954	107	810	1100	12/09/2024	12:17	LB133834
	Potassium	-36.7			0	0	12/09/2024	12:17	LB133834

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: New York City DEP of Environmental Protection/B **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	29.8	46.0	65	26	66	12/09/2024	12:17	LB133834
	Silver	195	201	97	170	232	12/09/2024	12:17	LB133834
	Sodium	5.96			0	0	12/09/2024	12:17	LB133834
	Thallium	81.8	108	76	68	148	12/09/2024	12:17	LB133834
	Vanadium	500	491	102	417	565	12/09/2024	12:17	LB133834
	Zinc	922	952	97	809	1095	12/09/2024	12:17	LB133834
ICSA	Aluminum	250000	255000	98	216000	294000	12/09/2024	12:21	LB133834
	Antimony	4.82			-50	50	12/09/2024	12:21	LB133834
	Arsenic	-56.9			-20	20	12/09/2024	12:21	LB133834
	Barium	-8.92	6.0	149	-94	106	12/09/2024	12:21	LB133834
	Beryllium	2.01			-6	6	12/09/2024	12:21	LB133834
	Cadmium	-1.42	1.0	142	-5	7	12/09/2024	12:21	LB133834
	Calcium	248000	245000	101	208000	282000	12/09/2024	12:21	LB133834
	Chromium	56.7	52.0	109	42	62	12/09/2024	12:21	LB133834
	Cobalt	-0.98			-30	30	12/09/2024	12:21	LB133834
	Copper	31.9	2.0	1595	-18	22	12/09/2024	12:21	LB133834
	Iron	95600	101000	95	85600	116500	12/09/2024	12:21	LB133834
	Lead	8.41			-12	12	12/09/2024	12:21	LB133834
	Magnesium	255000	255000	100	216000	294000	12/09/2024	12:21	LB133834
	Manganese	7.71	7.0	110	-13	27	12/09/2024	12:21	LB133834
	Nickel	34.2	2.0	1710	-38	42	12/09/2024	12:21	LB133834
	Potassium	-1160			0	0	12/09/2024	12:21	LB133834
	Selenium	-15.9			-20	20	12/09/2024	12:21	LB133834
	Silver	-9.38			-10	10	12/09/2024	12:21	LB133834
	Sodium	-653			0	0	12/09/2024	12:21	LB133834
	Thallium	-31.4			-40	40	12/09/2024	12:21	LB133834
	Vanadium	2.17			-40	40	12/09/2024	12:21	LB133834
	Zinc	155			-40	40	12/09/2024	12:21	LB133834
ICSAB	Aluminum	248000	247000	100	209000	285000	12/09/2024	12:25	LB133834
	Antimony	608	618	98	525	711	12/09/2024	12:25	LB133834
	Arsenic	25.9	104	25	88.4	120	12/09/2024	12:25	LB133834
	Barium	489	537	91	437	637	12/09/2024	12:25	LB133834
	Beryllium	535	495	108	420	570	12/09/2024	12:25	LB133834
	Cadmium	1000	972	103	826	1120	12/09/2024	12:25	LB133834
	Calcium	246000	235000	105	199000	271000	12/09/2024	12:25	LB133834
	Chromium	589	542	109	460	624	12/09/2024	12:25	LB133834
	Cobalt	499	476	105	404	548	12/09/2024	12:25	LB133834
	Copper	543	511	106	434	588	12/09/2024	12:25	LB133834
	Iron	95500	99300	96	84400	114500	12/09/2024	12:25	LB133834
	Lead	68.0	49.0	139	37	61	12/09/2024	12:25	LB133834
	Magnesium	252000	248000	102	210000	286000	12/09/2024	12:25	LB133834
	Manganese	525	507	104	430	584	12/09/2024	12:25	LB133834

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: New York City DEP of Environmental Protection/B **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Nickel	1040	954	109	810	1100	12/09/2024	12:25	LB133834
	Potassium	-1200			0	0	12/09/2024	12:25	LB133834
	Selenium	46.3	46.0	101	26	66	12/09/2024	12:25	LB133834
	Silver	182	201	90	170	232	12/09/2024	12:25	LB133834
	Sodium	-642			0	0	12/09/2024	12:25	LB133834
	Thallium	80.5	108	74	68	148	12/09/2024	12:25	LB133834
	Vanadium	513	491	104	417	565	12/09/2024	12:25	LB133834
	Zinc	1210	952	127	809	1095	12/09/2024	12:25	LB133834



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: New York City DEP of Environmental Prot **level:** low **sdg no.:** P4997
contract: NEWY17 **lab code:** CHEM **case no.:** P4997 **sas no.:** P4997
matrix: Water **sample id:** P5018-07 **client id:** 14B-(1-4)-COMPMS
Percent Solids for Sample: NA **Spiked ID:** P5018-07MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	717000		698000		1000	1910		P
Antimony	ug/L	75 - 125	347		25.0	U	400	87		P
Arsenic	ug/L	75 - 125	454		54.4		400	100		P
Barium	ug/L	75 - 125	661		543		100	118		P
Beryllium	ug/L	75 - 125	102		3.00	U	100	102		P
Cadmium	ug/L	75 - 125	99.0		2.25	J	100	97		P
Calcium	ug/L	75 - 125	37000		35400		500	315		P
Chromium	ug/L	75 - 125	205		17.8		200	93		P
Cobalt	ug/L	75 - 125	103		6.13	J	100	97		P
Copper	ug/L	75 - 125	580		465		150	77		P
Iron	ug/L	75 - 125	22300		23500		1500	-83		P
Lead	ug/L	75 - 125	458		17.2		500	88		P
Magnesium	ug/L	75 - 125	9600		8520		1000	108		P
Manganese	ug/L	75 - 125	16800		17600		100	-779		P
Mercury	ug/L	75 - 125	3.49		0.53		4.0	74	N	CV
Nickel	ug/L	75 - 125	258		18.3	J	250	96		P
Potassium	ug/L	75 - 125	9010		3270		5000	115		P
Selenium	ug/L	75 - 125	1020		9.57	J	1000	101		P
Silver	ug/L	75 - 125	43.7		2.13	J	37.5	111		P
Sodium	ug/L	75 - 125	38200		36300		1500	131		P
Thallium	ug/L	75 - 125	826		20.0	U	1000	83		P
Vanadium	ug/L	75 - 125	171		24.5		150	98		P
Zinc	ug/L	75 - 125	232		127		100	105		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: New York City DEP of Environmental Prot **level:** low **sdg no.:** P4997
contract: NEWY17 **lab code:** CHEM **case no.:** P4997 **sas no.:** P4997
matrix: Water **sample id:** P5018-07 **client id:** 14B-(1-4)-COMPMSD
Percent Solids for Sample: NA **Spiked ID:** P5018-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	712000		698000		1000	1428		P
Antimony	ug/L	75 - 125	344		25.0	U	400	86		P
Arsenic	ug/L	75 - 125	449		54.4		400	99		P
Barium	ug/L	75 - 125	646		543		100	103		P
Beryllium	ug/L	75 - 125	102		3.00	U	100	102		P
Cadmium	ug/L	75 - 125	97.6		2.25	J	100	95		P
Calcium	ug/L	75 - 125	36500		35400		500	224		P
Chromium	ug/L	75 - 125	204		17.8		200	93		P
Cobalt	ug/L	75 - 125	102		6.13	J	100	95		P
Copper	ug/L	75 - 125	573		465		150	72	N	P
Iron	ug/L	75 - 125	21800		23500		1500	-117		P
Lead	ug/L	75 - 125	451		17.2		500	87		P
Magnesium	ug/L	75 - 125	9520		8520		1000	99		P
Manganese	ug/L	75 - 125	16600		17600		100	-1054		P
Mercury	ug/L	75 - 125	3.62		0.53		4.0	77		CV
Nickel	ug/L	75 - 125	253		18.3	J	250	94		P
Potassium	ug/L	75 - 125	8840		3270		5000	111		P
Selenium	ug/L	75 - 125	1020		9.57	J	1000	101		P
Silver	ug/L	75 - 125	43.3		2.13	J	37.5	110		P
Sodium	ug/L	75 - 125	37200		36300		1500	60		P
Thallium	ug/L	75 - 125	802		20.0	U	1000	80		P
Vanadium	ug/L	75 - 125	169		24.5		150	96		P
Zinc	ug/L	75 - 125	230		127		100	103		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Matrix: Water

Level: LOW

Client ID: 14B-(1-4)-COMPA

Sample ID: P5018-07

Spiked ID: P5018-07A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	85 - 115	575		465		150	74		P
Mercury	ug/L	75 - 125	3.88		0.53		4.00	84		CV

A
B
C
D
E
F
G
H

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: New York City DEP of Environmental Prot **Level:** LOW **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
Matrix: Water **Sample ID:** P5018-07 **Client ID:** 14B-(1-4)-COMPDUP
Percent Solids for Sample: NA **Duplicate ID** P5018-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	698000		708000		1		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	54.4		54.6		0		P
Barium	ug/L	20	543		554		2		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	2.25	J	2.34	J	4		P
Calcium	ug/L	20	35400		35600		1		P
Chromium	ug/L	20	17.8		17.8		0		P
Cobalt	ug/L	20	6.13	J	5.97	J	3		P
Copper	ug/L	20	465		471		1		P
Iron	ug/L	20	23500		24000		2		P
Lead	ug/L	20	17.2		18.7		8		P
Magnesium	ug/L	20	8520		8570		1		P
Manganese	ug/L	20	17600		17900		2		P
Mercury	ug/L	20	0.53		0.50		6		CV
Nickel	ug/L	20	18.3	J	18.6	J	2		P
Potassium	ug/L	20	3270		3330		2		P
Selenium	ug/L	20	9.57	J	8.14	J	16		P
Silver	ug/L	20	2.13	J	2.27	J	6		P
Sodium	ug/L	20	36300		37100		2		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	24.5		24.9		2		P
Zinc	ug/L	20	127		128		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: New York City DEP of Environmental Prot **Level:** LOW **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997
Matrix: Water **Sample ID:** P5018-07MS **Client ID:** 14B-(1-4)-COMPMSD
Percent Solids for Sample: NA **Duplicate ID** P5018-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	717000		712000		1		P
Antimony	ug/L	20	347		344		1		P
Arsenic	ug/L	20	454		449		1		P
Barium	ug/L	20	661		646		2		P
Beryllium	ug/L	20	102		102		0		P
Cadmium	ug/L	20	99.0		97.6		1		P
Calcium	ug/L	20	37000		36500		1		P
Chromium	ug/L	20	205		204		0		P
Cobalt	ug/L	20	103		102		1		P
Copper	ug/L	20	580		573		1		P
Iron	ug/L	20	22300		21800		2		P
Lead	ug/L	20	458		451		2		P
Magnesium	ug/L	20	9600		9520		1		P
Manganese	ug/L	20	16800		16600		1		P
Mercury	ug/L	20	3.49		3.62		4		CV
Nickel	ug/L	20	258		253		2		P
Potassium	ug/L	20	9010		8840		2		P
Selenium	ug/L	20	1020		1020		0		P
Silver	ug/L	20	43.7		43.3		1		P
Sodium	ug/L	20	38200		37200		3		P
Thallium	ug/L	20	826		802		3		P
Vanadium	ug/L	20	171		169		1		P
Zinc	ug/L	20	232		230		1		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: New York City DEP of Environmental Prot **SDG No.:** P4997
Contract: NEWY17 **Lab Code:** CHEM **Case No.:** P4997 **SAS No.:** P4997

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165287BS							
Aluminum	ug/L	1000	983		98	85 - 115	P
Antimony	ug/L	400	377		94	85 - 115	P
Arsenic	ug/L	400	368		92	85 - 115	P
Barium	ug/L	100	96.0		96	85 - 115	P
Beryllium	ug/L	100	103		103	85 - 115	P
Cadmium	ug/L	100	93.5		94	85 - 115	P
Calcium	ug/L	500	505	J	101	85 - 115	P
Chromium	ug/L	200	196		98	85 - 115	P
Cobalt	ug/L	100	94.6		95	85 - 115	P
Copper	ug/L	150	149		99	85 - 115	P
Iron	ug/L	1500	1390		93	85 - 115	P
Lead	ug/L	500	465		93	85 - 115	P
Magnesium	ug/L	1000	979	J	98	85 - 115	P
Manganese	ug/L	100	99.5		100	85 - 115	P
Nickel	ug/L	250	238		95	85 - 115	P
Potassium	ug/L	5000	4470		89	85 - 115	P
Selenium	ug/L	1000	939		94	85 - 115	P
Silver	ug/L	37.5	35.9		96	85 - 115	P
Sodium	ug/L	1500	1300		87	85 - 115	P
Thallium	ug/L	1000	1040		104	85 - 115	P
Vanadium	ug/L	150	149		99	85 - 115	P
Zinc	ug/L	100	100		100	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165329BS Mercury	ug/L	4.0	3.91		98	85 - 115	CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

14B-(1-4)-COMPL

Lab Name: Chemtech Consulting Group

Contract: NEWY17

Lab Code: CHEM Lb No.: lb133790 Lab Sample ID : P5018-07L SDG No.: P4997

Matrix (soil/water): Water Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Aluminum	698000		722000		4		P
Antimony	25.0	U	125	U			P
Arsenic	54.4		53.2		2		P
Barium	543		532		2		P
Beryllium	3.00	U	15.0	U			P
Cadmium	2.25	J	1.70	J	25		P
Calcium	35400		37300		5		P
Chromium	17.8		19.3	J	9		P
Cobalt	6.13	J	6.52	J	6		P
Copper	465		520		12		P
Iron	23500		23500		0		P
Lead	17.2		22.8	J	33		P
Magnesium	8520		8940		5		P
Manganese	17600		18900		7		P
Mercury	0.53		0.38	J	29		CV
Nickel	18.3	J	20.0	J	9		P
Potassium	3270		2740	J	16		P
Selenium	9.57	J	50.0	U	100.0		P
Silver	2.13	J	25.0	U	100.0		P
Sodium	36300		31400		14		P
Thallium	20.0	U	100	U			P
Vanadium	24.5		24.1	J	1		P
Zinc	127		123		3		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	-0.0075970	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Case No.: P4997

SAS No.: P4997

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: New York City DEP of Environmental Prot

SDG No.: P4997

Contract: NEWY17

Lab Code: CHEM

Method: _____

Case No.: P4997 **SAS No.:** P4997

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB165287							
P4997-07	14B-(1-4)-COMP	SAM	WATER	11/27/2024	50.0	25.0	
P5018-07DUP	14B-(1-4)-COMPDUP	DUP	WATER	11/27/2024	50.0	25.0	
P5018-07MS	14B-(1-4)-COMPMS	MS	WATER	11/27/2024	50.0	25.0	
P5018-07MSD	14B-(1-4)-COMPMSD	MSD	WATER	11/27/2024	50.0	25.0	
PB165287BL	PB165287BL	MB	WATER	11/27/2024	50.0	25.0	
PB165287BS	PB165287BS	LCS	WATER	11/27/2024	50.0	25.0	

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>New York City DEP of Environmental Prot</u>	SDG No.:	<u>P4997</u>
Contract:	<u>NEWY17</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>P4997</u>
		SAS No.:	<u>P4997</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB165329							
P4997-07	14B-(1-4)-COMP	SAM	WATER	12/02/2024	30.0	30.0	
P5018-07DUP	14B-(1-4)-COMPDUP	DUP	WATER	12/02/2024	30.0	30.0	
P5018-07MS	14B-(1-4)-COMPMS	MS	WATER	12/02/2024	30.0	30.0	
P5018-07MSD	14B-(1-4)-COMPMSD	MSD	WATER	12/02/2024	30.0	30.0	
PB165329BL	PB165329BL	MB	WATER	12/02/2024	30.0	30.0	
PB165329BS	PB165329BS	LCS	WATER	12/02/2024	30.0	30.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: New York City DEP of Environmental Prot

Contract: NEWY17

Lab code: CHEM **Case no.:** P4997 **Sas no.:** P4997

Sdg no.: P4997

Instrument id number: **Method:**

Run number: LB133684

Start date: 12/02/2024 **End date:** 12/02/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1312	HG
S0.05	S0.05	1	1317	HG
S0.2	S0.2	1	1321	HG
S2.5	S2.5	1	1323	HG
S5	S5	1	1325	HG
S7.5	S7.5	1	1328	HG
S10	S10	1	1333	HG
ICV107	ICV107	1	1343	HG
ICB107	ICB107	1	1345	HG
CCV69	CCV69	1	1348	HG
CCB69	CCB69	1	1350	HG
CRA	CRA	1	1356	HG
PB165329BL	PB165329BL	1	1402	HG
PB165329BS	PB165329BS	1	1405	HG
P4997-07	14B-(1-4)-COMP	1	1407	HG
P5018-07DUP	14B-(1-4)-COMPDUP	1	1412	HG
P5018-07MS	14B-(1-4)-COMPMS	1	1414	HG
P5018-07MSD	14B-(1-4)-COMPMSD	1	1416	HG
CCV70	CCV70	1	1418	HG
CCB70	CCB70	1	1421	HG
P5018-07L	14B-(1-4)-COMPL	5	1423	HG
P5018-07A	14B-(1-4)-COMPA	1	1425	HG
CCV71	CCV71	1	1428	HG
CCB71	CCB71	1	1430	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: New York City DEP of Environmental Prot

Contract: NEWY17

Lab code: CHEM

Case no.: P4997

Sas no.: P4997

Sdg no.: P4997

Instrument id number: _____

Method: _____

Run number: LB133790

Start date: 12/06/2024

End date: 12/06/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1148	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1152	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1157	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1201	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1205	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1210	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1214	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1220	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1224	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1228	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1250	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1255	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA	ICSA	20	1300	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB	ICSAB	20	1304	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1309	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1313	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1338	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1342	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4997-07	14B-(1-4)-COMP	1	1347	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P5018-07DUP	14B-(1-4)-COMPDUP	1	1404	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P5018-07L	14B-(1-4)-COMPL	5	1408	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P5018-07MS	14B-(1-4)-COMPMS	1	1412	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P5018-07MSD	14B-(1-4)-COMPMSD	1	1417	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P5018-07A	14B-(1-4)-COMPA	1	1421	Cu
CCV03	CCV03	1	1442	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1446	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1540	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1544	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1635	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1639	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1727	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1731	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1821	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1827	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: New York City DEP of Environmental Prot **Contract:** NEWY17

Lab code: CHEM **Case no.:** P4997 **Sas no.:** P4997 **Sdg no.:** P4997

Instrument id number: **Method:** **Run number:** LB133834

Start date: 12/09/2024 **End date:** 12/09/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1106	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1110	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1114	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1118	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1123	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1127	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1135	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1149	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1154	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1159	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1208	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1217	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA	ICSA	20	1221	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB	ICSAB	20	1225	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1229	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1234	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB165287BL	PB165287BL	1	1246	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB165287BS	PB165287BS	1	1251	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1330	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1335	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1423	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1427	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1513	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1517	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1552	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1556	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

LAB CHRONICLE

OrderID:	P4997	OrderDate:	11/25/2024 1:27:00 PM
Client:	New York City DEP of Environmental Protection/BWS	Project:	Industrial Wastewater Discharge Permit - Fall 2024
Contact:	Nicholas Prokopowicz	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4997-01	14B-1	WATER			11/25/24 07:55			11/25/24
			Field pH	SM4500-H B			11/25/24 07:57	
			Field Temperature	SM2550-B			11/25/24 07:57	
			Non-Polar Material	1664A			12/02/24 09:40	
			TSS	SM2540 D			11/26/24 10:00	
P4997-02	14B-2	WATER			11/25/24 08:56			11/25/24
			Field pH	SM4500-H B			11/25/24 08:59	
			Field Temperature	SM2550-B			11/25/24 08:59	
			Non-Polar Material	1664A			12/02/24 09:40	
			TSS	SM2540 D			11/26/24 10:00	
P4997-03	14B-3	WATER			11/25/24 09:54			11/25/24
			Field pH	SM4500-H B			11/25/24 09:57	
			Field Temperature	SM2550-B			11/25/24 09:57	
			Non-Polar Material	1664A			12/02/24 09:40	
			TSS	SM2540 D			11/26/24 10:00	

LAB CHRONICLE

P4997-04	14B-4	WATER	11/25/24 10:54			11/25/24
		Field pH	SM4500-H B		11/25/24 10:56	
		Field Temperature	SM2550-B		11/25/24 10:56	
		Non-Polar Material	1664A		12/02/24 09:40	
		TSS	SM2540 D		11/26/24 10:00	
P4997-07	14B-(1-4)-COMP	WATER	11/25/24 14:55			11/25/24
		Cyanide	SM4500-CN C,E	12/05/24	12/05/24 13:20	
		Cyanide-Amenable	SM4500-CN B,G Cyanide-Amenable		12/05/24 00:00	
		Hexavalent Chromium	SM3500-Cr B		11/25/24 16:14	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24 07:55
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-1	SDG No.:	P4997
Lab Sample ID:	P4997-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Field pH	6.74		1	0	0	pH		11/25/24 07:57	SM4500-H B
Field Temperature	11.4		1	0	0	o C		11/25/24 07:57	SM 2550 B-10
Non-Polar Material	0.60	J	1	0.40	5.00	mg/L		12/02/24 09:40	1664A
TSS	4130		1	1.00	4.00	mg/L		11/26/24 10:00	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24 08:56
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-2	SDG No.:	P4997
Lab Sample ID:	P4997-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Field pH	6.68		1	0	0	pH		11/25/24 08:59	SM4500-H B
Field Temperature	12.6		1	0	0	o C		11/25/24 08:59	SM 2550 B-10
Non-Polar Material	0.50	J	1	0.40	5.00	mg/L		12/02/24 09:40	1664A
TSS	4410		1	1.00	4.00	mg/L		11/26/24 10:00	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24 09:54
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-3	SDG No.:	P4997
Lab Sample ID:	P4997-03	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Field pH	6.73		1	0	0	pH		11/25/24 09:57	SM4500-H B
Field Temperature	13.3		1	0	0	o C		11/25/24 09:57	SM 2550 B-10
Non-Polar Material	0.60	J	1	0.40	5.00	mg/L		12/02/24 09:40	1664A
TSS	4720		1	1.00	4.00	mg/L		11/26/24 10:00	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24 10:54
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-4	SDG No.:	P4997
Lab Sample ID:	P4997-04	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Field pH	6.76		1	0	0	pH		11/25/24 10:56	SM4500-H B
Field Temperature	13.6		1	0	0	o C		11/25/24 10:56	SM 2550 B-10
Non-Polar Material	0.80	J	1	0.40	5.00	mg/L		12/02/24 09:40	1664A
TSS	4480		1	1.00	4.00	mg/L		11/26/24 10:00	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	New York City DEP of Environmental Protection/BWS	Date Collected:	11/25/24 14:55
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Date Received:	11/25/24
Client Sample ID:	14B-(1-4)-COMP	SDG No.:	P4997
Lab Sample ID:	P4997-07	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.0045	J	1	0.00093	0.0050	mg/L	12/05/24 09:00	12/05/24 13:20	SM 4500-CN C-16 plus E-16
Cyanide-Amenable	0.0010	U	1	0.0010	0.0050	mg/L		12/05/24 00:00	SM 4500-CN B-16 plus G-16
Dissolved Hexavalent Chromium	0.0020	U	1	0.0020	0.010	mg/L		11/25/24 16:14	SM 3500-Cr B-11

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133615

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.503	0.5	101	95-105	11/25/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	11/25/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	11/25/2024

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133772

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.097	0.099	98	85-115	12/05/2024
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	12/05/2024
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	12/05/2024
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	12/05/2024

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133821

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.03	7	100	90-110	11/25/2024
Sample ID:	CCV1						
Field pH		pH	7.01	7.00	100	90-110	11/25/2024
Sample ID:	CCV2						
Field pH		pH	7.00	7.00	100	90-110	11/25/2024

Initial and Continuing Calibration Blank Summary

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133615

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	11/25/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	11/25/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	11/25/2024

Initial and Continuing Calibration Blank Summary

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133772

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024

Preparation Blank Summary

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P4997

Project: Industrial Wastewater Discharge Permit - Fall 2024

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB133615BL Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.002	0.01	11/25/2024
Sample ID: LB133644BL TSS	mg/L	< 2.0000	2.0000	U	1	4	11/26/2024
Sample ID: LB133678BL Non-Polar Material	mg/L	< 2.5000	2.5000	U	0.4	5.0	12/02/2024
Sample ID: PB165397BL Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024

Matrix Spike Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Sample ID:** P4997-04
Client ID: 14B-4MS **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Non-Polar Material	mg/L	78-114	20.9		0.80	J	20.0	1	101		12/02/2024

Matrix Spike Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Sample ID:** P4997-04
Client ID: 14B-4MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Non-Polar Material	mg/L	78-114	21.0		0.80	J	20.0	1	101		12/02/2024

Matrix Spike Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-07
Client ID:	14B-(1-4)-COMPMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.98		0.0020	U	1.0	2	98		11/25/2024
Cyanide	mg/L	75-125	0.040		0.0045	J	0.04	1	89		12/05/2024

Matrix Spike Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-07
Client ID:	14B-(1-4)-COMPMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.98		0.0020	U	1.0	2	98		11/25/2024
Cyanide	mg/L	75-125	0.040		0.0045	J	0.04	1	89		12/05/2024

Matrix Spike Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P5018-04
Client ID:	14B-4MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Non-Polar Material	mg/L	78-114	20.9		0.60	J	20.0	1	102		12/02/2024

Matrix Spike Summary

Client:

New York City DEP of Environmental Protection/BWS

SDG No.:

P4997

Project:

Industrial Wastewater Discharge Permit - Fall 2024

Sample ID:

P5018-04

Client ID:

14B-4MSD

Percent Solids for Spike Sample:

0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Non-Polar Material	mg/L	78-114	20.8		0.60	J	20.0	1	101		12/02/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-04
Client ID:	14B-4DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	4480		4520		1	0.89		11/26/2024
Field pH	pH	+/-20	6.76		6.74		1	0.3		11/25/2024
Field Temperature	o C	+/-20	13.6		13.6		1	0		11/25/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-04
Client ID:	14B-4MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Non-Polar Material	mg/L	+/-18	20.9		21.0		1	0.48		12/02/2024

Duplicate Sample Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Sample ID:** P4997-07
Client ID: 14B-(1-4)-COMPDUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0020	U	0.0020	U	1	0		11/25/2024
Cyanide	mg/L	+/-20	0.0045	J	0.0042	J	1	7		12/05/2024

Duplicate Sample Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Sample ID:** P4997-07
Client ID: 14B-(1-4)-COMPMSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.98		0.98		2	0.82		11/25/2024
Cyanide	mg/L	+/-20	0.040		0.040		1	0		12/05/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P5018-04
Client ID:	14B-4MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Non-Polar Material	mg/L	+/-18	20.9		20.8		1	0.48		12/02/2024

Laboratory Control Sample Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Run No.:** LB133615

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133615BS							
Hexavalent Chromium	mg/L	0.5	0.50		100	1	90-111	11/25/2024

Laboratory Control Sample Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Run No.:** LB133644

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133644BS							
TSS	mg/L	550	540		98	1	90-110	11/26/2024

Laboratory Control Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Run No.:	LB133678

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133678BS							
Non-Polar Material	mg/L	20.0	17.0		85	1	78-114	12/02/2024

Laboratory Control Sample Summary

Client: New York City DEP of Environmental Protection/BWS **SDG No.:** P4997
Project: Industrial Wastewater Discharge Permit - Fall 2024 **Run No.:** LB133772

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165397BS							
Cyanide	mg/L	0.1	0.095		95	1	85-115	12/05/2024