

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-04
Client ID:	14B-4DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	4480		4520		1	0.89		11/26/2024
Field pH	pH	+/-20	6.76		6.74		1	0.3		11/25/2024
Field Temperature	o C	+/-20	13.6		13.6		1	0		11/25/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-04
Client ID:	14B-4MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Non-Polar Material	mg/L	+/-18	20.9		21.0		1	0.48		12/02/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-07
Client ID:	14B-(1-4)-COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0020	U	0.0020	U	1	0		11/25/2024
Cyanide	mg/L	+/-20	0.0045	J	0.0042	J	1	7		12/05/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P4997-07
Client ID:	14B-(1-4)-COMPMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.98		0.98		2	0.82		11/25/2024
Cyanide	mg/L	+/-20	0.040		0.040		1	0		12/05/2024

Duplicate Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P4997
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Sample ID:	P5018-04
Client ID:	14B-4MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Non-Polar Material	mg/L	+/-18	20.9		20.8		1	0.48		12/02/2024