

SOP ID : M200.7-Trace Elements-22

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3


Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 11/27/2024 Time : 08:40 Temp : 95 °C

End Digest Date: 11/27/2024 Time : 11:50 Temp : 95 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: 

Supervisor Signature: 



Temp : 1. 95°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6000
LFS-2	0.25	M6009
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
1:1 HNO3	1.00	MP83122
1:1 HCL	0.50	MP83105
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

Hot Block # 1 Cell # 50 Temp: 95 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/27/24 12:00		
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
P4997-07	14B-(1-4)-COMP	N/A	50	25	Black	Yellow	Clear	Clear	N/A	6
P5018-07	14B-(1-4)-COMP	N/A	50	25	Black	Yellow	Clear	Clear	N/A	7
P5018-07DUP	14B-(1-4)-COMPDUP	N/A	50	25	Black	Yellow	Clear	Clear	N/A	8
P5018-07MS	14B-(1-4)-COMPMS	N/A	50	25	Black	Yellow	Clear	Clear	M6000,M6009	9
P5018-07MSD	14B-(1-4)-COMPMSD	N/A	50	25	Black	Yellow	Clear	Clear	M6000,M6009	10
PB165287BL	PBW287	N/A	50	25	Colorless	Colorless	Clear	Clear	N/A	11
PB165287BS	LCS287	N/A	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	12