

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : 540 DEGRAW ST, BROOKLYN, NY - E9309

ENTACT

606 E. Baltimore Pike

Floor 3

Media, PA - 19063

Phone No: 4844440702

ORDER ID : P5001

ATTENTION : Jarod Stanfield



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	5
2.1) SPLP VOA- Case Narrative	5
3) Qualifier Page	8
4) QA Checklist	9
5) SPLP VOA Data	10
6) Shipping Document	79
6.1) CHAIN OF CUSTODY	80
6.2) Lab Certificate	82

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

1

Laboratory Name : Alliance Technical Group Client : ENTACT
 Project Location : Brooklyn, NY Project Number : E9309
 Laboratory Sample ID(s) : P5001 Sampling Date(s) : 11/25/2024
 List DKQP Methods Used (e.g., 8260,8270, et Cetra) **1312,8260D**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	☒ Yes ☐ No
1A	Were the method specified handling, preservation, and holding time requirements met?	☒ Yes ☐ No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	☐ Yes ☐ No ☒ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	☒ Yes ☐ No
3	Were samples received at an appropriate temperature (4±2° C)?	☒ Yes ☐ No ☐ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	☐ Yes ☒ No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	☒ Yes ☐ No ☒ Yes ☐ No ☐ N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	☒ Yes ☐ No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	☐ Yes ☒ No

Notes: For all questions to which the response was “No” (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is “No”, the data package does not meet the requirements for “Data of Known Quality.”

Cover Page

Order ID : P5001

Project ID : 540 Degraw St, Brooklyn, NY - E9309

Client : ENTACT

Lab Sample Number

P5001-01
P5001-02
P5001-03
P5001-04
P5001-05
P5001-06
P5001-08
P5001-09
P5001-10
P5001-11

Client Sample Number

SPLP-C1-1
SPLP-C1-2
SPLP-C1-3
SPLP-C2-1
SPLP-C2-2
SPLP-C2-3
SPLP-C3-1
SPLP-C3-2
SPLP-C3-3
SPLP-C4-1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 12/4/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

ENTACT

Project Name: 540 Degraw St, Brooklyn, NY - E9309

Project # N/A

Chemtech Project # P5001

Test Name: SPLP VOA

A. Number of Samples and Date of Receipt:

10 Solid samples were received on 11/25/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SPLP VOA and SPLP ZHE Ext. This data package contains results for SPLP VOA.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of SPLP VOA was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for SPLP-C1-1 [Dibromofluoromethane - 70%] this compound meet the NJDKQP criteria but did not meet the in-house criteria, the failure samples in surrogates were reanalyzed to confirm the results as per method and reported in the data while,

SPLP-C1-1RE [Dibromofluoromethane - 66%], SPLP-C1-2 [Dibromofluoromethane - 68%], SPLP-C1-2RE [Dibromofluoromethane - 67%], SPLP-C1-3 [Dibromofluoromethane - 63%], SPLP-C1-3RE [Dibromofluoromethane - 63%], SPLP-C2-1 [Dibromofluoromethane - 53%], SPLP-C2-1RE [Dibromofluoromethane - 58%], SPLP-C3-1 [Dibromofluoromethane - 60%], SPLP-C3-1RE [Dibromofluoromethane - 61%], SPLP-C4-1 [Dibromofluoromethane - 66%] and SPLP-C4-1RE [Dibromofluoromethane - 69%] these compounds did not meet the NJDKQP criteria and in-house criteria, the failure samples in surrogates were reanalyzed to confirm the results as per method and reported in the data while,

SPLP-C2-2 [Dibromofluoromethane - 60%], SPLP-C2-3 [Dibromofluoromethane - 62%], SPLP-C3-2 [Dibromofluoromethane - 59%], SPLP-C3-3 [Dibromofluoromethane - 60%], these compounds did not meet the NJDKQP criteria and in-house criteria, due to high concentration samples required dilution therefore samples reanalyzed with dilution and reported.

The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.

The RPD for {VX1127WBSD01} with File ID: VX044041.D met criteria except for 1,1,2-Trichloroethane[21%], 2-Butanone[23%], 2-Hexanone[21%], Acetone[33%], Bromochloromethane[30%], Bromoform[21%], Chloromethane[21%], Dibromochloromethane[23%], Dichlorodifluoromethane[23%] and Methylcyclohexane[22%] these compounds did not meet the NJDKQP criteria and in-house criteria due to difference in results of BS-BSD.

The Blank Spike met requirements for all samples .
The Blank Spike Duplicate met requirements for all samples .
The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration method (82X1121W.M) for Bromoform this compound is passing on Quadratic Regression.
The Continuous Calibration met the requirements .
The Tuning criteria met requirements.

Samples SPLP-C2-2, SPLP-C2-3, SPLP-C3-2 and SPLP-C3-3 were diluted due to high concentrations.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.
Trip Blank was not provided with this set of samples.
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P5001

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 12/04/2024

Hit Summary Sheet
SW-846

SDG No.: P5001
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: SPLP-C1-1								
P5001-01	SPLP-C1-1	WATER	Acetone	8.70	J	1.40	25.0	ug/L
P5001-01	SPLP-C1-1	WATER	Methyl Acetate	2.20	J	0.60	5.00	ug/L
P5001-01	SPLP-C1-1	WATER	Benzene	4.30	J	0.16	5.00	ug/L
P5001-01	SPLP-C1-1	WATER	Toluene	17.9		0.18	5.00	ug/L
P5001-01	SPLP-C1-1	WATER	Ethyl Benzene	61.2		0.16	5.00	ug/L
P5001-01	SPLP-C1-1	WATER	m/p-Xylenes	30.0		0.31	10.0	ug/L
P5001-01	SPLP-C1-1	WATER	o-Xylene	20.7		0.14	5.00	ug/L
P5001-01	SPLP-C1-1	WATER	Isopropylbenzene	3.40	J	0.13	5.00	ug/L
Total Voc :				148				
Total Concentration:				148				
Client ID: SPLP-C1-1RE								
P5001-01RE	SPLP-C1-1RE	WATER	Acetone	8.80	J	1.40	25.0	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	Methyl Acetate	2.00	J	0.60	5.00	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	Benzene	4.20	J	0.16	5.00	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	Toluene	17.0		0.18	5.00	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	Ethyl Benzene	61.5		0.16	5.00	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	m/p-Xylenes	30.4		0.31	10.0	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	o-Xylene	21.4		0.14	5.00	ug/L
P5001-01RE	SPLP-C1-1RE	WATER	Isopropylbenzene	3.30	J	0.13	5.00	ug/L
Total Voc :				149				
Total Concentration:				149				
Client ID: SPLP-C1-2								
P5001-02	SPLP-C1-2	WATER	Acetone	9.80	J	1.40	25.0	ug/L
P5001-02	SPLP-C1-2	WATER	cis-1,2-Dichloroethene	2.60	J	0.25	5.00	ug/L
P5001-02	SPLP-C1-2	WATER	Benzene	23.6		0.16	5.00	ug/L
P5001-02	SPLP-C1-2	WATER	Toluene	56.6		0.18	5.00	ug/L
P5001-02	SPLP-C1-2	WATER	Ethyl Benzene	140		0.16	5.00	ug/L
P5001-02	SPLP-C1-2	WATER	m/p-Xylenes	66.3		0.31	10.0	ug/L
P5001-02	SPLP-C1-2	WATER	o-Xylene	46.3		0.14	5.00	ug/L
P5001-02	SPLP-C1-2	WATER	Isopropylbenzene	6.10		0.13	5.00	ug/L
Total Voc :				351				
Total Concentration:				351				
Client ID: SPLP-C1-2RE								
P5001-02RE	SPLP-C1-2RE	WATER	Acetone	10.2	J	1.40	25.0	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	cis-1,2-Dichloroethene	2.80	J	0.25	5.00	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	Benzene	24.3		0.16	5.00	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	Toluene	57.6		0.18	5.00	ug/L

Hit Summary Sheet

SW-846

SDG No.: P5001

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5001-02RE	SPLP-C1-2RE	WATER	Ethyl Benzene	130		0.16	5.00	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	m/p-Xylenes	64.5		0.31	10.0	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	o-Xylene	45.4		0.14	5.00	ug/L
P5001-02RE	SPLP-C1-2RE	WATER	Isopropylbenzene	6.30		0.13	5.00	ug/L
Total Voc :				341				
Total Concentration:				341				
Client ID:	SPLP-C1-3							
P5001-03	SPLP-C1-3	WATER	Acetone	10.3	J	1.40	25.0	ug/L
P5001-03	SPLP-C1-3	WATER	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
P5001-03	SPLP-C1-3	WATER	Benzene	10.3		0.16	5.00	ug/L
P5001-03	SPLP-C1-3	WATER	Toluene	41.7		0.18	5.00	ug/L
P5001-03	SPLP-C1-3	WATER	Ethyl Benzene	130		0.16	5.00	ug/L
P5001-03	SPLP-C1-3	WATER	m/p-Xylenes	62.6		0.31	10.0	ug/L
P5001-03	SPLP-C1-3	WATER	o-Xylene	43.6		0.14	5.00	ug/L
P5001-03	SPLP-C1-3	WATER	Isopropylbenzene	6.20		0.13	5.00	ug/L
Total Voc :				306				
Total Concentration:				306				
Client ID:	SPLP-C1-3RE							
P5001-03RE	SPLP-C1-3RE	WATER	Acetone	10.6	J	1.40	25.0	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	Benzene	9.80		0.16	5.00	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	Toluene	39.5		0.18	5.00	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	Ethyl Benzene	120		0.16	5.00	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	m/p-Xylenes	59.0		0.31	10.0	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	o-Xylene	41.0		0.14	5.00	ug/L
P5001-03RE	SPLP-C1-3RE	WATER	Isopropylbenzene	5.70		0.13	5.00	ug/L
Total Voc :				287				
Total Concentration:				287				
Client ID:	SPLP-C2-1							
P5001-04	SPLP-C2-1	WATER	Acetone	28.6		1.40	25.0	ug/L
P5001-04	SPLP-C2-1	WATER	cis-1,2-Dichloroethene	54.4		0.25	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	Benzene	8.10		0.16	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	Trichloroethene	21.2		0.32	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	Toluene	16.9		0.18	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	Ethyl Benzene	44.7		0.16	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	m/p-Xylenes	20.1		0.31	10.0	ug/L
P5001-04	SPLP-C2-1	WATER	o-Xylene	20.7		0.14	5.00	ug/L
P5001-04	SPLP-C2-1	WATER	Isopropylbenzene	3.00	J	0.13	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5001

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Voc :				218				
Total Concentration:				218				
Client ID:	SPLP-C2-1RE							
P5001-04RE	SPLP-C2-1RE	WATER	Acetone	28.0		1.40	25.0	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	cis-1,2-Dichloroethene	52.9		0.25	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	Benzene	7.50		0.16	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	Trichloroethene	19.4		0.32	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	Toluene	15.6		0.18	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	Ethyl Benzene	43.2		0.16	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	m/p-Xylenes	19.1		0.31	10.0	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	o-Xylene	20.3		0.14	5.00	ug/L
P5001-04RE	SPLP-C2-1RE	WATER	Isopropylbenzene	2.90	J	0.13	5.00	ug/L
Total Voc :				209				
Total Concentration:				209				
Client ID:	SPLP-C2-2							
P5001-05	SPLP-C2-2	WATER	Vinyl Chloride	1.00	J	0.34	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Acetone	22.7	J	1.40	25.0	ug/L
P5001-05	SPLP-C2-2	WATER	cis-1,2-Dichloroethene	100		0.25	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Benzene	45.4		0.16	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Trichloroethene	12.2		0.32	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Toluene	89.0		0.18	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Ethyl Benzene	250	E	0.16	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	m/p-Xylenes	110		0.31	10.0	ug/L
P5001-05	SPLP-C2-2	WATER	o-Xylene	84.2		0.14	5.00	ug/L
P5001-05	SPLP-C2-2	WATER	Isopropylbenzene	9.10		0.13	5.00	ug/L
Total Voc :				724				
Total Concentration:				724				
Client ID:	SPLP-C2-2DL							
P5001-05DL	SPLP-C2-2DL	WATER	cis-1,2-Dichloroethene	96.8	D	2.50	50.0	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	Benzene	41.7	JD	1.60	50.0	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	Trichloroethene	11.6	JD	3.20	50.0	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	Toluene	78.8	D	1.80	50.0	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	Ethyl Benzene	210	D	1.60	50.0	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	m/p-Xylenes	90.8	JD	3.10	100	ug/L
P5001-05DL	SPLP-C2-2DL	WATER	o-Xylene	74.2	D	1.40	50.0	ug/L
Total Voc :				604				
Total Concentration:				604				
Client ID:	SPLP-C2-3							

Hit Summary Sheet
SW-846

SDG No.: P5001

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5001-06	SPLP-C2-3	WATER	Vinyl Chloride	6.00		0.34	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Acetone	18.5	J	1.40	25.0	ug/L
P5001-06	SPLP-C2-3	WATER	cis-1,2-Dichloroethene	160	E	0.25	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Benzene	100		0.16	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Trichloroethene	16.9		0.32	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Toluene	140		0.18	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Ethyl Benzene	430	E	0.16	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	m/p-Xylenes	170		0.31	10.0	ug/L
P5001-06	SPLP-C2-3	WATER	o-Xylene	130		0.14	5.00	ug/L
P5001-06	SPLP-C2-3	WATER	Isopropylbenzene	17.5		0.13	5.00	ug/L
Total Voc :				1190				
Total Concentration:				1190				
Client ID:	SPLP-C2-3DL							
P5001-06DL	SPLP-C2-3DL	WATER	cis-1,2-Dichloroethene	150	D	2.50	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	Benzene	100	D	1.60	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	Trichloroethene	17.1	JD	3.20	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	Toluene	140	D	1.80	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	Ethyl Benzene	410	D	1.60	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	m/p-Xylenes	150	D	3.10	100	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	o-Xylene	130	D	1.40	50.0	ug/L
P5001-06DL	SPLP-C2-3DL	WATER	Isopropylbenzene	16.3	JD	1.30	50.0	ug/L
Total Voc :				1110				
Total Concentration:				1110				
Client ID:	SPLP-C3-1							
P5001-08	SPLP-C3-1	WATER	Acetone	17.4	J	1.40	25.0	ug/L
P5001-08	SPLP-C3-1	WATER	cis-1,2-Dichloroethene	7.50		0.25	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	Benzene	2.70	J	0.16	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	Trichloroethene	2.10	J	0.32	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	Toluene	8.70		0.18	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	Ethyl Benzene	33.8		0.16	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	m/p-Xylenes	15.6		0.31	10.0	ug/L
P5001-08	SPLP-C3-1	WATER	o-Xylene	13.5		0.14	5.00	ug/L
P5001-08	SPLP-C3-1	WATER	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
Total Voc :				104				
Total Concentration:				104				
Client ID:	SPLP-C3-1RE							
P5001-08RE	SPLP-C3-1RE	WATER	Acetone	19.0	J	1.40	25.0	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	cis-1,2-Dichloroethene	7.20		0.25	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5001

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5001-08RE	SPLP-C3-1RE	WATER	Benzene	2.70	J	0.16	5.00	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	Trichloroethene	2.00	J	0.32	5.00	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	Toluene	8.50		0.18	5.00	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	Ethyl Benzene	33.1		0.16	5.00	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	m/p-Xylenes	15.1		0.31	10.0	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	o-Xylene	13.3		0.14	5.00	ug/L
P5001-08RE	SPLP-C3-1RE	WATER	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
Total Voc :				103				
Total Concentration:				103				
Client ID:	SPLP-C3-2							
P5001-09	SPLP-C3-2	WATER	Acetone	46.4		1.40	25.0	ug/L
P5001-09	SPLP-C3-2	WATER	2-Butanone	7.40	J	1.30	25.0	ug/L
P5001-09	SPLP-C3-2	WATER	cis-1,2-Dichloroethene	69.3		0.25	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	Benzene	92.6		0.16	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	Trichloroethene	1.80	J	0.32	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	Toluene	270	E	0.18	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	Ethyl Benzene	460	E	0.16	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	m/p-Xylenes	220		0.31	10.0	ug/L
P5001-09	SPLP-C3-2	WATER	o-Xylene	150	E	0.14	5.00	ug/L
P5001-09	SPLP-C3-2	WATER	Isopropylbenzene	17.8		0.13	5.00	ug/L
Total Voc :				1340				
Total Concentration:				1340				
Client ID:	SPLP-C3-2DL							
P5001-09DL	SPLP-C3-2DL	WATER	cis-1,2-Dichloroethene	58.3	D	2.50	50.0	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	Benzene	82.4	D	1.60	50.0	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	Toluene	230	D	1.80	50.0	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	Ethyl Benzene	380	D	1.60	50.0	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	m/p-Xylenes	180	D	3.10	100	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	o-Xylene	130	D	1.40	50.0	ug/L
P5001-09DL	SPLP-C3-2DL	WATER	Isopropylbenzene	14.7	JD	1.30	50.0	ug/L
Total Voc :				1080				
Total Concentration:				1080				
Client ID:	SPLP-C3-3							
P5001-10	SPLP-C3-3	WATER	Acetone	12.1	J	1.40	25.0	ug/L
P5001-10	SPLP-C3-3	WATER	cis-1,2-Dichloroethene	8.00		0.25	5.00	ug/L
P5001-10	SPLP-C3-3	WATER	Benzene	52.0		0.16	5.00	ug/L
P5001-10	SPLP-C3-3	WATER	Toluene	74.3		0.18	5.00	ug/L
P5001-10	SPLP-C3-3	WATER	Ethyl Benzene	320	E	0.16	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5001
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5001-10	SPLP-C3-3	WATER	m/p-Xylenes	120		0.31	10.0	ug/L
P5001-10	SPLP-C3-3	WATER	o-Xylene	100		0.14	5.00	ug/L
P5001-10	SPLP-C3-3	WATER	Isopropylbenzene	14.1		0.13	5.00	ug/L
Total Voc :				701				
Total Concentration:				701				
Client ID:	SPLP-C3-3DL							
P5001-10DL	SPLP-C3-3DL	WATER	Benzene	44.4	JD	1.60	50.0	ug/L
P5001-10DL	SPLP-C3-3DL	WATER	Toluene	60.1	D	1.80	50.0	ug/L
P5001-10DL	SPLP-C3-3DL	WATER	Ethyl Benzene	250	D	1.60	50.0	ug/L
P5001-10DL	SPLP-C3-3DL	WATER	m/p-Xylenes	97.6	JD	3.10	100	ug/L
P5001-10DL	SPLP-C3-3DL	WATER	o-Xylene	81.9	D	1.40	50.0	ug/L
P5001-10DL	SPLP-C3-3DL	WATER	Isopropylbenzene	10.3	JD	1.30	50.0	ug/L
Total Voc :				544				
Total Concentration:				544				
Client ID:	SPLP-C4-1							
P5001-11	SPLP-C4-1	WATER	Acetone	17.8	J	1.40	25.0	ug/L
P5001-11	SPLP-C4-1	WATER	cis-1,2-Dichloroethene	4.50	J	0.25	5.00	ug/L
P5001-11	SPLP-C4-1	WATER	Ethyl Benzene	3.30	J	0.16	5.00	ug/L
P5001-11	SPLP-C4-1	WATER	o-Xylene	1.20	J	0.14	5.00	ug/L
Total Voc :				26.8				
Total Concentration:				26.8				
Client ID:	SPLP-C4-1RE							
P5001-11RE	SPLP-C4-1RE	WATER	Acetone	16.9	J	1.40	25.0	ug/L
P5001-11RE	SPLP-C4-1RE	WATER	cis-1,2-Dichloroethene	4.30	J	0.25	5.00	ug/L
P5001-11RE	SPLP-C4-1RE	WATER	Ethyl Benzene	2.30	J	0.16	5.00	ug/L
Total Voc :				23.5				
Total Concentration:				23.5				



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-1	SDG No.:	P5001
Lab Sample ID:	P5001-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044031.D	1		11/27/24 11:36	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	8.70	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	2.20	J	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	4.30	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	17.9		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-1	SDG No.:	P5001
Lab Sample ID:	P5001-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044031.D	1		11/27/24 11:36	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	61.2		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	30.0		0.31	10.0	ug/L
95-47-6	o-Xylene	20.7		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	3.40	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.6		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	34.8		70 (75) - 130 (124)	70%	SPK: 50
2037-26-5	Toluene-d8	51.6		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	113000	5.544			
540-36-3	1,4-Difluorobenzene	218000	6.757			
3114-55-4	Chlorobenzene-d5	202000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	89400	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-1		SDG No.:	P5001	
Lab Sample ID:	P5001-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044031.D	1		11/27/24 11:36	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044069.D	1		12/02/24 16:24	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	8.80	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	2.00	J	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	4.20	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	17.0		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044069.D	1		12/02/24 16:24	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	61.5		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	30.4		0.31	10.0	ug/L
95-47-6	o-Xylene	21.4		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	3.30	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.7		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	33.1	*	70 (75) - 130 (124)	66%	SPK: 50
2037-26-5	Toluene-d8	48.9		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.55			
540-36-3	1,4-Difluorobenzene	232000	6.757			
3114-55-4	Chlorobenzene-d5	204000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	98500	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044069.D	1		12/02/24 16:24	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-2	SDG No.:	P5001
Lab Sample ID:	P5001-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044032.D	1		11/27/24 11:59	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	9.80	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.60	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	23.6		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	56.6		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-2	SDG No.:	P5001
Lab Sample ID:	P5001-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044032.D	1		11/27/24 11:59	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	140		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	66.3		0.31	10.0	ug/L
95-47-6	o-Xylene	46.3		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	6.10		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.2		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	33.8	*	70 (75) - 130 (124)	68%	SPK: 50
2037-26-5	Toluene-d8	49.4		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.3		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	123000	5.543			
540-36-3	1,4-Difluorobenzene	238000	6.757			
3114-55-4	Chlorobenzene-d5	208000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	98100	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-2		SDG No.:	P5001	
Lab Sample ID:	P5001-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044032.D	1		11/27/24 11:59	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-2RE	SDG No.:	P5001
Lab Sample ID:	P5001-02RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044070.D	1		12/02/24 16:47	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	10.2	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.80	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	24.3		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	57.6		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-2RE		SDG No.:	P5001	
Lab Sample ID:	P5001-02RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044070.D	1		12/02/24 16:47	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	130		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	64.5		0.31	10.0	ug/L
95-47-6	o-Xylene	45.4		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	6.30		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.0		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	33.7	*	70 (75) - 130 (124)	67%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	120000	5.543			
540-36-3	1,4-Difluorobenzene	236000	6.757			
3114-55-4	Chlorobenzene-d5	212000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	96600	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-2RE		SDG No.:	P5001	
Lab Sample ID:	P5001-02RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044070.D	1		12/02/24 16:47	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-3	SDG No.:	P5001
Lab Sample ID:	P5001-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044033.D	1		11/27/24 12:22	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	10.3	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	10.3		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	41.7		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-3		SDG No.:	P5001	
Lab Sample ID:	P5001-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044033.D	1		11/27/24 12:22	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	130		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	62.6		0.31	10.0	ug/L
95-47-6	o-Xylene	43.6		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	6.20		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.4		70 (74) - 130 (125)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	31.5	*	70 (75) - 130 (124)	63%	SPK: 50
2037-26-5	Toluene-d8	50.2		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	121000	5.55			
540-36-3	1,4-Difluorobenzene	227000	6.757			
3114-55-4	Chlorobenzene-d5	204000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	99000	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-3		SDG No.:	P5001	
Lab Sample ID:	P5001-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044033.D	1		11/27/24 12:22	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-3RE	SDG No.:	P5001
Lab Sample ID:	P5001-03RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044071.D	1		12/02/24 17:10	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	10.6	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	9.80		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	39.5		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C1-3RE	SDG No.:	P5001
Lab Sample ID:	P5001-03RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044071.D	1		12/02/24 17:10	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	120		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	59.0		0.31	10.0	ug/L
95-47-6	o-Xylene	41.0		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	5.70		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.9		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	31.7	*	70 (75) - 130 (124)	63%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	122000	5.544			
540-36-3	1,4-Difluorobenzene	237000	6.757			
3114-55-4	Chlorobenzene-d5	212000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	101000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C1-3RE		SDG No.:	P5001	
Lab Sample ID:	P5001-03RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044071.D	1		12/02/24 17:10	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-1	SDG No.:	P5001
Lab Sample ID:	P5001-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044034.D	1		11/27/24 12:45	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	28.6		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	54.4		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	8.10		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	21.2		0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	16.9		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-1	SDG No.:	P5001
Lab Sample ID:	P5001-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044034.D	1		11/27/24 12:45	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	44.7		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	20.1		0.31	10.0	ug/L
95-47-6	o-Xylene	20.7		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	3.00	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.7		70 (74) - 130 (125)	97%	SPK: 50
1868-53-7	Dibromofluoromethane	26.5	*	70 (75) - 130 (124)	53%	SPK: 50
2037-26-5	Toluene-d8	50.5		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.4		70 (77) - 130 (121)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	117000	5.544			
540-36-3	1,4-Difluorobenzene	223000	6.757			
3114-55-4	Chlorobenzene-d5	200000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	95800	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-1		SDG No.:	P5001	
Lab Sample ID:	P5001-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044034.D	1		11/27/24 12:45	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-1RE	SDG No.:	P5001
Lab Sample ID:	P5001-04RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044072.D	1		12/02/24 17:32	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	28.0		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	52.9		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	7.50		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	19.4		0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	15.6		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-1RE	SDG No.:	P5001
Lab Sample ID:	P5001-04RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044072.D	1		12/02/24 17:32	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	43.2		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	19.1		0.31	10.0	ug/L
95-47-6	o-Xylene	20.3		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	2.90	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.7		70 (74) - 130 (125)	101%	SPK: 50
1868-53-7	Dibromofluoromethane	28.9	*	70 (75) - 130 (124)	58%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.5		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	119000	5.544			
540-36-3	1,4-Difluorobenzene	231000	6.757			
3114-55-4	Chlorobenzene-d5	205000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	97300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-04RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044072.D	1		12/02/24 17:32	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-2	SDG No.:	P5001
Lab Sample ID:	P5001-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044035.D	1		11/27/24 13:08	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	1.00	J	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	22.7	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	100		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	45.4		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	12.2		0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	89.0		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-2	SDG No.:	P5001
Lab Sample ID:	P5001-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044035.D	1		11/27/24 13:08	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	250	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	110		0.31	10.0	ug/L
95-47-6	o-Xylene	84.2		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	9.10		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.7		70 (74) - 130 (125)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	29.9	*	70 (75) - 130 (124)	60%	SPK: 50
2037-26-5	Toluene-d8	49.0		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.3		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	117000	5.543			
540-36-3	1,4-Difluorobenzene	228000	6.757			
3114-55-4	Chlorobenzene-d5	194000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	96300	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-2		SDG No.:	P5001	
Lab Sample ID:	P5001-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044035.D	1		11/27/24 13:08	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-2DL	SDG No.:	P5001
Lab Sample ID:	P5001-05DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044065.D	10		12/02/24 14:52	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.10	UD	2.10	50.0	ug/L
74-87-3	Chloromethane	3.50	UD	3.50	50.0	ug/L
75-01-4	Vinyl Chloride	3.40	UD	3.40	50.0	ug/L
74-83-9	Bromomethane	13.6	UD	13.6	50.0	ug/L
75-00-3	Chloroethane	5.60	UD	5.60	50.0	ug/L
75-69-4	Trichlorofluoromethane	3.40	UD	3.40	50.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	UD	2.50	50.0	ug/L
75-35-4	1,1-Dichloroethene	2.60	UD	2.60	50.0	ug/L
67-64-1	Acetone	13.9	UD	13.9	250	ug/L
75-15-0	Carbon Disulfide	3.20	UD	3.20	50.0	ug/L
1634-04-4	Methyl tert-butyl Ether	1.60	UD	1.60	50.0	ug/L
79-20-9	Methyl Acetate	6.00	UD	6.00	50.0	ug/L
75-09-2	Methylene Chloride	3.20	UD	3.20	50.0	ug/L
156-60-5	trans-1,2-Dichloroethene	2.50	UD	2.50	50.0	ug/L
75-34-3	1,1-Dichloroethane	2.30	UD	2.30	50.0	ug/L
110-82-7	Cyclohexane	16.2	UD	16.2	50.0	ug/L
78-93-3	2-Butanone	13.0	UD	13.0	250	ug/L
56-23-5	Carbon Tetrachloride	2.50	UD	2.50	50.0	ug/L
156-59-2	cis-1,2-Dichloroethene	96.8	D	2.50	50.0	ug/L
74-97-5	Bromochloromethane	1.80	UD	1.80	50.0	ug/L
67-66-3	Chloroform	2.60	UD	2.60	50.0	ug/L
71-55-6	1,1,1-Trichloroethane	1.90	UD	1.90	50.0	ug/L
108-87-2	Methylcyclohexane	1.90	UD	1.90	50.0	ug/L
71-43-2	Benzene	41.7	JD	1.60	50.0	ug/L
107-06-2	1,2-Dichloroethane	2.40	UD	2.40	50.0	ug/L
79-01-6	Trichloroethene	11.6	JD	3.20	50.0	ug/L
78-87-5	1,2-Dichloropropane	1.90	UD	1.90	50.0	ug/L
75-27-4	Bromodichloromethane	2.40	UD	2.40	50.0	ug/L
108-10-1	4-Methyl-2-Pentanone	7.50	UD	7.50	250	ug/L
108-88-3	Toluene	78.8	D	1.80	50.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-2DL	SDG No.:	P5001
Lab Sample ID:	P5001-05DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044065.D	10		12/02/24 14:52	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	2.10	UD	2.10	50.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	1.80	UD	1.80	50.0	ug/L
79-00-5	1,1,2-Trichloroethane	2.10	UD	2.10	50.0	ug/L
591-78-6	2-Hexanone	11.3	UD	11.3	250	ug/L
124-48-1	Dibromochloromethane	1.80	UD	1.80	50.0	ug/L
106-93-4	1,2-Dibromoethane	1.60	UD	1.60	50.0	ug/L
127-18-4	Tetrachloroethene	2.50	UD	2.50	50.0	ug/L
108-90-7	Chlorobenzene	1.30	UD	1.30	50.0	ug/L
100-41-4	Ethyl Benzene	210	D	1.60	50.0	ug/L
179601-23-1	m/p-Xylenes	90.8	JD	3.10	100	ug/L
95-47-6	o-Xylene	74.2	D	1.40	50.0	ug/L
100-42-5	Styrene	1.60	UD	1.60	50.0	ug/L
75-25-2	Bromoform	2.10	UD	2.10	50.0	ug/L
98-82-8	Isopropylbenzene	1.30	UD	1.30	50.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.70	UD	2.70	50.0	ug/L
541-73-1	1,3-Dichlorobenzene	2.40	UD	2.40	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	2.70	UD	2.70	50.0	ug/L
95-50-1	1,2-Dichlorobenzene	1.90	UD	1.90	50.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	UD	4.60	50.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	4.20	UD	4.20	50.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	5.10	UD	5.10	50.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.9		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	45.7		70 (75) - 130 (124)	91%	SPK: 50
2037-26-5	Toluene-d8	49.3		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	123000	5.544			
540-36-3	1,4-Difluorobenzene	242000	6.757			
3114-55-4	Chlorobenzene-d5	211000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	99300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-2DL		SDG No.:	P5001	
Lab Sample ID:	P5001-05DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044065.D	10		12/02/24 14:52	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-3	SDG No.:	P5001
Lab Sample ID:	P5001-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044036.D	1		11/27/24 13:31	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	6.00		0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	18.5	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	160	E	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	100		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	16.9		0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	140		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-3	SDG No.:	P5001
Lab Sample ID:	P5001-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044036.D	1		11/27/24 13:31	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	430	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	170		0.31	10.0	ug/L
95-47-6	o-Xylene	130		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	17.5		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.8		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	30.8	*	70 (75) - 130 (124)	62%	SPK: 50
2037-26-5	Toluene-d8	50.5		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.1		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.55			
540-36-3	1,4-Difluorobenzene	225000	6.757			
3114-55-4	Chlorobenzene-d5	200000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	94700	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-3		SDG No.:	P5001	
Lab Sample ID:	P5001-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044036.D	1		11/27/24 13:31	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-3DL	SDG No.:	P5001
Lab Sample ID:	P5001-06DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044066.D	10		12/02/24 15:15	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.10	UD	2.10	50.0	ug/L
74-87-3	Chloromethane	3.50	UD	3.50	50.0	ug/L
75-01-4	Vinyl Chloride	3.40	UD	3.40	50.0	ug/L
74-83-9	Bromomethane	13.6	UD	13.6	50.0	ug/L
75-00-3	Chloroethane	5.60	UD	5.60	50.0	ug/L
75-69-4	Trichlorofluoromethane	3.40	UD	3.40	50.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	UD	2.50	50.0	ug/L
75-35-4	1,1-Dichloroethene	2.60	UD	2.60	50.0	ug/L
67-64-1	Acetone	13.9	UD	13.9	250	ug/L
75-15-0	Carbon Disulfide	3.20	UD	3.20	50.0	ug/L
1634-04-4	Methyl tert-butyl Ether	1.60	UD	1.60	50.0	ug/L
79-20-9	Methyl Acetate	6.00	UD	6.00	50.0	ug/L
75-09-2	Methylene Chloride	3.20	UD	3.20	50.0	ug/L
156-60-5	trans-1,2-Dichloroethene	2.50	UD	2.50	50.0	ug/L
75-34-3	1,1-Dichloroethane	2.30	UD	2.30	50.0	ug/L
110-82-7	Cyclohexane	16.2	UD	16.2	50.0	ug/L
78-93-3	2-Butanone	13.0	UD	13.0	250	ug/L
56-23-5	Carbon Tetrachloride	2.50	UD	2.50	50.0	ug/L
156-59-2	cis-1,2-Dichloroethene	150	D	2.50	50.0	ug/L
74-97-5	Bromochloromethane	1.80	UD	1.80	50.0	ug/L
67-66-3	Chloroform	2.60	UD	2.60	50.0	ug/L
71-55-6	1,1,1-Trichloroethane	1.90	UD	1.90	50.0	ug/L
108-87-2	Methylcyclohexane	1.90	UD	1.90	50.0	ug/L
71-43-2	Benzene	100	D	1.60	50.0	ug/L
107-06-2	1,2-Dichloroethane	2.40	UD	2.40	50.0	ug/L
79-01-6	Trichloroethene	17.1	JD	3.20	50.0	ug/L
78-87-5	1,2-Dichloropropane	1.90	UD	1.90	50.0	ug/L
75-27-4	Bromodichloromethane	2.40	UD	2.40	50.0	ug/L
108-10-1	4-Methyl-2-Pentanone	7.50	UD	7.50	250	ug/L
108-88-3	Toluene	140	D	1.80	50.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C2-3DL	SDG No.:	P5001
Lab Sample ID:	P5001-06DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044066.D	10		12/02/24 15:15	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	2.10	UD	2.10	50.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	1.80	UD	1.80	50.0	ug/L
79-00-5	1,1,2-Trichloroethane	2.10	UD	2.10	50.0	ug/L
591-78-6	2-Hexanone	11.3	UD	11.3	250	ug/L
124-48-1	Dibromochloromethane	1.80	UD	1.80	50.0	ug/L
106-93-4	1,2-Dibromoethane	1.60	UD	1.60	50.0	ug/L
127-18-4	Tetrachloroethene	2.50	UD	2.50	50.0	ug/L
108-90-7	Chlorobenzene	1.30	UD	1.30	50.0	ug/L
100-41-4	Ethyl Benzene	410	D	1.60	50.0	ug/L
179601-23-1	m/p-Xylenes	150	D	3.10	100	ug/L
95-47-6	o-Xylene	130	D	1.40	50.0	ug/L
100-42-5	Styrene	1.60	UD	1.60	50.0	ug/L
75-25-2	Bromoform	2.10	UD	2.10	50.0	ug/L
98-82-8	Isopropylbenzene	16.3	JD	1.30	50.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.70	UD	2.70	50.0	ug/L
541-73-1	1,3-Dichlorobenzene	2.40	UD	2.40	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	2.70	UD	2.70	50.0	ug/L
95-50-1	1,2-Dichlorobenzene	1.90	UD	1.90	50.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	UD	4.60	50.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	4.20	UD	4.20	50.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	5.10	UD	5.10	50.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.2		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	45.6		70 (75) - 130 (124)	91%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	120000	5.55			
540-36-3	1,4-Difluorobenzene	232000	6.757			
3114-55-4	Chlorobenzene-d5	204000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	95400	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C2-3DL		SDG No.:	P5001	
Lab Sample ID:	P5001-06DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044066.D	10		12/02/24 15:15	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-1	SDG No.:	P5001
Lab Sample ID:	P5001-08	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044037.D	1		11/27/24 13:54	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.4	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	7.50		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.70	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	2.10	J	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	8.70		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-1	SDG No.:	P5001
Lab Sample ID:	P5001-08	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044037.D	1		11/27/24 13:54	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	33.8		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	15.6		0.31	10.0	ug/L
95-47-6	o-Xylene	13.5		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.6		70 (74) - 130 (125)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	29.9	*	70 (75) - 130 (124)	60%	SPK: 50
2037-26-5	Toluene-d8	49.4		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.544			
540-36-3	1,4-Difluorobenzene	227000	6.757			
3114-55-4	Chlorobenzene-d5	201000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	94800	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-1		SDG No.:	P5001	
Lab Sample ID:	P5001-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044037.D	1		11/27/24 13:54	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-1RE	SDG No.:	P5001
Lab Sample ID:	P5001-08RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044073.D	1		12/02/24 17:55	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	19.0	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	7.20		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.70	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	2.00	J	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	8.50		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-1RE	SDG No.:	P5001
Lab Sample ID:	P5001-08RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044073.D	1		12/02/24 17:55	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	33.1		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	15.1		0.31	10.0	ug/L
95-47-6	o-Xylene	13.3		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.1		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	30.4	*	70 (75) - 130 (124)	61%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	114000	5.55			
540-36-3	1,4-Difluorobenzene	222000	6.757			
3114-55-4	Chlorobenzene-d5	199000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	93000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-08RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044073.D	1		12/02/24 17:55	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-2	SDG No.:	P5001
Lab Sample ID:	P5001-09	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044038.D	1		11/27/24 14:16	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	46.4		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	7.40	J	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	69.3		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	92.6		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	1.80	J	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	270	E	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-2	SDG No.:	P5001
Lab Sample ID:	P5001-09	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044038.D	1		11/27/24 14:16	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	460	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	220		0.31	10.0	ug/L
95-47-6	o-Xylene	150	E	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	17.8		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.2		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	29.7	*	70 (75) - 130 (124)	59%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.544			
540-36-3	1,4-Difluorobenzene	228000	6.757			
3114-55-4	Chlorobenzene-d5	201000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	95300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-2		SDG No.:	P5001	
Lab Sample ID:	P5001-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044038.D	1		11/27/24 14:16	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-2DL	SDG No.:	P5001
Lab Sample ID:	P5001-09DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044067.D	10		12/02/24 15:38	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.10	UD	2.10	50.0	ug/L
74-87-3	Chloromethane	3.50	UD	3.50	50.0	ug/L
75-01-4	Vinyl Chloride	3.40	UD	3.40	50.0	ug/L
74-83-9	Bromomethane	13.6	UD	13.6	50.0	ug/L
75-00-3	Chloroethane	5.60	UD	5.60	50.0	ug/L
75-69-4	Trichlorofluoromethane	3.40	UD	3.40	50.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	UD	2.50	50.0	ug/L
75-35-4	1,1-Dichloroethene	2.60	UD	2.60	50.0	ug/L
67-64-1	Acetone	13.9	UD	13.9	250	ug/L
75-15-0	Carbon Disulfide	3.20	UD	3.20	50.0	ug/L
1634-04-4	Methyl tert-butyl Ether	1.60	UD	1.60	50.0	ug/L
79-20-9	Methyl Acetate	6.00	UD	6.00	50.0	ug/L
75-09-2	Methylene Chloride	3.20	UD	3.20	50.0	ug/L
156-60-5	trans-1,2-Dichloroethene	2.50	UD	2.50	50.0	ug/L
75-34-3	1,1-Dichloroethane	2.30	UD	2.30	50.0	ug/L
110-82-7	Cyclohexane	16.2	UD	16.2	50.0	ug/L
78-93-3	2-Butanone	13.0	UD	13.0	250	ug/L
56-23-5	Carbon Tetrachloride	2.50	UD	2.50	50.0	ug/L
156-59-2	cis-1,2-Dichloroethene	58.3	D	2.50	50.0	ug/L
74-97-5	Bromochloromethane	1.80	UD	1.80	50.0	ug/L
67-66-3	Chloroform	2.60	UD	2.60	50.0	ug/L
71-55-6	1,1,1-Trichloroethane	1.90	UD	1.90	50.0	ug/L
108-87-2	Methylcyclohexane	1.90	UD	1.90	50.0	ug/L
71-43-2	Benzene	82.4	D	1.60	50.0	ug/L
107-06-2	1,2-Dichloroethane	2.40	UD	2.40	50.0	ug/L
79-01-6	Trichloroethene	3.20	UD	3.20	50.0	ug/L
78-87-5	1,2-Dichloropropane	1.90	UD	1.90	50.0	ug/L
75-27-4	Bromodichloromethane	2.40	UD	2.40	50.0	ug/L
108-10-1	4-Methyl-2-Pentanone	7.50	UD	7.50	250	ug/L
108-88-3	Toluene	230	D	1.80	50.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-2DL	SDG No.:	P5001
Lab Sample ID:	P5001-09DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044067.D	10		12/02/24 15:38	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	2.10	UD	2.10	50.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	1.80	UD	1.80	50.0	ug/L
79-00-5	1,1,2-Trichloroethane	2.10	UD	2.10	50.0	ug/L
591-78-6	2-Hexanone	11.3	UD	11.3	250	ug/L
124-48-1	Dibromochloromethane	1.80	UD	1.80	50.0	ug/L
106-93-4	1,2-Dibromoethane	1.60	UD	1.60	50.0	ug/L
127-18-4	Tetrachloroethene	2.50	UD	2.50	50.0	ug/L
108-90-7	Chlorobenzene	1.30	UD	1.30	50.0	ug/L
100-41-4	Ethyl Benzene	380	D	1.60	50.0	ug/L
179601-23-1	m/p-Xylenes	180	D	3.10	100	ug/L
95-47-6	o-Xylene	130	D	1.40	50.0	ug/L
100-42-5	Styrene	1.60	UD	1.60	50.0	ug/L
75-25-2	Bromoform	2.10	UD	2.10	50.0	ug/L
98-82-8	Isopropylbenzene	14.7	JD	1.30	50.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.70	UD	2.70	50.0	ug/L
541-73-1	1,3-Dichlorobenzene	2.40	UD	2.40	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	2.70	UD	2.70	50.0	ug/L
95-50-1	1,2-Dichlorobenzene	1.90	UD	1.90	50.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	UD	4.60	50.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	4.20	UD	4.20	50.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	5.10	UD	5.10	50.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.1		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		70 (75) - 130 (124)	95%	SPK: 50
2037-26-5	Toluene-d8	51.2		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	128000	5.544			
540-36-3	1,4-Difluorobenzene	243000	6.757			
3114-55-4	Chlorobenzene-d5	216000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	101000	12.018			

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-2DL	SDG No.:	P5001
Lab Sample ID:	P5001-09DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044067.D	10		12/02/24 15:38	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-3	SDG No.:	P5001
Lab Sample ID:	P5001-10	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044039.D	1		11/27/24 14:39	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	12.1	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	8.00		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	52.0		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	74.3		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-3	SDG No.:	P5001
Lab Sample ID:	P5001-10	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044039.D	1		11/27/24 14:39	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	320	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	120		0.31	10.0	ug/L
95-47-6	o-Xylene	100		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	14.1		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.4		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	30.1	*	70 (75) - 130 (124)	60%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	115000	5.549			
540-36-3	1,4-Difluorobenzene	221000	6.757			
3114-55-4	Chlorobenzene-d5	196000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	90500	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-3		SDG No.:	P5001	
Lab Sample ID:	P5001-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044039.D	1		11/27/24 14:39	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-3DL		SDG No.:	P5001	
Lab Sample ID:	P5001-10DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044068.D	10		12/02/24 16:01	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.10	UD	2.10	50.0	ug/L
74-87-3	Chloromethane	3.50	UD	3.50	50.0	ug/L
75-01-4	Vinyl Chloride	3.40	UD	3.40	50.0	ug/L
74-83-9	Bromomethane	13.6	UD	13.6	50.0	ug/L
75-00-3	Chloroethane	5.60	UD	5.60	50.0	ug/L
75-69-4	Trichlorofluoromethane	3.40	UD	3.40	50.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	UD	2.50	50.0	ug/L
75-35-4	1,1-Dichloroethene	2.60	UD	2.60	50.0	ug/L
67-64-1	Acetone	13.9	UD	13.9	250	ug/L
75-15-0	Carbon Disulfide	3.20	UD	3.20	50.0	ug/L
1634-04-4	Methyl tert-butyl Ether	1.60	UD	1.60	50.0	ug/L
79-20-9	Methyl Acetate	6.00	UD	6.00	50.0	ug/L
75-09-2	Methylene Chloride	3.20	UD	3.20	50.0	ug/L
156-60-5	trans-1,2-Dichloroethene	2.50	UD	2.50	50.0	ug/L
75-34-3	1,1-Dichloroethane	2.30	UD	2.30	50.0	ug/L
110-82-7	Cyclohexane	16.2	UD	16.2	50.0	ug/L
78-93-3	2-Butanone	13.0	UD	13.0	250	ug/L
56-23-5	Carbon Tetrachloride	2.50	UD	2.50	50.0	ug/L
156-59-2	cis-1,2-Dichloroethene	2.50	UD	2.50	50.0	ug/L
74-97-5	Bromochloromethane	1.80	UD	1.80	50.0	ug/L
67-66-3	Chloroform	2.60	UD	2.60	50.0	ug/L
71-55-6	1,1,1-Trichloroethane	1.90	UD	1.90	50.0	ug/L
108-87-2	Methylcyclohexane	1.90	UD	1.90	50.0	ug/L
71-43-2	Benzene	44.4	JD	1.60	50.0	ug/L
107-06-2	1,2-Dichloroethane	2.40	UD	2.40	50.0	ug/L
79-01-6	Trichloroethene	3.20	UD	3.20	50.0	ug/L
78-87-5	1,2-Dichloropropane	1.90	UD	1.90	50.0	ug/L
75-27-4	Bromodichloromethane	2.40	UD	2.40	50.0	ug/L
108-10-1	4-Methyl-2-Pentanone	7.50	UD	7.50	250	ug/L
108-88-3	Toluene	60.1	D	1.80	50.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C3-3DL	SDG No.:	P5001
Lab Sample ID:	P5001-10DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044068.D	10		12/02/24 16:01	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	2.10	UD	2.10	50.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	1.80	UD	1.80	50.0	ug/L
79-00-5	1,1,2-Trichloroethane	2.10	UD	2.10	50.0	ug/L
591-78-6	2-Hexanone	11.3	UD	11.3	250	ug/L
124-48-1	Dibromochloromethane	1.80	UD	1.80	50.0	ug/L
106-93-4	1,2-Dibromoethane	1.60	UD	1.60	50.0	ug/L
127-18-4	Tetrachloroethene	2.50	UD	2.50	50.0	ug/L
108-90-7	Chlorobenzene	1.30	UD	1.30	50.0	ug/L
100-41-4	Ethyl Benzene	250	D	1.60	50.0	ug/L
179601-23-1	m/p-Xylenes	97.6	JD	3.10	100	ug/L
95-47-6	o-Xylene	81.9	D	1.40	50.0	ug/L
100-42-5	Styrene	1.60	UD	1.60	50.0	ug/L
75-25-2	Bromoform	2.10	UD	2.10	50.0	ug/L
98-82-8	Isopropylbenzene	10.3	JD	1.30	50.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.70	UD	2.70	50.0	ug/L
541-73-1	1,3-Dichlorobenzene	2.40	UD	2.40	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	2.70	UD	2.70	50.0	ug/L
95-50-1	1,2-Dichlorobenzene	1.90	UD	1.90	50.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	4.60	UD	4.60	50.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	4.20	UD	4.20	50.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	5.10	UD	5.10	50.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.2		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	45.2		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	49.9		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	126000	5.544			
540-36-3	1,4-Difluorobenzene	247000	6.757			
3114-55-4	Chlorobenzene-d5	215000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	102000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C3-3DL		SDG No.:	P5001	
Lab Sample ID:	P5001-10DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044068.D	10		12/02/24 16:01	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C4-1	SDG No.:	P5001
Lab Sample ID:	P5001-11	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044040.D	1		11/27/24 15:02	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.8	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	4.50	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C4-1	SDG No.:	P5001
Lab Sample ID:	P5001-11	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044040.D	1		11/27/24 15:02	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	3.30	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	1.20	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.1		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	33.2	*	70 (75) - 130 (124)	66%	SPK: 50
2037-26-5	Toluene-d8	50.5		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		70 (77) - 130 (121)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	114000	5.544			
540-36-3	1,4-Difluorobenzene	223000	6.757			
3114-55-4	Chlorobenzene-d5	203000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	88700	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C4-1		SDG No.:	P5001	
Lab Sample ID:	P5001-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044040.D	1		11/27/24 15:02	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/25/24
Client Sample ID:	SPLP-C4-1RE	SDG No.:	P5001
Lab Sample ID:	P5001-11RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044074.D	1		12/02/24 18:18	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	16.9	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	4.30	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/25/24	
Client Sample ID:	SPLP-C4-1RE		SDG No.:	P5001	
Lab Sample ID:	P5001-11RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044074.D	1		12/02/24 18:18	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	2.30	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	U	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	34.5	*	70 (75) - 130 (124)	69%	SPK: 50
2037-26-5	Toluene-d8	49.4		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	119000	5.55			
540-36-3	1,4-Difluorobenzene	232000	6.757			
3114-55-4	Chlorobenzene-d5	201000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	88000	12.018			

Report of Analysis

Client:	ENTACT			Date Collected:	11/25/24		
Project:	540 Degraw St, Brooklyn, NY - E9309			Date Received:	11/25/24		
Client Sample ID:	SPLP-C4-1RE			SDG No.:	P5001		
Lab Sample ID:	P5001-11RE			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	SPLP VOA		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044074.D	1		12/02/24 18:18	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	P5001	OrderDate:	11/25/2024 3:29:57 PM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	L51

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5001-01	SPLP-C1-1	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-01RE	SPLP-C1-1RE	Water			11/25/24			11/25/24
			SPLP VOA	8260D			12/02/24	
P5001-02	SPLP-C1-2	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-02RE	SPLP-C1-2RE	Water			11/25/24			11/25/24
			SPLP VOA	8260D			12/02/24	
P5001-03	SPLP-C1-3	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-03RE	SPLP-C1-3RE	Water			11/25/24			11/25/24
			SPLP VOA	8260D			12/02/24	
P5001-04	SPLP-C2-1	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-04RE	SPLP-C2-1RE	Water			11/25/24			11/25/24
			SPLP VOA	8260D			12/02/24	
P5001-05	SPLP-C2-2	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-05DL	SPLP-C2-2DL	Water			11/25/24			11/25/24
			SPLP VOA	8260D			12/02/24	
P5001-06	SPLP-C2-3	Water			11/25/24			11/25/24
			SPLP VOA	8260D			11/27/24	
P5001-06DL	SPLP-C2-3DL	Water			11/25/24			11/25/24

LAB CHRONICLE

			SPLP VOA	8260D	12/02/24	
P5001-08	SPLP-C3-1	Water			11/25/24	11/25/24
			SPLP VOA	8260D	11/27/24	
P5001-08RE	SPLP-C3-1RE	Water			11/25/24	11/25/24
			SPLP VOA	8260D	12/02/24	
P5001-09	SPLP-C3-2	Water			11/25/24	11/25/24
			SPLP VOA	8260D	11/27/24	
P5001-09DL	SPLP-C3-2DL	Water			11/25/24	11/25/24
			SPLP VOA	8260D	12/02/24	
P5001-10	SPLP-C3-3	Water			11/25/24	11/25/24
			SPLP VOA	8260D	11/27/24	
P5001-10DL	SPLP-C3-3DL	Water			11/25/24	11/25/24
			SPLP VOA	8260D	12/02/24	
P5001-11	SPLP-C4-1	Water			11/25/24	11/25/24
			SPLP VOA	8260D	11/27/24	
P5001-11RE	SPLP-C4-1RE	Water			11/25/24	11/25/24
			SPLP VOA	8260D	12/02/24	



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 788-9222

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

P5001

COC Number: 2042105

Page 1 of 1

6

6.1

CLIENT INFORMATION

PROJECT INFORMATION

BILLING INFORMATION

COMPANY: ENTACT, LLC

ADDRESS: 150 Bay Street, Suite 806

CITY: Jersey City STATE: NJ ZIP: 07302

ATTENTION: Jarod Stanfield

PHONE: 570-886-0442

FAX:

PROJECT NAME: 540 Degraw St Brooklyn, NY

PROJECT #: E9309

LOCATION: Brooklyn, NY

PROJECT MANAGER: Jarod Stanfield

E-MAIL: jstanfield@entact.com

PHONE: 570-886-0442

FAX:

BILL TO: ENTACT, LLC

PO# E9309

ADDRESS: 999 Oakmont Plaza Drive, Suite 300

CITY: Westmont

STATE: IL ZIP: 60559

ATTENTION: Wendy Murray

PHONE: 800-936-8228

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX: 5 DAYS*
HARD COPY: 5 DAYS*
EDD 5 DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

- ☐ RESEULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☐ EDD Format _____

SPLP for VOCs

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		E									
1.	SPLP-C1-1	Soil		X	11/25	10:00	1	X									
2.	SPLP-C1-2	Soil		X	11/25	10:00	1	X									
3.	SPLP-C1-3	Soil		X	11/25	10:00	1	X									
4.	SPLP-C2-1	Soil		X	11/25	10:00	1	X									
5.	SPLP-C2-2	Soil		X	11/25	10:00	1	X									
6.	SPLP-C2-3	Soil		X	11/25	10:00	1	X									
7.	SPLP-C2-3	Soil		X	11/25	10:00	1	X									
8.	SPLP-C3-1	Soil		X	11/25	10:00	1	X									
9.	SPLP-C3-2	Soil		X	11/25	10:00	1	X									
10.	SPLP-C3-3	Soil		X	11/25	10:00	1	X									

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER 1. Jarod Stanfield	DATE/TIME 11/25 12:30	RECEIVED BY 1. [Signature]	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☒ Cooler Temp 3.0°C ☐ Ice in Cooler?: _____
RELINQUISHED BY 2. [Signature]	DATE/TIME	RECEIVED BY 2. [Signature]	Comments:
RELINQUISHED BY 3. [Signature]	DATE/TIME 11-25-24 15:15	RECEIVED FOR LAB BY 3. [Signature]	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight ALLIANCE: ☐ Picked Up ☐ Overnight
Page 1 of 2			Shipment Complete ☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

P5001

80 of 82



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 788-9222

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

P5001

COC Number: 2042105

Page 2 of 2

CLIENT INFORMATION

COMPANY: ENTACT, LLC

ADDRESS: 150 Bay Street, Suite 806

CITY Jersey City STATE: NJ ZIP: 07302

ATTENTION: Jarod Stanfield

PHONE: 570-886-0442

FAX:

PROJECT INFORMATION

PROJECT NAME: 540 Degraw St Brooklyn, NY

PROJECT #: E9309

LOCATION: Brooklyn, NY

PROJECT MANAGER: Jarod Stanfield

E-MAIL: jstanfield@entact.com

PHONE: 570-886-0442

FAX:

BILLING INFORMATION

BILL TO: ENTACT, LLC

PO# E9309

ADDRESS: 999 Oakmont Plaza Drive, Suite 300

CITY: Westmont

STATE: IL ZIP: 60559

ATTENTION: Wendy Murray

PHONE: 800-936-8228

ANALYSIS

SPLP for VOCs

10

11

12

13

14

15

16

PRESERVATIVES

COMMENTS

<- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

CHEMTECH
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPESAMPLE
COLLECTION

of Bottles

COMP

GRAB

DATE

TIME

E

1

2

3

4

5

6

7

8

9

1.

SPLP-C4-1

Soil

X

11/25

10:00

1

X

2.

3.

4.

5.

6.

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER
1. Jarod Stanfield

DATE/TIME

11/25 12:30

RECEIVED BY

1.

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

RELINQUISHED BY

DATE/TIME

11-25-24 1515

RECEIVED FOR LAB BY

3.

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant

Cooler Temp 3.0°C

☐ Ice in Cooler?:

Comments:

Page 2 of 2

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ OvernightALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488