

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5001 SAS No.: P5001 SDG NO.: P5001
 Lab File ID: VX044052.D Date Analyzed: 12/02/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	140021	5.54	242098	6.75	204952	10.05
UPPER LIMIT	280042	6.043	484196	7.25	409904	10.549
LOWER LIMIT	70010.5	5.043	121049	6.25	102476	9.549
EPA SAMPLE NO.						
SPLP-C1-1RE	118265	5.55	232463	6.76	203672	10.05
SPLP-C1-2RE	119853	5.54	235670	6.76	212261	10.06
SPLP-C1-3RE	121878	5.54	237284	6.76	212345	10.05
SPLP-C2-1RE	119282	5.54	230844	6.76	204538	10.05
SPLP-C2-2DL	122610	5.54	241771	6.76	210916	10.05
SPLP-C2-3DL	120449	5.55	231562	6.76	204334	10.06
SPLP-C3-1RE	113564	5.55	222024	6.76	199117	10.05
SPLP-C3-2DL	127859	5.54	242833	6.76	215764	10.05
SPLP-C3-3DL	126077	5.54	246510	6.76	214521	10.05
SPLP-C4-1RE	118857	5.55	232004	6.76	201444	10.06
VX1202WBL01	137622	5.54	262897	6.76	221008	10.05
VX1202WBS01	133564	5.55	236701	6.76	191057	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.