

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5001 SAS No.: P5001 SDG NO.: P5001
 Lab File ID: VX044024.D Date Analyzed: 11/27/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	86408	12.018				
UPPER LIMIT	172816	12.518				
LOWER LIMIT	43204	11.518				
EPA SAMPLE NO.						
SPLP-C1-1	89439	12.02				
SPLP-C1-2	98083	12.02				
SPLP-C1-3	98952	12.02				
SPLP-C2-1	95817	12.02				
SPLP-C2-2	96320	12.02				
SPLP-C2-3	94742	12.02				
SPLP-C3-1	94799	12.02				
SPLP-C3-2	95305	12.02				
SPLP-C3-3	90506	12.02				
SPLP-C4-1	88732	12.02				
VX1127WBL01	92049	12.02				
VX1127WBS01	74629	12.02				
VX1127WBSD01	92085	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.