

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5001 SAS No.: P5001 SDG NO.: P5001
 Lab File ID: VX044052.D Date Analyzed: 12/02/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	98861	12.018				
UPPER LIMIT	197722	12.518				
LOWER LIMIT	49430.5	11.518				
EPA SAMPLE NO.						
SPLP-C1-1RE	98468	12.02				
SPLP-C1-2RE	96574	12.02				
SPLP-C1-3RE	101267	12.02				
SPLP-C2-1RE	97308	12.02				
SPLP-C2-2DL	99346	12.02				
SPLP-C2-3DL	95432	12.02				
SPLP-C3-1RE	92954	12.02				
SPLP-C3-2DL	100587	12.02				
SPLP-C3-3DL	101769	12.02				
SPLP-C4-1RE	88047	12.02				
VX1202WBL01	93273	12.02				
VX1202WBS01	90073	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.