

Continuing Calibration Report for SequenceID : FC112724AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067870.D

Aliphatic C9-C12	9816162.000	60.000	3.186	6.461	163602.700	165791.359	1.320
Aliphatic C12-C16	6888015.000	40.000	6.462	9.849	172200.375	170384.422	-1.066
Aliphatic C16-C21	10183410.000	60.000	9.850	13.205	169723.500	169621.553	-0.060
Aliphatic C21-C28	12700436.000	80.000	13.206	16.857	158755.450	157280.060	-0.938
Aliphatic C28-C40	16216453.000	120.000	16.858	21.700	135137.108	121111.336	-11.581
Aliphatic EPH	55804476.000	360.000	3.186	21.700	155012.433	150155.324	-3.235